

7470A


LB125903

INSTRUMENT ID: 012

Linear

 μ Abs.:

46517



A= 0.0000e+000

B= 2.1483e-004

C= 5.3654e-002

Rho= 0.9994675

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	7.11
0.0	0.000	0.071	0.071	82	0.000	82	0				1
0.2	0.200	0.222	0.022	783	0.0 %	783					11
2.5	2.500	2.316	-0.184	10533	0.0 %	10533					7
5.0	5.000	5.171	0.171	23821	0.0 %	23821					3
7.5	7.500	7.372	-0.128	34068	0.0 %	34068					2
10.0	10.000	10.047	0.047	46517	0.0 %	46517					0

LB125903 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jun 2023 10:39:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	82	-	0.0000	7470A	PPB	12 Jun 2023 10:56:34	S	Std
0.2 - 1	50.2	783	-	0.2000	7470A	PPB	12 Jun 2023 10:58:51	S	Std
2.5 - 1	52.5	10533	-	2.5000	7470A	PPB	12 Jun 2023 11:01:08	S	Std
5.0 - 1	55	23821	-	5.0000	7470A	PPB	12 Jun 2023 11:03:25	S	Std
7.5 - 1	57.5	34068	-	7.5000	7470A	PPB	12 Jun 2023 11:08:17	S	Std
10.0 - 1	518	46517	-	10.0000	7470A	PPB	12 Jun 2023 11:13:11	S	Std
ICV16 - 1	ICV16	19100	4.1569	-	7470A	PPB	12 Jun 2023 11:16:01	U	SMPL
ICB16 - 1	ICB16	-140	0.0236	-	7470A	PPB	12 Jun 2023 11:18:16	U	SMPL
CCV18 - 1	CCV18	22662	4.9221	-	7470A	PPB	12 Jun 2023 11:20:32	U	SMPL
CCB18 - 1	CCB18	-147	0.0221	-	7470A	PPB	12 Jun 2023 11:22:49	U	SMPL
CRA - 1	CRA	665	0.1965	-	7470A	PPB	12 Jun 2023 11:25:06	U	SMPL
HighStd - 1	HighStd	45188	9.7614	-	7470A	PPB	12 Jun 2023 11:27:22	U	SMPL
ChkStd - 1	ChkStd	31247	6.7664	-	7470A	PPB	12 Jun 2023 11:29:37	U	SMPL
PB153335BL - 1	PBW	-221	0.0062	-	7470A	PPB	12 Jun 2023 11:31:55	U	SMPL
PB153335BS - 1	LCSW	15820	3.4523	-	7470A	PPB	12 Jun 2023 11:34:13	U	SMPL
O3092-02 - 1	TP-1	-124	0.0270	-	7470A	PPB	12 Jun 2023 11:36:30	U	SMPL
O3092-05 - 1	TP-2	-32	0.0468	-	7470A	PPB	12 Jun 2023 11:38:47	U	SMPL
O3096-02 - 1	TP-24-0A	-15	0.0504	-	7470A	PPB	12 Jun 2023 11:41:04	U	SMPL
O3096-05 - 1	TP-25-0A	42	0.0627	-	7470A	PPB	12 Jun 2023 11:43:22	U	SMPL
O3109-02 - 1	SP-1-COMP	60	0.0665	-	7470A	PPB	12 Jun 2023 11:45:37	U	SMPL
CCV19 - 1	CCV19	22107	4.8029	-	7470A	PPB	12 Jun 2023 11:47:53	U	SMPL
CCB19 - 1	CCB19	-201	0.0105	-	7470A	PPB	12 Jun 2023 11:50:09	U	SMPL
O3109-02DUP - 1	SP-1-COMP	24	0.0588	-	7470A	PPB	12 Jun 2023 11:52:26	U	SMPL
O3109-02MS - 1	SP-1-COMPMS	18582	4.0456	-	7470A	PPB	12 Jun 2023 11:54:41	U	SMPL
O3109-02MSD - 1	SP-1-COMPMSD	17455	3.8035	-	7470A	PPB	12 Jun 2023 11:56:57	U	SMPL
PB153303TB - 1	PB153303TB	-64	0.0399	-	7470A	PPB	12 Jun 2023 11:59:13	U	SMPL
O3109-02LX5 - 1		-7	0.0522	-	7470A	PPB	12 Jun 2023 12:01:30	U	SMPL
O3109-02A - 1		16803	3.6634	-	7470A	PPB	12 Jun 2023 12:03:47	U	SMPL
PB153257TB - 1	PB153195TB	-142	0.0231	-	7470A	PPB	12 Jun 2023 12:06:03	U	SMPL
CCV20 - 1	CCV20	22597	4.9081	-	7470A	PPB	12 Jun 2023 12:08:22	U	SMPL
CCB20 - 1	CCB20	-177	0.0156	-	7470A	PPB	12 Jun 2023 12:10:39	U	SMPL