

LB125907


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

38858



A= 0.0000e+000

B= 2.6014e-004 slope

C= -5.6106e-002 y-intercept

Rho= 0.9994854

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% 1)
0.00	0.000	-0.045	-0.045	43	0.000	43					1
0.05	0.050					302	0				-9
0.20	0.200	0.183	-0.017	920	0.0 %	920					2
2.50	2.500	2.547	0.047	10005	0.0 %	10005					4
5.00	5.000	5.177	0.177	20115	0.0 %	20115					3
7.50	7.500	7.286	-0.214	28225	0.0 %	28225					1
10.0	10.000	10.052	0.052	38858	0.0 %	38858					

LB125907

INTRUMENT ID :CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	43	-		0 SFAM01.1	PPB	6/12/2023 13:52	Std
	0.2 S01	920	-		0.2 SFAM01.1	PPB	6/12/2023 14:02	Std
	2.5 S02	10005	-		2.5 SFAM01.1	PPB	6/12/2023 14:04	Std
	5 S03	20115	-		5 SFAM01.1	PPB	6/12/2023 14:11	Std
	7.5 S04	28225	-		7.5 SFAM01.1	PPB	6/12/2023 14:18	Std
	10 S05	38858	-		10 SFAM01.1	PPB	6/12/2023 14:28	Std
ICV017	ICV017	15744	4.0395 -		SFAM01.1	PPB	6/12/2023 14:39	SMPL
ICB017	ICB017	11	-0.0532 -		SFAM01.1	PPB	6/12/2023 14:41	SMPL
CCV021	CCV021	20307	5.2265 -		SFAM01.1	PPB	6/12/2023 14:43	SMPL
CCB021	CCB021	5	-0.0548 -		SFAM01.1	PPB	6/12/2023 14:46	SMPL
PB153383BL	PBS383	6	-0.0545 -		SFAM01.1	PPB	6/12/2023 14:48	SMPL
O3057-01	GBY63	20853	5.3685 -		SFAM01.1	PPB	6/12/2023 14:50	SMPL
O3057-02	GBY64	9151	2.3244 -		SFAM01.1	PPB	6/12/2023 14:52	SMPL
O3057-03	GBY65	11714	2.9911 -		SFAM01.1	PPB	6/12/2023 14:55	SMPL
O3057-04	GBY66	14815	3.7978 -		SFAM01.1	PPB	6/12/2023 14:57	SMPL
O3057-05	GBY67	12280	3.1384 -		SFAM01.1	PPB	6/12/2023 14:59	SMPL
O3057-06	GBY68	6279	1.5773 -		SFAM01.1	PPB	6/12/2023 15:02	SMPL
O3057-07	GBY69	8731	2.2152 -		SFAM01.1	PPB	6/12/2023 15:04	SMPL
O3057-08	GBY70	2537	0.6039 -		SFAM01.1	PPB	6/12/2023 15:06	SMPL
O3057-09	GBY71	515	0.0779 -		SFAM01.1	PPB	6/12/2023 15:08	SMPL
O3057-10	GBY72	496	0.0729 -		SFAM01.1	PPB	6/12/2023 15:11	SMPL
O3057-11	GBY72D	511	0.0768 -		SFAM01.1	PPB	6/12/2023 15:13	SMPL
O3057-13	GBY73	718	0.1307 -		SFAM01.1	PPB	6/12/2023 15:17	SMPL
O3057-14	GBY74	176	-0.0103 -		SFAM01.1	PPB	6/12/2023 15:20	SMPL
O3057-15	GBY75	200	-0.0041 -		SFAM01.1	PPB	6/12/2023 15:22	SMPL
O3057-16	GBY76	496	0.0729 -		SFAM01.1	PPB	6/12/2023 15:24	SMPL
O3057-17	GBY77	1467	0.3255 -		SFAM01.1	PPB	6/12/2023 15:26	SMPL
O3057-18	GBY78	189	-0.0069 -		SFAM01.1	PPB	6/12/2023 15:29	SMPL
O3057-19	GBY79	592	0.0979 -		SFAM01.1	PPB	6/12/2023 15:31	SMPL
O3057-12	GBY72S	11318	2.8881 -		SFAM01.1	PPB	6/12/2023 15:34	SMPL
CCV022	CCV022	20206	5.2002 -		SFAM01.1	PPB	6/12/2023 15:36	SMPL
CCB022	CCB022	-47	-0.0683 -		SFAM01.1	PPB	6/12/2023 15:38	SMPL
O3057-20	GBY80	539	0.0841 -		SFAM01.1	PPB	6/12/2023 15:41	SMPL
O3057-21	GBY81	168	-0.0124 -		SFAM01.1	PPB	6/12/2023 15:43	SMPL
PB153384BL	PBS384	13	-0.0527 -		SFAM01.1	PPB	6/12/2023 15:45	SMPL
O3058-01	GBY82	347	0.0342 -		SFAM01.1	PPB	6/12/2023 15:48	SMPL
O3058-02	GBY83	509	0.0763 -		SFAM01.1	PPB	6/12/2023 15:50	SMPL
O3058-03	GBY84	331	0.03 -		SFAM01.1	PPB	6/12/2023 15:52	SMPL
O3058-04	GBY85	346	0.0339 -		SFAM01.1	PPB	6/12/2023 15:54	SMPL
O3058-05	GBY86	374	0.0412 -		SFAM01.1	PPB	6/12/2023 15:57	SMPL
O3058-06	GBY87	654	0.114 -		SFAM01.1	PPB	6/12/2023 15:59	SMPL
O3058-07	GBY88	362	0.0381 -		SFAM01.1	PPB	6/12/2023 16:01	SMPL
O3058-08	GBY89	470	0.0662 -		SFAM01.1	PPB	6/12/2023 16:03	SMPL
O3058-09	GBY90	701	0.1262 -		SFAM01.1	PPB	6/12/2023 16:06	SMPL
O3058-10	GBY91	339	0.0321 -		SFAM01.1	PPB	6/12/2023 16:08	SMPL
O3058-11	GBY92	443	0.0591 -		SFAM01.1	PPB	6/12/2023 16:10	SMPL

LB125907

INTRUMENT ID :CV1

O3058-12	GBY92D	451	0.0612 -	SFAM01.1	PPB	6/12/2023 16:13 SMPL
O3058-13	GBY92S	10361	2.6392 -	SFAM01.1	PPB	6/12/2023 16:17 SMPL
O3058-14	GBY93	516	0.0781 -	SFAM01.1	PPB	6/12/2023 16:20 SMPL
O3058-15	GBY94	294	0.0204 -	SFAM01.1	PPB	6/12/2023 16:22 SMPL
O3058-16	GBY95	441	0.0586 -	SFAM01.1	PPB	6/12/2023 16:24 SMPL
O3058-17	GBY96	302	0.0225 -	SFAM01.1	PPB	6/12/2023 16:27 SMPL
CCV023	CCV023	20229	5.2062 -	SFAM01.1	PPB	6/12/2023 16:29 SMPL
CCB023	CCB023	5	-0.0548 -	SFAM01.1	PPB	6/12/2023 16:34 SMPL
O3058-18	GBY97	345	0.0336 -	SFAM01.1	PPB	6/12/2023 16:36 SMPL
O3058-19	GBY98	349	0.0347 -	SFAM01.1	PPB	6/12/2023 16:38 SMPL
O3058-20	GBY99	290	0.0193 -	SFAM01.1	PPB	6/12/2023 16:41 SMPL
PB153385BL	PBS385	89	-0.033 -	SFAM01.1	PPB	6/12/2023 16:43 SMPL
O3059-01	GBYA1	260	0.0115 -	SFAM01.1	PPB	6/12/2023 16:45 SMPL
O3059-02	GBYA2	381	0.043 -	SFAM01.1	PPB	6/12/2023 16:47 SMPL
O3059-04	GBYA4	421	0.0534 -	SFAM01.1	PPB	6/12/2023 16:50 SMPL
O3059-05	GBYA4D	380	0.0427 -	SFAM01.1	PPB	6/12/2023 16:52 SMPL
O3059-06	GBYA4S	10874	2.7726 -	SFAM01.1	PPB	6/12/2023 16:54 SMPL
O3059-07	GBYA5	293	0.0201 -	SFAM01.1	PPB	6/12/2023 16:56 SMPL
O3059-09	GBYA7	255	0.0102 -	SFAM01.1	PPB	6/12/2023 16:59 SMPL
O3059-10	GBYA8	478	0.0682 -	SFAM01.1	PPB	6/12/2023 17:01 SMPL
O3059-11	GBYA9	368	0.0396 -	SFAM01.1	PPB	6/12/2023 17:03 SMPL
O3059-12	GBYB0	11903	3.0403 -	SFAM01.1	PPB	6/12/2023 17:06 SMPL
O3059-13	GBYB1	6768	1.7045 -	SFAM01.1	PPB	6/12/2023 17:08 SMPL
O3059-14	GBYB2	7558	1.91 -	SFAM01.1	PPB	6/12/2023 17:10 SMPL
O3059-15	GBYB3	23426	6.0379 -	SFAM01.1	PPB	6/12/2023 17:12 SMPL
O3059-16	GBYB4	10018	2.5499 -	SFAM01.1	PPB	6/12/2023 17:15 SMPL
O3059-17	GBYB5	7567	1.9124 -	SFAM01.1	PPB	6/12/2023 17:17 SMPL
O3059-18	GBYB6	529	0.0815 -	SFAM01.1	PPB	6/12/2023 17:19 SMPL
CCV024	CCV024	20534	5.2856 -	SFAM01.1	PPB	6/12/2023 17:21 SMPL
CCB024	CCB024	-3	-0.0569 -	SFAM01.1	PPB	6/12/2023 17:26 SMPL
O3059-19	GBYB7	1110	0.2326 -	SFAM01.1	PPB	6/12/2023 17:29 SMPL
O3059-20	GBYB8	1612	0.3632 -	SFAM01.1	PPB	6/12/2023 17:31 SMPL
O3059-21	GBYB9	1655	0.3744 -	SFAM01.1	PPB	6/12/2023 17:33 SMPL
CCV025	CCV025	20342	5.2356 -	SFAM01.1	PPB	6/12/2023 17:35 SMPL
CCB025	CCB025	-15	-0.06 -	SFAM01.1	PPB	6/12/2023 17:38 SMPL