



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125914

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/13/2023	09:15
2	CAL2	1	Water	NA	NA	20.8	6.99	06/13/2023	09:16
3	CAL3	1	Water	NA	NA	20.7	10.03	06/13/2023	09:17
4	ICV	1	Water	NA	NA	20.7	6.99	06/13/2023	09:20
5	CCV1	1	Water	NA	NA	20.7	2.01	06/13/2023	09:22
6	O3153-04	1	Water	NA	NA	26.2	7.25	06/13/2023	09:25
7	O3153-04DUP	1	Water	NA	NA	26.3	7.26	06/13/2023	09:26
8	O3169-02	1	Water	NA	NA	26.6	8.19	06/13/2023	09:33
9	O3170-01	1	Water	NA	NA	26.3	7.42	06/13/2023	09:37
10	O3170-02	1	Water	NA	NA	26.5	6.08	06/13/2023	09:38
11	CCV2	1	Water	NA	NA	20.8	12.02	06/13/2023	09:40

VB 125914

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o3169 w

WorkList ID : 170810

Department : Wet-Chemistry

Date : 06-13-2023 09:05:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3153-04	A NO3-TRANS-WATER	Water	pH	Cool 4 deg C	PSEG03	D21	06/12/2023	9040C
O3169-02	B T-3089	Water	pH	Cool 4 deg C	PSEG03	D32	06/12/2023	9040C
O3170-01	C AUD-1917	Water	pH	Cool 4 deg C	PSEG03	D31	06/12/2023	9040C
O3170-02	C AUD-2012	Water	pH	Cool 4 deg C	PSEG03	D31	06/12/2023	9040C

Date/Time

06-13-23

09:10

Raw Sample Received by:

JP (WCE)

Raw Sample Relinquished by:

CP SM

Date/Time

06-13-23

12:00

Raw Sample Received by:

CP SM

Raw Sample Relinquished by:

JP (WCE)