



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125915

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/13/2023	09:15
2	CAL2	1	Water	NA	NA	20.8	6.99	06/13/2023	09:16
3	CAL3	1	Water	NA	NA	20.7	10.03	06/13/2023	09:17
4	ICV	1	Water	NA	NA	20.7	6.99	06/13/2023	09:20
5	CCV1	1	Water	NA	NA	20.6	2.02	06/13/2023	09:50
6	O3170-03	1	Solid	20.02	20	25.3	7.08	06/13/2023	10:00
7	O3170-03DUP	1	Solid	20.01	20	25.4	7.09	06/13/2023	10:01
8	CCV2	1	Water	NA	NA	20.8	12.03	06/13/2023	10:05

VB 125915

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o3170 WorkList ID : 170809 Department : Wet-Chemistry Date : 06-13-2023 09:04:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3170-03	AUD-1971	Solid	pH	Cool 4 deg C	PSEG03	D31	06/12/2023	9045D

Date/Time 06.13.23 09:15
Raw Sample Received by: JN/WCI
Raw Sample Relinquished by: CP SM

Date/Time 06.13.23 12:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JN/WCI