

LB125917


245.1

INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

40004



A= 0.0000e+000

B= 2.5117e-004

C= -4.2407e-003

Rho= 0.9999485

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | -0.006 | -0.006 | -8    | 0.000      | -8    |       |       |       |       | 0/0 |
| 0.05   | 0.050  | 0.056  | 0.006  | 238   | 0.0 %      | 238   | 0     |       |       |       | 12  |
| 0.2    | 0.200  | 0.191  | -0.009 | 776   | 0.0 %      | 776   |       |       |       |       | -5  |
| 2.5    | 2.500  | 2.559  | 0.059  | 10206 | 0.0 %      | 10206 |       |       |       |       | 2   |
| 5.0    | 5.000  | 4.959  | -0.041 | 19759 | 0.0 %      | 19759 |       |       |       |       | -1  |
| 7.5    | 7.500  | 7.449  | -0.051 | 29672 | 0.0 %      | 29672 |       |       |       |       | -1  |
| 10.0   | 10.000 | 10.044 | 0.044  | 40004 | 0.0 %      | 40004 |       |       |       |       | 0   |

## LB125917 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Jun 2023 10:39:13

| Sample ID       | Extended ID | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 56          | -8     | -       | 0.0000   | 245.1  | PPB   | 13 Jun 2023 10:52:07 | S    | Std  |
| 0.05 - 1        | 50.05       | 238    | -       | 0.0500   | 245.1  | PPB   | 13 Jun 2023 10:57:00 | S    | Std  |
| 0.2 - 1         | 50.2        | 776    | -       | 0.2000   | 245.1  | PPB   | 13 Jun 2023 11:02:09 | S    | Std  |
| 2.5 - 1         | 52.5        | 10206  | -       | 2.5000   | 245.1  | PPB   | 13 Jun 2023 11:04:26 | S    | Std  |
| 5.0 - 1         | 55          | 19759  | -       | 5.0000   | 245.1  | PPB   | 13 Jun 2023 11:08:57 | S    | Std  |
| 7.5 - 1         | 57.5        | 29672  | -       | 7.5000   | 245.1  | PPB   | 13 Jun 2023 11:18:34 | S    | Std  |
| 10.0 - 1        | 50          | 40004  | -       | 10.0000  | 245.1  | PPB   | 13 Jun 2023 11:45:59 | S    | Std  |
| ICV18 - 1       | ICV18       | 16571  | 4.1579  | -245.1   |        | PPB   | 13 Jun 2023 11:58:03 | U    | SMPL |
| ICB18 - 1       | ICB18       | -116   | -0.0334 | -245.1   |        | PPB   | 13 Jun 2023 12:00:17 | U    | SMPL |
| CCV26 - 1       | CCV26       | 19536  | 4.9027  | -245.1   |        | PPB   | 13 Jun 2023 12:02:33 | U    | SMPL |
| CCB26 - 1       | CCB26       | -122   | -0.0349 | -245.1   |        | PPB   | 13 Jun 2023 12:04:49 | U    | SMPL |
| CRA - 1         | CRA         | 181    | 0.0412  | -245.1   |        | PPB   | 13 Jun 2023 12:09:40 | U    | SMPL |
| HighStd - 1     | HighStd     | 38736  | 9.7251  | -245.1   |        | PPB   | 13 Jun 2023 12:11:55 | U    | SMPL |
| ChkStd - 1      | ChkStd      | 26753  | 6.7154  | -245.1   |        | PPB   | 13 Jun 2023 12:14:10 | U    | SMPL |
| PB153401BL - 1  | PBW         | -69    | -0.0216 | -245.1   |        | PPB   | 13 Jun 2023 12:16:29 | U    | SMPL |
| PB153401BS - 1  | LCSW        | 14951  | 3.7510  | -245.1   |        | PPB   | 13 Jun 2023 12:21:26 | U    | SMPL |
| O3069-11 - 1    | TWP08       | 1624   | 0.4037  | -245.1   |        | PPB   | 13 Jun 2023 12:28:15 | U    | SMPL |
| O3069-11DUP - 1 | TWP08DUP    | 1425   | 0.3537  | -245.1   |        | PPB   | 13 Jun 2023 12:30:31 | U    | SMPL |
| O3069-11MS - 1  | TWP08MS     | 17581  | 4.4116  | -245.1   |        | PPB   | 13 Jun 2023 12:42:53 | U    | SMPL |
| O3069-11MSD - 1 | TWP08MSD    | 17422  | 4.3717  | -245.1   |        | PPB   | 13 Jun 2023 12:45:09 | U    | SMPL |
| CCV27 - 1       | CCV27       | 19464  | 4.8846  | -245.1   |        | PPB   | 13 Jun 2023 12:47:26 | U    | SMPL |
| CCB27 - 1       | CCB27       | -51    | -0.0171 | -245.1   |        | PPB   | 13 Jun 2023 12:49:42 | U    | SMPL |
| O3069-11LX5 - 1 |             | 468    | 0.1133  | -245.1   |        | PPB   | 13 Jun 2023 12:54:37 | U    | SMPL |
| O3069-11A - 1   |             | 19132  | 4.8012  | -245.1   |        | PPB   | 13 Jun 2023 12:56:52 | U    | SMPL |
| CCV28 - 1       | CCV28       | 19408  | 4.8705  | -245.1   |        | PPB   | 13 Jun 2023 12:59:08 | U    | SMPL |
| CCB28 - 1       | CCB28       | -69    | -0.0216 | -245.1   |        | PPB   | 13 Jun 2023 13:01:25 | U    | SMPL |