



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/14/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 06/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:50
Out Date: 06/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125918

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3144-01	EZAP9	1	1.13	8.44	9.57	8.08	82.3	
O3144-02	EZAQ0	2	1.15	8.52	9.67	8.43	85.4	
O3144-03	EZAQ1	3	1.14	8.49	9.63	8.35	84.9	
O3144-04	EZAQ2	4	1.18	8.48	9.66	8.61	87.6	
O3144-05	EZAQ3	5	1.11	8.74	9.85	8.81	88.1	
O3144-06	EZAQ4	6	1.18	8.47	9.65	8.44	85.7	
O3144-07	EZAQ5	7	1.19	8.46	9.65	8.73	89.1	
O3144-08	EZAQ6	8	1.15	8.51	9.66	8.73	89.1	
O3144-09	EZAQ7	9	1.18	8.44	9.62	8.44	86.0	
O3144-10	EZAQ8	10	1.14	8.52	9.66	8.58	87.3	
O3144-11	EZAQ9	11	1.19	8.59	9.78	8.77	88.2	
O3144-12	EZAR0	12	1.13	8.68	9.81	8.67	86.9	
O3144-13	EZAR1	13	1.16	8.63	9.79	8.48	84.8	
O3144-14	EZAR2	14	1.15	8.79	9.94	8.87	87.8	
O3144-15	EZAR3	15	1.15	8.53	9.68	8.39	84.9	
O3144-16	EZAR4	16	1.18	8.42	9.6	8.54	87.4	
O3144-17	EZAR5	17	1.12	8.48	9.6	8.63	88.6	
O3144-18	EZAR6	18	1.18	8.52	9.7	8.45	85.3	
O3144-19	EZAR7	19	1.16	8.80	9.96	8.52	83.6	
O3144-20	EZAS1	20	1.14	8.47	9.61	8.24	83.8	
O3144-21	EZAS1D	21	1.14	8.47	9.61	8.24	83.8	
O3144-22	EZAS1S	22	1.14	8.47	9.61	8.24	83.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$