



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125922

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/13/2023	12:25
2	CAL2	1	Water	NA	NA	20.7	7.00	06/13/2023	12:26
3	CAL3	1	Water	NA	NA	20.7	10.02	06/13/2023	12:28
4	ICV	1	Water	NA	NA	20.8	6.98	06/13/2023	12:30
5	CCV1	1	Water	NA	NA	20.7	2.01	06/13/2023	12:33
6	O3175-01	1	Water	NA	NA	22.1	4.53	06/13/2023	12:37
7	O3175-01DUP	1	Water	NA	NA	22.2	4.54	06/13/2023	12:38
8	O3175-02	1	Water	NA	NA	22.2	4.70	06/13/2023	12:40
9	O3175-03	1	Water	NA	NA	22.1	4.56	06/13/2023	12:45
10	O3176-01	1	Water	NA	NA	22.2	4.59	06/13/2023	12:50
11	CCV2	1	Water	NA	NA	20.7	12.02	06/13/2023	12:55

125922

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 03175

WorkList ID : 170837

Department : Wet-Chemistry

Date : 06-13-2023 12:10:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
03175-01	A CELL-A-POOL-WATER	Water	pH	Cool 4 deg C	STER07	D11	06/09/2023	9040C
03175-02	A CELL-B/C-POOL-WATER	Water	pH	Cool 4 deg C	STER07	D11	06/09/2023	9040C
03175-03	A CELL-D-POOL-WATER	Water	pH	Cool 4 deg C	STER07	D11	06/09/2023	9040C
03176-01	A POOL-WATER	Water	pH	Cool 4 deg C	STER08	D11	06/08/2023	9040C

Date/Time 06-13-23 12:20
 Raw Sample Received by: J.C. (sam)
 Raw Sample Relinquished by: J.C. (sam)

Date/Time 06-13-23 15:00
 Raw Sample Received by: J.C. (sam)
 Raw Sample Relinquished by: J.C. (sam)