

L9125926


7470A

INSTRUMENT ID: EV1

Linear

μ Abs.:

49237



A= 0.0000e+000

B= 2.0430e-004

C= -8.4626e-003

Rho= 0.9999399

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/21
0.0	0.000	0.034	0.034	210	0.000	210	0				-
0.2	0.200	0.194	-0.006	990	0.0 %	990					3
2.5	2.500	2.476	-0.024	12163	0.0 %	12163					-
5.0	5.000	5.017	0.017	24596	0.0 %	24596					0
7.5	7.500	7.428	-0.072	36399	0.0 %	36399					-
10.0	10.000	10.051	0.051	49237	0.0 %	49237					-

LB125926 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jun 2023 13:36:58

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	210	-	0.0000	7470A	PPB	13 Jun 2023 13:38:58	S	Std
0.2 - 1	50.2	990	-	0.2000	7470A	PPB	13 Jun 2023 13:41:16	S	Std
2.5 - 1	52.5	12163	-	2.5000	7470A	PPB	13 Jun 2023 13:43:33	S	Std
5.0 - 1	55	24596	-	5.0000	7470A	PPB	13 Jun 2023 13:45:50	S	Std
7.5 - 1	52.5	36399	-	7.5000	7470A	PPB	13 Jun 2023 13:50:40	S	Std
10.0 - 1	50	49237	-	10.0000	7470A	PPB	13 Jun 2023 13:55:29	S	Std
ICV19 - 1	ICV19	19632	4.0024	-	7470A	PPB	13 Jun 2023 13:58:13	U	SMPL
ICB19 - 1	ICB19	-22	-0.0130	-	7470A	PPB	13 Jun 2023 14:00:28	U	SMPL
CCV29 - 1	CCV29	24886	5.0758	-	7470A	PPB	13 Jun 2023 14:02:45	U	SMPL
CCB29 - 1	CCB29	1	-0.0083	-	7470A	PPB	13 Jun 2023 14:05:00	U	SMPL
CRA - 1	CRA	924	0.1803	-	7470A	PPB	13 Jun 2023 14:07:16	U	SMPL
HighStd - 1	HighStd	48798	9.9611	-	7470A	PPB	13 Jun 2023 14:09:32	U	SMPL
ChkStd - 1	ChkStd	33450	6.8255	-	7470A	PPB	13 Jun 2023 14:11:47	U	SMPL
PB153369BL - 1	PBW	-148	-0.0387	-	7470A	PPB	13 Jun 2023 14:14:05	U	SMPL
PB153369BS - 1	LCSW	20598	4.1998	-	7470A	PPB	13 Jun 2023 14:16:23	U	SMPL
O3108-01 - 1	OU3-FC-BACKFILL-01R	512	0.0961	-	7470A	PPB	13 Jun 2023 14:23:45	U	SMPL
O3108-01DUP - 1	OU3-FC-BACKFILL-01RDUP	330	0.0590	-	7470A	PPB	13 Jun 2023 14:26:02	U	SMPL
O3108-01MS - 1	OU3-FC-BACKFILL-01RMS	23295	4.7508	-	7470A	PPB	13 Jun 2023 14:28:19	U	SMPL
O3108-01MSD - 1	OU3-FC-BACKFILL-01RMSD	23410	4.7743	-	7470A	PPB	13 Jun 2023 14:30:36	U	SMPL
O3108-02 - 1	OU3-FC-BACKFILL-02R	957	0.1871	-	7470A	PPB	13 Jun 2023 14:32:52	U	SMPL
CCV30 - 1	CCV30	24324	4.9610	-	7470A	PPB	13 Jun 2023 14:35:09	U	SMPL
CCB30 - 1	CCB30	80	0.0079	-	7470A	PPB	13 Jun 2023 14:37:24	U	SMPL
PB153371BL - 1	PBW	123	0.0167	-	7470A	PPB	13 Jun 2023 14:39:40	U	SMPL
PB153371BS - 1	LCSW	21534	4.3910	-	7470A	PPB	13 Jun 2023 14:41:56	U	SMPL
O3114-02 - 1	MARION-SOIL-0609	25	-0.0034	-	7470A	PPB	13 Jun 2023 14:44:12	U	SMPL
O3117-03 - 1	TP-26OA	91	0.0101	-	7470A	PPB	13 Jun 2023 14:46:28	U	SMPL
O3117-06 - 1	TP-2GA	131	0.0183	-	7470A	PPB	13 Jun 2023 14:48:44	U	SMPL
O3122-01 - 1	M-060823RB01	239	0.0404	-	7470A	PPB	13 Jun 2023 14:51:00	U	SMPL
O3136-03 - 1	ELI-SUB-LD-CON	160	0.0242	-	7470A	PPB	13 Jun 2023 14:53:17	U	SMPL
O3140-03 - 1	TP-3-GA	205	0.0334	-	7470A	PPB	13 Jun 2023 14:55:34	U	SMPL
O3140-06 - 1	TP-6-GA	159	0.0240	-	7470A	PPB	13 Jun 2023 14:57:52	U	SMPL
O3140-06DUP - 1	TP-6-GADUP	140	0.0201	-	7470A	PPB	13 Jun 2023 15:00:08	U	SMPL
CCV31 - 1	CCV31	24269	4.9498	-	7470A	PPB	13 Jun 2023 15:02:23	U	SMPL
CCB31 - 1	CCB31	73	0.0065	-	7470A	PPB	13 Jun 2023 15:04:39	U	SMPL
O3140-06MS - 1	TP-6-GAMS	21830	4.4515	-	7470A	PPB	13 Jun 2023 15:06:56	U	SMPL
O3140-06MSD - 1	TP-6-GAMSD	21232	4.3293	-	7470A	PPB	13 Jun 2023 15:09:12	U	SMPL
PB153372BL - 1	PBW	-13	-0.0111	-	7470A	PPB	13 Jun 2023 15:11:29	U	SMPL
PB153372BS - 1	LCSW	19700	4.0163	-	7470A	PPB	13 Jun 2023 15:13:45	U	SMPL
O3072-04 - 1	30998	-61	-0.0209	-	7470A	PPB	13 Jun 2023 15:16:01	U	SMPL
O3072-04DUP - 1	30998DUP	77	0.0073	-	7470A	PPB	13 Jun 2023 15:18:18	U	SMPL
O3072-04MS - 1	30998MS	21226	4.3281	-	7470A	PPB	13 Jun 2023 15:27:55	U	SMPL
O3072-04MSD - 1	30998MSD	21176	4.3178	-	7470A	PPB	13 Jun 2023 15:30:12	U	SMPL
O3137-01 - 1	ACEDEMY-STREET-DRUM	10017	2.0380	-	7470A	PPB	13 Jun 2023 15:32:31	U	SMPL
CCV32 - 1	CCV32	23931	4.8807	-	7470A	PPB	13 Jun 2023 15:48:10	U	SMPL
CCB32 - 1	CCB32	-69	-0.0226	-	7470A	PPB	13 Jun 2023 15:50:25	U	SMPL
PB153369TB - 1	PB153334TB	49	0.0015	-	7470A	PPB	13 Jun 2023 15:52:43	U	SMPL
PB153337TB - 1	PB153337TB	141	0.0203	-	7470A	PPB	13 Jun 2023 15:54:59	U	SMPL
PB153345TB - 1	PB153345TB	176	0.0275	-	7470A	PPB	13 Jun 2023 15:57:15	U	SMPL
O3108-01LX5 - 1		-13	-0.0111	-	7470A	PPB	13 Jun 2023 16:02:40	U	SMPL
O3108-01A - 1		21288	4.3407	-	7470A	PPB	13 Jun 2023 16:04:56	U	SMPL
O3140-06LX5 - 1		-95	-0.0279	-	7470A	PPB	13 Jun 2023 16:07:12	U	SMPL
O3140-06A - 1		19329	3.9405	-	7470A	PPB	13 Jun 2023 16:09:30	U	SMPL
O3072-04LX5 - 1		-88	-0.0264	-	7470A	PPB	13 Jun 2023 16:11:46	U	SMPL
O3072-04A - 1		23367	4.7655	-	7470A	PPB	13 Jun 2023 16:14:04	U	SMPL
CCV33 - 1	CCV33	25441	5.1892	-	7470A	PPB	13 Jun 2023 16:16:21	U	SMPL
CCB33 - 1	CCB33	24	-0.0036	-	7470A	PPB	13 Jun 2023 16:18:39	U	SMPL