

LB125930

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

6/13/2023 13:55

Reviewed by : 12 Instrument ID : Konelab

Test: Total CN

Sample Id	^{m/l} Result	Dil. 1 +	Response	Errors
ICV1	94.373	0.0	0.070	
ICB1	1.261	0.0	0.002	
CCV1	245.180	0.0	0.182	
CCB1	0.858	0.0	0.001	
PB153387BL	1.100	0.0	0.001	
O3082-02	1.188	0.0	0.002	
O3082-02DUP	1.224	0.0	0.002	
CCV2	251.905	0.0	0.187	
CCB2	1.444	0.0	0.002	

N	9
Mean	66.504
SD	107.6333
CV%	161.85

Aquakem v. 7.2AQ1

Results from time period:

Tue Jun 13 13:29:26 2023

Tue Jun 13 13:54:39 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	1.1354	µg/l	6/13/2023 10:55:12	
5.OPPBCN	A	Total CN	P	5.7778	µg/l	6/13/2023 10:55:13	
10PPBCN	A	Total CN	P	10.1307	µg/l	6/13/2023 10:55:14	
50PPBCN	A	Total CN	P	47.7658	µg/l	6/13/2023 10:55:15	
100PPBCN	A	Total CN	P	98.2459	µg/l	6/13/2023 10:55:16	
250PPBCN	A	Total CN	P	252.7612	µg/l	6/13/2023 10:55:17	
500PPBCN	A	Total CN	P	499.1833	µg/l	6/13/2023 10:55:18	
ICV1	S	Total CN	P	94.3731	µg/l	6/13/2023 13:29:27	
ICB1	S	Total CN	P	1.2613	µg/l	6/13/2023 13:29:28	
CCV1	S	Total CN	P	245.1801	µg/l	6/13/2023 13:29:30	
CCB1	S	Total CN	P	0.8578	µg/l	6/13/2023 13:29:33	
PB153387BL	S	Total CN	P	1.1004	µg/l	6/13/2023 13:29:35	
O3082-02	S	Total CN	P	1.1878	µg/l	6/13/2023 13:54:32	
O3082-02DUP	S	Total CN	P	1.2242	µg/l	6/13/2023 13:54:33	
CCV2	S	Total CN	P	251.9048	µg/l	6/13/2023 13:54:37	
CCB2	S	Total CN	P	1.4437	µg/l	6/13/2023 13:54:39	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

6/13/2023 10:55

Reviewed by : R Instrument ID : Konelab


Test Total CN

Accepted 6/13/2023 10:55

Factor 1355
Bias 0.001

Coeff. of det. 0.999911

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	1.1354	0.0000	-
2	5.0PPBCN	0.005	5.7778	5.0000	15.6
3	10PPBCN	0.008	10.1307	10.0000	1.3
4	50PPBCN	0.036	47.7658	50.0000	-4.5
5	100PPBCN	0.073	98.2459	100.0000	-1.8
6	250PPBCN	0.187	252.7612	250.0000	1.1
7	500PPBCN	0.369	499.1833	500.0000	-0.2

R
6/13/23