



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125936

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	06/14/2023	10:00
2	O3151-04		1	12	100.00	>212.0	>212.0	06/14/2023	10:30
3	O3151-04DUP		1	13	100.00	>212.0	>212.0	06/14/2023	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

06125936

WorkList Name : fp s-6-13

WorkList ID : 170803

Department : Wet-Chemistry

Date : 06-13-2023 08:29:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3151-04	31023	Solid	Flash Point	Cool 4 deg C	PSEG03	D22	06/12/2023	1010B

Date/Time 06/14/2023 09:50
 Raw Sample Received by: RM Cwe
 Raw Sample Relinquished by: [Signature]

Date/Time 06/14/2023 11:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: RM Cwe