

LB125937


7471B

INSTRUMENT 1A: CV1

Linear

 μ Abs.:

6542E



A= 0.0000e+000

B= 1.5256e-004

C= 5.1861e-002

Rho= 0.9998080

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.091	0.091	258	0.000	258					
0.2	0.200	0.229	0.029	1164	0.0 %	1164					
2.5	2.500	2.377	-0.123	15241	0.0 %	15241					
5.0	5.000	4.932	-0.068	31988	0.0 %	31988					
7.5	7.500	7.537	0.037	49060	0.0 %	49060					
10.0	10.000	10.033	0.033	65425	0.0 %	65425					

0/01
15
15
100

LB125937 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jun 2023 16:42:17

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	258	-	0.0000	7471B	PPB	13 Jun 2023 16:43:24	S	Std
0.2 - 1	50-2	1164	-	0.2000	7471B	PPB	13 Jun 2023 16:47:42	S	Std
2.5 - 1	52-5	15241	-	2.5000	7471B	PPB	13 Jun 2023 16:50:00	S	Std
5.0 - 1	55	31988	-	5.0000	7471B	PPB	13 Jun 2023 16:52:17	S	Std
7.5 - 1	57-5	49060	-	7.5000	7471B	PPB	13 Jun 2023 16:54:35	S	Std
10.0 - 1	510	65425	-	10.0000	7471B	PPB	13 Jun 2023 16:59:33	S	Std
ICV20 - 1	ICV20	27851	4.3009	-	7471B	PPB	13 Jun 2023 17:06:51	U	SMPL
ICB20 - 1	ICB20	-64	0.0421	-	7471B	PPB	13 Jun 2023 17:09:07	U	SMPL
CCV34 - 1	CCV34	31978	4.9306	-	7471B	PPB	13 Jun 2023 17:11:23	U	SMPL
CCB34 - 1	CCB34	-87	0.0386	-	7471B	PPB	13 Jun 2023 17:13:39	U	SMPL
CRA - 1	CRA	1136	0.2252	-	7471B	PPB	13 Jun 2023 17:15:56	U	SMPL
HighStd - 1	HighStd	64925	9.9571	-	7471B	PPB	13 Jun 2023 17:18:11	U	SMPL
ChkStd - 1	ChkStd	49324	7.5770	-	7471B	PPB	13 Jun 2023 17:20:26	U	SMPL
PB153391BL - 1	PBS	-199	0.0215	-	7471B	PPB	13 Jun 2023 17:22:45	U	SMPL
PB153391BS - 1	LCSS	22638	3.5056	-	7471B	PPB	13 Jun 2023 17:27:37	U	SMPL
O3114-01 - 1	MARION-SOIL-0609	2208	0.3887	-	7471B	PPB	13 Jun 2023 17:29:53	U	SMPL
O3117-01 - 1	TP-26OA	7314	1.1677	-	7471B	PPB	13 Jun 2023 17:32:13	U	SMPL
O3117-04 - 1	TP-2GA	26128	4.0381	-	7471B	PPB	13 Jun 2023 17:34:30	U	SMPL
O3136-01 - 1	ELI-SUB-LD-CON	510	0.1297	-	7471B	PPB	13 Jun 2023 17:36:48	U	SMPL
CCV35 - 1	CCV35	31487	4.8557	-	7471B	PPB	13 Jun 2023 17:41:21	U	SMPL
CCB35 - 1	CCB35	-95	0.0374	-	7471B	PPB	13 Jun 2023 17:43:36	U	SMPL
O3140-01 - 1	TP-3-GA	26157	4.0425	-	7471B	PPB	13 Jun 2023 17:52:43	U	SMPL
O3140-04 - 1	TP-6-GA	10506	1.6547	-	7471B	PPB	13 Jun 2023 17:55:02	U	SMPL
O3141-01 - 1	CMC-EME-Topsoil-001	1546	0.2877	-	7471B	PPB	13 Jun 2023 17:57:20	U	SMPL
O3141-02 - 1	CMC-EME-Topsoil-002	1612	0.2978	-	7471B	PPB	13 Jun 2023 17:59:37	U	SMPL
O3141-03 - 1	CMC-EME-Topsoil-003	2008	0.3582	-	7471B	PPB	13 Jun 2023 18:01:54	U	SMPL
O3141-04 - 1	CMC-EME-Topsoil-004	1761	0.3205	-	7471B	PPB	13 Jun 2023 18:04:11	U	SMPL
O3155-01 - 1	WC-1	10945	1.7217	-	7471B	PPB	13 Jun 2023 18:06:27	U	SMPL
CCV36 - 1	CCV36	31669	4.8834	-	7471B	PPB	13 Jun 2023 18:08:43	U	SMPL
CCB36 - 1	CCB36	-126	0.0326	-	7471B	PPB	13 Jun 2023 18:10:59	U	SMPL