

L3125945


7471B

INSTRUMENT ID: CVL

Linear

μ Abs.:

6871E



A= 0.0000e+000

B= 1.4344e-004

C= -6.0760e-003

Rho= 0.9995039

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.0	0.000	0.027	0.027	232	0.000	232					
0.2	0.200	0.174	-0.026	1253	0.0 %	1253					-13
2.5	2.500	2.462	-0.038	17207	0.0 %	17207					-2
5.0	5.000	4.954	-0.046	34580	0.0 %	34580					1
7.5	7.500	7.732	0.232	53945	0.0 %	53945					3
10.0	10.000	9.851	-0.149	68716	0.0 %	68716					-1

LB125945 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Jun 2023 10:44:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	232	-	0.0000	7471B	PPB	14 Jun 2023 11:00:20	S	Std
0.2 - 1	50-2	1253	-	0.2000	7471B	PPB	14 Jun 2023 11:05:41	S	Std
2.5 - 1	52-5	17207	-	2.5000	7471B	PPB	14 Jun 2023 11:07:58	S	Std
5.0 - 1	55	34580	-	5.0000	7471B	PPB	14 Jun 2023 11:10:15	S	Std
7.5 - 1	57-5	53945	-	7.5000	7471B	PPB	14 Jun 2023 11:15:14	S	Std
10.0 - 1	51N	68716	-	10.0000	7471B	PPB	14 Jun 2023 11:20:08	S	Std
ICV21 - 1	ICV21	28322	4.0565	-	7471B	PPB	14 Jun 2023 11:22:51	U	SMPL
ICB21 - 1	ICB21	-38	-0.0115	-	7471B	PPB	14 Jun 2023 11:25:08	U	SMPL
CCV38 - 1	CCV38	34400	4.9284	-	7471B	PPB	14 Jun 2023 11:27:25	U	SMPL
CCB38 - 1	CCB38	-39	-0.0117	-	7471B	PPB	14 Jun 2023 11:29:40	U	SMPL
CRA - 1	CRA	1183	0.1636	-	7471B	PPB	14 Jun 2023 11:31:57	U	SMPL
HighStd - 1	HighStd	67589	9.6891	-	7471B	PPB	14 Jun 2023 11:34:12	U	SMPL
ChkStd - 1	ChkStd	49875	7.1482	-	7471B	PPB	14 Jun 2023 11:36:28	U	SMPL
PB153429BL - 1	PB153429BL	601	0.0801	-	7471B	PPB	14 Jun 2023 11:38:47	U	SMPL
PB153429BS - 1	PB153429BS	26083	3.7354	-	7471B	PPB	14 Jun 2023 11:41:06	U	SMPL
O3154-01 - 1	CHRT-24051	4925	0.7004	-	7471B	PPB	14 Jun 2023 11:43:23	U	SMPL
O3154-03 - 1	CHRT-26598	997	0.1369	-	7471B	PPB	14 Jun 2023 11:45:42	U	SMPL
O3160-01 - 1	PNNL-B-0609-1	9421	1.3453	-	7471B	PPB	14 Jun 2023 11:48:01	U	SMPL
O3160-03 - 1	PNNL-B-0609-2	10743	1.5349	-	7471B	PPB	14 Jun 2023 11:50:19	U	SMPL
O3164-01 - 1	SOIL-001	52315	7.4982	-	7471B	PPB	14 Jun 2023 11:52:36	U	SMPL
CCV39 - 1	CCV39	32027	4.5880	-	7471B	PPB	14 Jun 2023 11:54:52	U	SMPL
CCB39 - 1	CCB39	-151	-0.0277	-	7471B	PPB	14 Jun 2023 11:57:11	U	SMPL
O3166-01 - 1	TEST-PIT-6-12-23	7456	1.0634	-	7471B	PPB	14 Jun 2023 11:59:28	U	SMPL
O3167-01 - 1	TP-4-GA	5535	0.7879	-	7471B	PPB	14 Jun 2023 12:01:44	U	SMPL
O3170-03 - 1	AUD-1971	1928	0.2705	-	7471B	PPB	14 Jun 2023 12:04:00	U	SMPL
O3172-01 - 1	BUR-1578	2903	0.4103	-	7471B	PPB	14 Jun 2023 12:06:17	U	SMPL
O3186-01 - 1	OR-03-061323	7230	1.0310	-	7471B	PPB	14 Jun 2023 12:08:34	U	SMPL
O3186-01DUP - 1	OR-03-061323DUP	7323	1.0444	-	7471B	PPB	14 Jun 2023 12:10:50	U	SMPL
O3186-01MS - 1	OR-03-061323MS	30165	4.3209	-	7471B	PPB	14 Jun 2023 12:13:07	U	SMPL
O3186-01MSD - 1	OR-03-061323MSD	31609	4.5280	-	7471B	PPB	14 Jun 2023 12:15:24	U	SMPL
O3214-01 - 1	HYATT-PIPE-TP	34536	4.9479	-	7471B	PPB	14 Jun 2023 12:17:43	U	SMPL
O3221-02 - 1	RT-4194	10793	1.5421	-	7471B	PPB	14 Jun 2023 12:20:01	U	SMPL
CCV40 - 1	CCV40	37917	5.4329	-	7471B	PPB	14 Jun 2023 12:25:04	U	SMPL
CCB40 - 1	CCB40	-136	-0.0256	-	7471B	PPB	14 Jun 2023 12:27:19	U	SMPL
O3186-01LX5 - 1		1407	0.1957	-	7471B	PPB	14 Jun 2023 12:29:37	U	SMPL
O3186-01A - 1		31092	4.4539	-	7471B	PPB	14 Jun 2023 12:31:53	U	SMPL
O3138-01 - 1		7600	1.0841	-	7471B	PPB	14 Jun 2023 12:34:09	U	SMPL
O3138-01DUP - 1		7274	1.0373	-	7471B	PPB	14 Jun 2023 12:36:27	U	SMPL
O3138-01MS - 1		32253	4.6204	-	7471B	PPB	14 Jun 2023 12:43:48	U	SMPL
O3138-01MSD - 1		31603	4.5272	-	7471B	PPB	14 Jun 2023 12:46:06	U	SMPL
O3138-01LX5 - 1		1391	0.1935	-	7471B	PPB	14 Jun 2023 12:55:07	U	SMPL
O3138-01A - 1		37895	5.4297	-	7471B	PPB	14 Jun 2023 12:57:24	U	SMPL
O3157-01 - 1		4757	0.6763	-	7471B	PPB	14 Jun 2023 12:59:41	U	SMPL
CCV41 - 1	CCV41	37924	5.4339	-	7471B	PPB	14 Jun 2023 13:01:59	U	SMPL
CCB41 - 1	CCB41	-6	-0.0069	-	7471B	PPB	14 Jun 2023 13:04:15	U	SMPL