



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/15/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 13:45
In Date: 06/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 06/15/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125951

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3180-01	EZBA6	1	1.12	8.85	9.97	9.38	93.3	
O3180-02	EZBA7	2	1.06	8.79	9.85	9.03	90.7	
O3180-03	EZBA8	3	1.08	8.78	9.86	8.7	86.8	
O3180-04	EZBA9	4	1.13	8.74	9.87	9.28	93.2	
O3180-05	EZBB0	5	1.1	8.87	9.97	9.06	89.7	
O3180-06	EZBB1	6	1.09	8.85	9.94	8.67	85.6	
O3180-07	EZBB2	7	1.09	8.88	9.97	9.71	97.1	
O3180-08	EZBB3	8	1.09	8.73	9.82	9.55	96.9	
O3180-09	EZBB4	9	1.1	8.71	9.81	9.05	91.3	
O3180-10	EZBB5	10	1.1	8.76	9.86	8.95	89.6	
O3180-11	EZBB6	11	1.09	8.73	9.82	9.16	92.4	
O3180-12	EZBB7	12	1.11	8.79	9.9	9.11	91.0	
O3180-13	EZBB8	13	1.08	8.84	9.92	8.76	86.9	
O3180-14	EZBB9	14	1.13	8.70	9.83	8.96	90.0	
O3180-15	EZBC0	15	1.1	8.77	9.87	8.81	87.9	
O3180-16	EZBC1	16	1.09	8.81	9.9	8.55	84.7	
O3180-17	EZBC2	17	1.07	8.84	9.91	8.84	87.9	
O3180-18	EZBC3	18	1.1	8.72	9.82	8.48	84.6	
O3180-19	EZBC4	19	1.09	8.89	9.98	8.67	85.3	
O3180-20	EZBC5	20	1.11	8.77	9.88	8.29	81.9	
O3180-21	EZBC5D	21	1.11	8.77	9.88	8.29	81.9	
O3180-22	EZBC5S	22	1.11	8.77	9.88	8.29	81.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$