



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB125966

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/15/2023	10:30
2	CAL2	1	Water	NA	NA	20.8	7.00	06/15/2023	10:31
3	CAL3	1	Water	NA	NA	20.7	10.02	06/15/2023	10:33
4	ICV	1	Water	NA	NA	20.8	6.98	06/15/2023	10:35
5	CCV1	1	Water	NA	NA	20.7	2.01	06/15/2023	10:40
6	O3233-04	1	Water	NA	NA	21.9	5.54	06/15/2023	10:45
7	O3233-04DUP	1	Water	NA	NA	21.8	5.55	06/15/2023	10:46
8	O3233-05	1	Water	NA	NA	22.6	7.56	06/15/2023	10:49
9	CCV2	1	Water	NA	NA	20.8	12.02	06/15/2023	11:00

125966

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRISIVITY O3233 w WorkList ID : 170907 Department : Wet-Chemistry Date : 06-15-2023 10:16:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3233-04	ELI-0042	Water	Corrosivity	Cool 4 deg C	PSEG03	G11	06/14/2023	9040C
O3233-05	ELI-0047	Water	Corrosivity	Cool 4 deg C	PSEG03	G11	06/14/2023	9040C

06.15.23 10:20

Date/Time
Raw Sample Received by: JP (wcc)
Raw Sample Relinquished by: JP (wcc)

Date/Time
Raw Sample Received by: JP (wcc)
Raw Sample Relinquished by: JP (wcc)