



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 6/16/2023

OVENTEMP IN Celsius(°C): 108
Time IN: 14:25
In Date: 06/15/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 06/16/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125968

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3190-01	EZBR9	1	1.13	8.73	9.86	8.81	88.0	
O3190-02	EZBS0	2	1.12	8.86	9.98	9.46	94.1	
O3190-03	EZBS1	3	1.14	8.73	9.87	9.21	92.4	
O3190-04	EZBS2	4	1.12	8.70	9.82	8.86	89.0	
O3190-05	EZBS3	5	1.15	8.71	9.86	9.28	93.3	
O3190-06	EZBS4	6	1.13	8.78	9.91	9.18	91.7	
O3190-07	EZBS5	7	1.14	8.79	9.93	9.02	89.6	
O3190-08	EZBS6	8	1.15	8.82	9.97	9.36	93.1	
O3190-09	EZBS7	9	1.13	8.84	9.97	9.25	91.9	
O3190-10	EZBS8	10	1.14	8.83	9.97	9.22	91.5	
O3190-11	EZBS9	11	1.15	8.68	9.83	8.86	88.8	
O3190-12	EZBS9D	12	1.15	8.68	9.83	8.86	88.8	
O3190-13	EZBS9S	13	1.15	8.68	9.83	8.86	88.8	
O3190-14	EZCG2	14	1.14	8.79	9.93	9.14	91.0	
O3190-15	EZCG3	15	1.14	8.70	9.84	8.74	87.4	
O3190-16	EZCG4	16	1.13	8.83	9.96	8.52	83.7	
O3190-17	EZCG5	17	1.13	8.85	9.98	9.44	93.9	
O3190-18	EZCG6	18	1.15	8.82	9.97	8.97	88.7	
O3190-19	EZCG7	19	1.14	8.81	9.95	8.71	85.9	
O3190-20	EZCG8	20	1.14	8.83	9.97	9.14	90.6	
O3190-21	EZCG9	21	1.14	8.84	9.98	8.82	86.9	
O3190-22	EZCH0	22	1.14	8.74	9.88	8.7	86.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$