

LB125972


7470A

INSTRUMENT ID: CY1

Linear

μ Abs.:

4968E



A= 0.0000e+000

B= 2.0008e-004

C= -1.1900e-001

Rho= 0.9993286

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.015	0.015	671	0.000	671	0				-
0.2	0.200	0.192	-0.008	1553	0.0 %	1553					-4
2.5	2.500	2.443	-0.057	12805	0.0 %	12805					-2
5.0	5.000	4.958	-0.042	25377	0.0 %	25377					-1
7.5	7.500	7.770	0.270	39428	0.0 %	39428					4
10.0	10.000	9.822	-0.178	49686	0.0 %	49686					-2

LB125972 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jun 2023 11:34:27

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	671	-	0.0000	7470A	PPB	15 Jun 2023 11:40:22	S	Std
0.2 - 1	502	1553	-	0.2000	7470A	PPB	15 Jun 2023 11:42:39	S	Std
2.5 - 1	52.5	12805	-	2.5000	7470A	PPB	15 Jun 2023 11:44:57	S	Std
5.0 - 1	52.5	25377	-	5.0000	7470A	PPB	15 Jun 2023 11:47:14	S	Std
7.5 - 1	57.5	39428	-	7.5000	7470A	PPB	15 Jun 2023 11:49:31	S	Std
10.0 - 1	50	49686	-	10.0000	7470A	PPB	15 Jun 2023 11:51:48	S	Std
ICV23 - 1	ICV23	20115	3.9056	-	7470A	PPB	15 Jun 2023 11:59:17	U	SMPL
ICB23 - 1	ICB23	409	-0.0372	-	7470A	PPB	15 Jun 2023 12:01:32	U	SMPL
CCV42 - 1	CCV42	24681	4.8191	-	7470A	PPB	15 Jun 2023 12:03:50	U	SMPL
CCB42 - 1	CCB42	370	-0.0450	-	7470A	PPB	15 Jun 2023 12:06:05	U	SMPL
CRA - 1	CRA	1647	0.2105	-	7470A	PPB	15 Jun 2023 12:11:14	U	SMPL
HighStd - 1	HighStd	49341	9.7530	-	7470A	PPB	15 Jun 2023 12:13:28	U	SMPL
ChkStd - 1	ChkStd	36645	7.2128	-	7470A	PPB	15 Jun 2023 12:15:44	U	SMPL
PB153441BL - 1	PBW	289	-0.0612	-	7470A	PPB	15 Jun 2023 12:23:24	U	SMPL
PB153441BS - 1	LCSW	19786	3.8397	-	7470A	PPB	15 Jun 2023 12:25:41	U	SMPL
O3157-05 - 1	SU-622-COMP-01	84	-0.1022	-	7470A	PPB	15 Jun 2023 12:27:57	U	SMPL
O3160-02 - 1	PNNL-B-0609-1	193	-0.0804	-	7470A	PPB	15 Jun 2023 12:30:15	U	SMPL
O3160-04 - 1	PNNL-B-0609-2	351	-0.0488	-	7470A	PPB	15 Jun 2023 12:32:33	U	SMPL
O3163-01 - 1	149TH-ST-DRUM-15-POLY	420	-0.0350	-	7470A	PPB	15 Jun 2023 12:34:50	U	SMPL
O3163-02 - 1	149TH-ST-DRUM-16-PPE	310	-0.0570	-	7470A	PPB	15 Jun 2023 12:37:05	U	SMPL
CCV43 - 1	CCV43	25112	4.9053	-	7470A	PPB	15 Jun 2023 12:39:21	U	SMPL
CCB43 - 1	CCB43	386	-0.0418	-	7470A	PPB	15 Jun 2023 12:41:36	U	SMPL
O3164-03 - 1	SOIL-001	340	-0.0510	-	7470A	PPB	15 Jun 2023 12:43:52	U	SMPL
O3166-04 - 1	TEST-PIT-6-12-23	238	-0.0714	-	7470A	PPB	15 Jun 2023 12:46:08	U	SMPL
O3167-03 - 1	TP-4-GA	277	-0.0636	-	7470A	PPB	15 Jun 2023 12:48:23	U	SMPL
O3169-01 - 1	D-3090	366	-0.0458	-	7470A	PPB	15 Jun 2023 12:50:39	U	SMPL
O3214-03 - 1	HYATT-PIPE-TP	1113	0.1037	-	7470A	PPB	15 Jun 2023 12:58:02	U	SMPL
O3214-03DUP - 1	HYATT-PIPE-TPDUP	1114	0.1039	-	7470A	PPB	15 Jun 2023 13:00:17	U	SMPL
O3214-03MS - 1	HYATT-PIPE-TPMS	17621	3.4066	-	7470A	PPB	15 Jun 2023 13:02:35	U	SMPL
O3214-03MSD - 1	HYATT-PIPE-TPMSD	17402	3.3627	-	7470A	PPB	15 Jun 2023 13:04:52	U	SMPL
PB153462BL - 1	PBW	278	-0.0634	-	7470A	PPB	15 Jun 2023 13:07:09	U	SMPL
PB153462BS - 1	LCSW	20640	4.0106	-	7470A	PPB	15 Jun 2023 13:09:25	U	SMPL
CCV44 - 1	CCV44	24657	4.8143	-	7470A	PPB	15 Jun 2023 13:11:41	U	SMPL
CCB44 - 1	CCB44	264	-0.0662	-	7470A	PPB	15 Jun 2023 13:13:58	U	SMPL
O3170-01 - 1	AUD-1917	737	0.0285	-	7470A	PPB	15 Jun 2023 13:16:15	U	SMPL
O3170-01DUP - 1	AUD-1917DUP	506	-0.0178	-	7470A	PPB	15 Jun 2023 13:18:31	U	SMPL
O3170-01MS - 1	AUD-1917MS	10320	1.9458	-	7470A	PPB	15 Jun 2023 13:20:47	U	SMPL
O3170-01MSD - 1	AUD-1917MSD	8413	1.5642	-	7470A	PPB	15 Jun 2023 13:23:03	U	SMPL
O3170-02 - 1	AUD-2012	259	-0.0672	-	7470A	PPB	15 Jun 2023 13:25:23	U	SMPL
O3241-02 - 1	40508	1018	0.0847	-	7470A	PPB	15 Jun 2023 13:27:41	U	SMPL
CCV45 - 1	CCV45	24527	4.7883	-	7470A	PPB	15 Jun 2023 13:43:33	U	SMPL
CCB45 - 1	CCB45	461	-0.0268	-	7470A	PPB	15 Jun 2023 13:45:50	U	SMPL
PB153392TB - 1	PB153392TB	124	-0.0942	-	7470A	PPB	15 Jun 2023 13:48:08	U	SMPL
O3214-03LX5 - 1		345	-0.0500	-	7470A	PPB	15 Jun 2023 13:50:24	U	SMPL
O3214-03A - 1		19068	3.6961	-	7470A	PPB	15 Jun 2023 13:52:40	U	SMPL
O3170-01LX5 - 1		459	-0.0272	-	7470A	PPB	15 Jun 2023 13:54:56	U	SMPL
O3170-01A - 1		10745	2.0308	-	7470A	PPB	15 Jun 2023 13:57:13	U	SMPL
CCV46 - 1	CCV46	26860	5.2551	-	7470A	PPB	15 Jun 2023 14:12:03	U	SMPL
CCB46 - 1	CCB46	315	-0.0560	-	7470A	PPB	15 Jun 2023 14:14:19	U	SMPL