

LB125980


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

60068



A= 0.0000e+000

B= 1.6792e-004 slope

C= -5.3500e-002 y-intercept

Rho= 0.9999523

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/10/11
0.00	0.000	0.062	0.062	689	0.000	689					
0.05	0.050					302	0				
0.20	0.200	0.184	-0.016	1416	0.0 %	1416					
2.50	2.500	2.456	-0.044	14945	0.0 %	14945					
5.00	5.000	4.983	-0.017	29996	0.0 %	29996					
7.50	7.500	7.481	-0.019	44868	0.0 %	44868					
10.0	10.000	10.033	0.033	60068	0.0 %	60068					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	689	-		0 SFAM01.1	PPB	6/15/2023 14:56	Std
	0.2 S01	1416	-		0.2 SFAM01.1	PPB	6/15/2023 14:58	Std
	2.5 S02	14945	-		2.5 SFAM01.1	PPB	6/15/2023 15:00	Std
	5 S03	29996	-		5 SFAM01.1	PPB	6/15/2023 15:04	Std
	7.5 S04	44868	-		7.5 SFAM01.1	PPB	6/15/2023 15:07	Std
	10 S05	60068	-		10 SFAM01.1	PPB	6/15/2023 15:09	Std
ICV024	ICV024	23722	3.9299 -		SFAM01.1	PPB	6/15/2023 15:15	SMPL
ICB024	ICB024	423	0.0175 -		SFAM01.1	PPB	6/15/2023 15:20	SMPL
CCV047	CCV047	29948	4.9754 -		SFAM01.1	PPB	6/15/2023 15:22	SMPL
CCB047	CCB047	237	-0.0137 -		SFAM01.1	PPB	6/15/2023 15:24	SMPL
PB153473BL	PBS473	617	0.0501 -		SFAM01.1	PPB	6/15/2023 15:26	SMPL
O3060-01	GBYC0	933	0.1032 -		SFAM01.1	PPB	6/15/2023 15:29	SMPL
O3060-02	GBYC1	935	0.1035 -		SFAM01.1	PPB	6/15/2023 15:31	SMPL
O3060-03	GBYC2	1617	0.218 -		SFAM01.1	PPB	6/15/2023 15:33	SMPL
O3060-04	GBYC3	9695	1.5745 -		SFAM01.1	PPB	6/15/2023 15:35	SMPL
O3060-06	GBYC5	6288	1.0024 -		SFAM01.1	PPB	6/15/2023 15:38	SMPL
O3060-09	GBYC8	3681	0.5646 -		SFAM01.1	PPB	6/15/2023 15:40	SMPL
O3060-10	GBYC9	1169	0.1428 -		SFAM01.1	PPB	6/15/2023 15:58	SMPL
O3060-11	GBYD0	965	0.1085 -		SFAM01.1	PPB	6/15/2023 16:00	SMPL
O3060-12	GBYD1	644	0.0546 -		SFAM01.1	PPB	6/15/2023 16:02	SMPL
O3060-13	GBYD2	367	0.0081 -		SFAM01.1	PPB	6/15/2023 16:05	SMPL
O3060-14	GBYD3	781	0.0776 -		SFAM01.1	PPB	6/15/2023 16:07	SMPL
CCV048	CCV048	29504	4.9008 -		SFAM01.1	PPB	6/15/2023 16:09	SMPL
CCB048	CCB048	555	0.0397 -		SFAM01.1	PPB	6/15/2023 16:11	SMPL
O3060-15	GBYD4	410	0.0153 -		SFAM01.1	PPB	6/15/2023 16:14	SMPL
O3060-16	GBYD4D	422	0.0174 -		SFAM01.1	PPB	6/15/2023 16:16	SMPL
O3060-17	GBYD4S	14234	2.3367 -		SFAM01.1	PPB	6/15/2023 16:18	SMPL
O3060-18	GBYD5	749	0.0723 -		SFAM01.1	PPB	6/15/2023 16:20	SMPL
O3060-19	GBYD6	816	0.0835 -		SFAM01.1	PPB	6/15/2023 16:23	SMPL
O3060-20	GBYD7	443	0.0209 -		SFAM01.1	PPB	6/15/2023 16:25	SMPL
O3060-21	GBYD8	421	0.0172 -		SFAM01.1	PPB	6/15/2023 16:27	SMPL
PB153474BL	PBS474	621	0.0508 -		SFAM01.1	PPB	6/15/2023 16:30	SMPL
CCV049	CCV049	29543	4.9074 -		SFAM01.1	PPB	6/15/2023 16:32	SMPL
CCB049	CCB049	62	-0.0431 -		SFAM01.1	PPB	6/15/2023 16:34	SMPL
O3061-01	GBYD9	756	0.0734 -		SFAM01.1	PPB	6/15/2023 16:36	SMPL
O3061-02	GBYE0	6440	1.0279 -		SFAM01.1	PPB	6/15/2023 16:39	SMPL
O3061-04	GBYE2	5899	0.9371 -		SFAM01.1	PPB	6/15/2023 16:41	SMPL
O3061-05	GBYE3	6206	0.9886 -		SFAM01.1	PPB	6/15/2023 16:43	SMPL
O3061-06	GBYE3D	6323	1.0083 -		SFAM01.1	PPB	6/15/2023 16:45	SMPL
O3061-07	GBYE3S	19715	3.2571 -		SFAM01.1	PPB	6/15/2023 16:48	SMPL
PB153475BL	PBS475	540	0.0372 -		SFAM01.1	PPB	6/15/2023 16:52	SMPL
O3063-01	MEZEK9	894	0.0966 -		SFAM01.1	PPB	6/15/2023 16:55	SMPL
O3063-02	MEZEL0	643	0.0545 -		SFAM01.1	PPB	6/15/2023 16:57	SMPL
O3063-03	MEZEL1	517	0.0333 -		SFAM01.1	PPB	6/15/2023 16:59	SMPL
O3063-04	MEZEL1D	475	0.0263 -		SFAM01.1	PPB	6/15/2023 17:01	SMPL
O3063-05	MEZEL1S	15439	2.539 -		SFAM01.1	PPB	6/15/2023 17:04	SMPL

LB125980

INSTRUMENT ID : CV1

O3063-06	MEZEL2	594	0.0462 -	SFAM01.1	PPB	6/15/2023 17:06 SMPL
O3063-07	MEZEL3	352	0.0056 -	SFAM01.1	PPB	6/15/2023 17:08 SMPL
O3063-08	MEZEL4	354	0.0059 -	SFAM01.1	PPB	6/15/2023 17:10 SMPL
O3063-09	MEZEL5	315	-0.0006 -	SFAM01.1	PPB	6/15/2023 17:13 SMPL
O3063-10	MEZEM9	290	-0.0048 -	SFAM01.1	PPB	6/15/2023 17:15 SMPL
O3061-08	GBYE4	26591	4.4117 -	SFAM01.1	PPB	6/15/2023 17:20 SMPL
CCV050	CCV050	29689	4.9319 -	SFAM01.1	PPB	6/15/2023 17:24 SMPL
CCB050	CCB050	83	-0.0396 -	SFAM01.1	PPB	6/15/2023 17:26 SMPL