



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125988

Analyst By : jignesh

Supervisor Review By : Sohil

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	06/15/2023	17:10
2	CAL2	1	Water	NA	NA	20.8	6.99	06/15/2023	17:11
3	CAL3	1	Water	NA	NA	20.7	10.02	06/15/2023	17:12
4	ICV	1	Water	NA	NA	20.8	7.01	06/15/2023	17:15
5	CCV1	1	Water	NA	NA	20.7	2.01	06/15/2023	17:20
6	O3241-02	1	Water	NA	NA	23.9	8.60	06/15/2023	17:22
7	O3276-01	1	Water	NA	NA	22.3	4.42	06/15/2023	17:25
8	O3278-02	1	Water	NA	NA	20.2	6.17	06/15/2023	17:30
9	O3278-02DUP	1	Water	NA	NA	20.2	6.18	06/15/2023	17:31
10	CCV2	1	Water	NA	NA	20.8	12.02	06/15/2023	17:33

VB 125988

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o3278 WorkList ID : 170933 Department : Wet-Chemistry Date : 06-15-2023 16:55:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3241-02	C 40508	Water	pH	Cool 4 deg C	PSEG03	G11	06/14/2023	9040C
O3276-01	A POOL-WATER	Water	pH	Cool 4 deg C	STER15	D32	06/12/2023	9040C
O3278-02	D RW5A-IDWGW-20230614	Water	pH	Cool 4 deg C	TETR06	G11	06/14/2023	9040C

Date/Time 06-15-23 17:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 06-15-23 18:00
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)