


LB125991

7471B

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 1.3944e-004

C= -1.4110e-002

Rho= 0.9998762

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	268	0.000	268				
0.2	0.200	0.194	-0.006	1491	0.0 %	1491				
2.5	2.500	2.417	-0.083	17436	0.0 %	17436				
5.0	5.000	5.058	0.058	36373	0.0 %	36373				
7.5	7.500	7.573	0.073	54410	0.0 %	54410				
10.0	10.000	9.935	-0.065	71355	0.0 %	71355				

LB125991 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Jun 2023 09:32:19

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	268	-	0.000	7471B	PPB	16 Jun 2023 09:41:45	S	Std
0.2 - 1	50.2	1491	-	0.200	7471B	PPB	16 Jun 2023 09:46:36	S	Std
2.5 - 1	52.5	17436	-	2.500	7471B	PPB	16 Jun 2023 09:48:53	S	Std
5.0 - 1	85	36373	-	5.000	7471B	PPB	16 Jun 2023 09:51:10	S	Std
7.5 - 1	53.5	54410	-	7.500	7471B	PPB	16 Jun 2023 09:53:28	S	Std
10.0 - 1	51.0	71355	-	10.000	7471B	PPB	16 Jun 2023 09:55:46	S	Std
ICV25 - 1	ICV25	26347	3.6597	-	7471B	PPB	16 Jun 2023 09:58:42	U	SMPL
ICB25 - 1	ICB25	-33	-0.0187	-	7471B	PPB	16 Jun 2023 10:00:57	U	SMPL
CCV50 - 1	CCV50	36102	5.0199	-	7471B	PPB	16 Jun 2023 10:03:15	U	SMPL
CCB50 - 1	CCB50	58	-0.0060	-	7471B	PPB	16 Jun 2023 10:05:30	U	SMPL
CRA - 1	CRA	1490	0.1937	-	7471B	PPB	16 Jun 2023 10:07:48	U	SMPL
HighStd - 1	HighStd	71160	9.9083	-	7471B	PPB	16 Jun 2023 10:10:03	U	SMPL
ChkStd - 1	ChkStd	51056	7.1050	-	7471B	PPB	16 Jun 2023 10:12:19	U	SMPL
PB153481BL - 1	PBS	349	0.0346	-	7471B	PPB	16 Jun 2023 10:14:38	U	SMPL
PB153481BS - 1	LCSS	28431	3.9502	-	7471B	PPB	16 Jun 2023 10:20:36	U	SMPL
O3222-01 - 1	COMP-SB-33	27460	3.8148	-	7471B	PPB	16 Jun 2023 10:22:52	U	SMPL
O3222-01DUP - 1	COMP-SB-33DUP	27383	3.8041	-	7471B	PPB	16 Jun 2023 10:25:11	U	SMPL
O3222-01MS - 1	COMP-SB-33MS	53279	7.4150	-	7471B	PPB	16 Jun 2023 10:27:31	U	SMPL
O3222-01MSD - 1	COMP-SB-33MSD	55707	7.7535	-	7471B	PPB	16 Jun 2023 10:29:50	U	SMPL
O3222-07 - 1	COMP-SB-26	80755	11.2462	-	7471B	PPB	16 Jun 2023 10:32:08	U	SMPL
CCV51 - 1	CCV51	35845	4.9840	-	7471B	PPB	16 Jun 2023 10:34:26	U	SMPL
CCB51 - 1	CCB51	5	-0.0134	-	7471B	PPB	16 Jun 2023 10:36:46	U	SMPL
O3222-13 - 1	COMPOSITE-SB-06	22366	3.1046	-	7471B	PPB	16 Jun 2023 10:39:03	U	SMPL
O3222-19 - 1	COMPOSITE-SB-16	25907	3.5983	-	7471B	PPB	16 Jun 2023 10:41:18	U	SMPL
O3223-01 - 1	COMPOSIT-SB-04	37707	5.2437	-	7471B	PPB	16 Jun 2023 10:43:36	U	SMPL
O3223-07 - 1	COMPOSIT-SB-05	32253	4.4832	-	7471B	PPB	16 Jun 2023 10:45:53	U	SMPL
O3223-13 - 1	COMPOSIT-SB-11	32233	4.4804	-	7471B	PPB	16 Jun 2023 10:48:12	U	SMPL
O3223-19 - 1	COMPOSIT-SB-03	24763	3.4388	-	7471B	PPB	16 Jun 2023 10:50:31	U	SMPL
O3224-01 - 1	COMPOSIT-SB-01/02	15248	2.1120	-	7471B	PPB	16 Jun 2023 10:52:50	U	SMPL
O3224-07 - 1	COMPOSIT-SB-10	31605	4.3928	-	7471B	PPB	16 Jun 2023 10:55:09	U	SMPL
O3224-12 - 1	COMPOSIT-SB-15	21608	2.9989	-	7471B	PPB	16 Jun 2023 10:57:27	U	SMPL
PB153482BL - 1	PBS	428	0.0456	-	7471B	PPB	16 Jun 2023 10:59:45	U	SMPL
CCV52 - 1	CCV52	36702	5.1035	-	7471B	PPB	16 Jun 2023 11:02:03	U	SMPL
CCB52 - 1	CCB52	123	0.0030	-	7471B	PPB	16 Jun 2023 11:04:18	U	SMPL
PB153482BS - 1	LCSS	24542	3.4080	-	7471B	PPB	16 Jun 2023 11:06:36	U	SMPL
O3229-01 - 1	OR-402-COMP-09	1499	0.1949	-	7471B	PPB	16 Jun 2023 11:08:52	U	SMPL
O3229-08 - 1	OR-402-COMP-10	772	0.0935	-	7471B	PPB	16 Jun 2023 11:11:10	U	SMPL
O3235-01 - 1	A-MANHOLE	1244	0.1594	-	7471B	PPB	16 Jun 2023 11:13:26	U	SMPL
O3235-01DUP - 1	A-MANHOLEDUP	1304	0.1677	-	7471B	PPB	16 Jun 2023 11:15:42	U	SMPL
O3235-01MS - 1	A-MANHOLEMS	27374	3.8029	-	7471B	PPB	16 Jun 2023 11:17:58	U	SMPL
O3235-01MSD - 1	A-MANHOLEMSD	29006	4.0304	-	7471B	PPB	16 Jun 2023 11:20:16	U	SMPL
O3235-03 - 1	B-PLEASANT	3093	0.4172	-	7471B	PPB	16 Jun 2023 11:22:35	U	SMPL
O3235-05 - 1	C-ANDREW	2217	0.2950	-	7471B	PPB	16 Jun 2023 11:24:53	U	SMPL
O3241-01 - 1	30852	21824	3.0290	-	7471B	PPB	16 Jun 2023 11:27:10	U	SMPL
CCV53 - 1	CCV53	36256	5.0413	-	7471B	PPB	16 Jun 2023 11:29:26	U	SMPL
CCB53 - 1	CCB53	203	0.0142	-	7471B	PPB	16 Jun 2023 11:31:44	U	SMPL
O3255-01 - 1	COMPOSITE-SB-24	20298	2.8162	-	7471B	PPB	16 Jun 2023 11:34:01	U	SMPL
O3255-07 - 1	COMPOSITE-SB-34	22023	3.0567	-	7471B	PPB	16 Jun 2023 11:36:17	U	SMPL
O3256-01 - 1	COMPOSITE-TP6-01-1-5	30505	4.2394	-	7471B	PPB	16 Jun 2023 11:38:34	U	SMPL
O3260-01 - 1	WC-1	5393	0.7379	-	7471B	PPB	16 Jun 2023 11:40:52	U	SMPL
O3222-01LX5 - 1		4418	0.6019	-	7471B	PPB	16 Jun 2023 11:43:10	U	SMPL
O3222-01A - 1		55723	7.7558	-	7471B	PPB	16 Jun 2023 11:45:28	U	SMPL
O3235-01LX5 - 1		2	-0.0138	-	7471B	PPB	16 Jun 2023 11:47:46	U	SMPL
O3235-01A - 1		29763	4.1360	-	7471B	PPB	16 Jun 2023 11:50:06	U	SMPL
O3222-07DLX2 - 1		39903	5.5499	-	7471B	PPB	16 Jun 2023 11:52:23	U	SMPL
CCV54 - 1	CCV54	36475	5.0719	-	7471B	PPB	16 Jun 2023 11:54:41	U	SMPL
CCB54 - 1	CCB54	161	0.0083	-	7471B	PPB	16 Jun 2023 11:56:59	U	SMPL