

LB125995


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50227



A= 0.0000e+000

B= 2.0057e-004

C= -6.9376e-002

Rho= 0.9999937

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% (1)
0.0	0.000	0.020	0.020	444	0.000	444	0				1
0.2	0.200	0.198	-0.002	1335	0.0 %	1335					1
2.5	2.500	2.475	-0.025	12687	0.0 %	12687					1
5.0	5.000	5.000	0.000	25276	0.0 %	25276					0
7.5	7.500	7.502	0.002	37748	0.0 %	37748					0
10.0	10.000	10.005	0.005	50227	0.0 %	50227					6

LB125995 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Jun 2023 12:14:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	444	-	0.0000	7470A	PPB	16 Jun 2023 12:15:59	S	Std
0.2 - 1	50.2	1335	-	0.2000	7470A	PPB	16 Jun 2023 12:18:17	S	Std
2.5 - 1	52.5	12687	-	2.5000	7470A	PPB	16 Jun 2023 12:20:34	S	Std
5.0 - 1	55	25276	-	5.0000	7470A	PPB	16 Jun 2023 12:22:52	S	Std
7.5 - 1	57.5	37748	-	7.5000	7470A	PPB	16 Jun 2023 12:27:48	S	Std
10.0 - 1	50	50227	-	10.0000	7470A	PPB	16 Jun 2023 12:35:46	S	Std
ICV26 - 1	ICV26	19622	3.8662	-	7470A	PPB	16 Jun 2023 12:38:35	U	SMPL
ICB26 - 1	ICB26	187	-0.0319	-	7470A	PPB	16 Jun 2023 12:40:50	U	SMPL
CCV55 - 1	CCV55	24522	4.8490	-	7470A	PPB	16 Jun 2023 12:43:07	U	SMPL
CCB55 - 1	CCB55	179	-0.0335	-	7470A	PPB	16 Jun 2023 12:45:22	U	SMPL
CRA - 1	CRA	1213	0.1739	-	7470A	PPB	16 Jun 2023 12:47:39	U	SMPL
HighStd - 1	HighStd	49398	9.8384	-	7470A	PPB	16 Jun 2023 12:49:55	U	SMPL
ChkStd - 1	ChkStd	34694	6.8892	-	7470A	PPB	16 Jun 2023 12:52:11	U	SMPL
PB153479BL - 1	PBW	115	-0.0463	-	7470A	PPB	16 Jun 2023 12:54:28	U	SMPL
PB153479BS - 1	LCSW	20268	3.9958	-	7470A	PPB	16 Jun 2023 12:56:47	U	SMPL
O3222-02 - 1	COMP-SB-33	423	0.0155	-	7470A	PPB	16 Jun 2023 12:59:04	U	SMPL
O3222-08 - 1	COMP-SB-26	532	0.0373	-	7470A	PPB	16 Jun 2023 13:01:23	U	SMPL
O3222-14 - 1	COMPOSITE-SB-06	556	0.0421	-	7470A	PPB	16 Jun 2023 13:03:40	U	SMPL
O3222-20 - 1	COMPOSITE-SB-16	494	0.0297	-	7470A	PPB	16 Jun 2023 13:05:57	U	SMPL
O3223-02 - 1	COMPOSIT-SB-04	548	0.0405	-	7470A	PPB	16 Jun 2023 13:08:12	U	SMPL
CCV56 - 1	CCV56	25402	5.0255	-	7470A	PPB	16 Jun 2023 13:10:28	U	SMPL
CCB56 - 1	CCB56	177	-0.0339	-	7470A	PPB	16 Jun 2023 13:12:43	U	SMPL
O3223-08 - 1	COMPOSIT-SB-05	469	0.0247	-	7470A	PPB	16 Jun 2023 13:15:00	U	SMPL
O3223-14 - 1	COMPOSIT-SB-11	555	0.0419	-	7470A	PPB	16 Jun 2023 13:17:15	U	SMPL
O3223-20 - 1	COMPOSIT-SB-03	575	0.0460	-	7470A	PPB	16 Jun 2023 13:19:31	U	SMPL
O3224-02 - 1	COMPOSIT-SB-01/02	679	0.0668	-	7470A	PPB	16 Jun 2023 13:21:47	U	SMPL
O3224-08 - 1	COMPOSIT-SB-10	389	0.0086	-	7470A	PPB	16 Jun 2023 13:24:03	U	SMPL
O3224-13 - 1	COMPOSIT-SB-15	312	-0.0068	-	7470A	PPB	16 Jun 2023 13:26:20	U	SMPL
O3228-01 - 1	SOM-COMP-1	289	-0.0114	-	7470A	PPB	16 Jun 2023 13:28:37	U	SMPL
O3228-02 - 1	SOM-COMP-2	376	0.0060	-	7470A	PPB	16 Jun 2023 13:30:54	U	SMPL
O3228-03 - 1	SOM-COMP-3	414	0.0137	-	7470A	PPB	16 Jun 2023 13:33:12	U	SMPL
O3229-02 - 1	OR-402-COMP-09	445	0.0199	-	7470A	PPB	16 Jun 2023 13:35:28	U	SMPL
CCV57 - 1	CCV57	26590	5.2638	-	7470A	PPB	16 Jun 2023 13:37:44	U	SMPL
CCB57 - 1	CCB57	-44	-0.0782	-	7470A	PPB	16 Jun 2023 13:39:59	U	SMPL
O3229-09 - 1	OR-402-COMP-10	748	0.0807	-	7470A	PPB	16 Jun 2023 13:42:16	U	SMPL
O3233-01 - 1	ELI-COMP-1	604	0.0518	-	7470A	PPB	16 Jun 2023 13:44:32	U	SMPL
O3233-02 - 1	ELI-COMP-2	778	0.0867	-	7470A	PPB	16 Jun 2023 13:46:48	U	SMPL
O3233-02DUP - 1	ELI-COMP-2DUP	716	0.0742	-	7470A	PPB	16 Jun 2023 13:49:04	U	SMPL
O3233-02MS - 1	ELI-COMP-2MS	21605	4.2639	-	7470A	PPB	16 Jun 2023 13:51:20	U	SMPL
O3233-02MSD - 1	ELI-COMP-2MSD	18436	3.6283	-	7470A	PPB	16 Jun 2023 13:53:37	U	SMPL
O3233-03 - 1	ELI-COMP-3	2779	0.4880	-	7470A	PPB	16 Jun 2023 13:55:55	U	SMPL
O3169-02 - 1	T-3089	11495	2.2362	-	7470A	PPB	16 Jun 2023 14:02:49	U	SMPL
CCV58 - 1	CCV58	25637	5.0726	-	7470A	PPB	16 Jun 2023 14:05:06	U	SMPL
CCB58 - 1	CCB58	157	-0.0379	-	7470A	PPB	16 Jun 2023 14:07:23	U	SMPL
O3233-04 - 1	ELI-0042	357	0.0022	-	7470A	PPB	16 Jun 2023 14:09:40	U	SMPL
O3233-04DUP - 1	ELI-0042DUP	-72	-0.0838	-	7470A	PPB	16 Jun 2023 14:11:56	U	SMPL
O3233-04MS - 1	ELI-0042MS	19469	3.8355	-	7470A	PPB	16 Jun 2023 14:14:12	U	SMPL
O3233-04MSD - 1	ELI-0042MSD	18466	3.6344	-	7470A	PPB	16 Jun 2023 14:16:28	U	SMPL
O3233-05 - 1	ELI-0047	1105	0.1523	-	7470A	PPB	16 Jun 2023 14:18:46	U	SMPL
PB153480BL - 1	PBW	293	-0.0106	-	7470A	PPB	16 Jun 2023 14:21:06	U	SMPL
PB153480BS - 1	LCSW	22212	4.3857	-	7470A	PPB	16 Jun 2023 14:23:23	U	SMPL
PB153433TB - 1	PB153433TB	342	-0.0008	-	7470A	PPB	16 Jun 2023 14:25:40	U	SMPL
PB153456TB - 1	PB153456TB	339	-0.0014	-	7470A	PPB	16 Jun 2023 14:27:57	U	SMPL
CCV59 - 1	CCV59	24060	4.7563	-	7470A	PPB	16 Jun 2023 14:30:14	U	SMPL
CCB59 - 1	CCB59	197	-0.0299	-	7470A	PPB	16 Jun 2023 14:32:31	U	SMPL
O3233-02LX5 - 1		167	-0.0359	-	7470A	PPB	16 Jun 2023 14:34:48	U	SMPL
O3233-02A - 1		20759	4.0943	-	7470A	PPB	16 Jun 2023 14:37:05	U	SMPL
O3233-04LX5 - 1		151	-0.0391	-	7470A	PPB	16 Jun 2023 14:39:22	U	SMPL
O3233-04A - 1		21495	4.2419	-	7470A	PPB	16 Jun 2023 14:41:39	U	SMPL
CCV60 - 1	CCV60	25835	5.1124	-	7470A	PPB	16 Jun 2023 14:43:56	U	SMPL
CCB60 - 1	CCB60	236	-0.0220	-	7470A	PPB	16 Jun 2023 14:46:13	U	SMPL