



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125997

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	06/16/2023	11:50
2	O3233-04		1	12	100.00	>212.0	>212.0	06/16/2023	12:20
3	O3233-04DUP		1	14	100.00	>212.0	>212.0	06/16/2023	12:50
4	O3233-05		1	13	100.00	>212.0	>212.0	06/16/2023	13:20
5	O3241-02		1	14	100.00	>212.0	>212.0	06/16/2023	13:50
6	O3251-01		1	14	100.00	>212.0	>212.0	06/16/2023	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66125997

WorkList Name : fp 6-15 WorkList ID : 170896 Department : Wet-Chemistry Date : 06-15-2023 08:29:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3251-01 J	MW-01	Water	Flash Point	Cool 4 deg C	LIRO01	G11	06/14/2023	1010B
O3233-04 B	ELI-0042	Water	Flash Point	Cool 4 deg C	PSEG03	G11	06/14/2023	1010B
O3233-05 B	ELI-0047	Water	Flash Point	Cool 4 deg C	PSEG03	G11	06/14/2023	1010B
O3241-02 F	40508	Water	Flash Point	Cool 4 deg C	PSEG03	G11	06/14/2023	1010B

Date/Time 06/16/2023 11:45
 Raw Sample Received by: RY GWC
 Raw Sample Relinquished by: AM CQ2C1

Date/Time 06/16/2023 14:30
 Raw Sample Received by: JM CQ2C1
 Raw Sample Relinquished by: RY GWC