



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126001

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/16/2023	14:00
2	CAL2	1	Water	NA	NA	20.7	7.01	06/16/2023	14:01
3	CAL3	1	Water	NA	NA	20.7	10.02	06/16/2023	14:02
4	ICV	1	Water	NA	NA	20.7	7.01	06/16/2023	14:05
5	CCV1	1	Water	NA	NA	20.6	2.01	06/16/2023	14:10
6	O3289-03	1	Solid	20.02	20	22.6	10.28	06/16/2023	14:15
7	O3289-06	1	Solid	20.03	20	22.6	9.01	06/16/2023	14:20
8	O3289-09	1	Solid	20.02	20	22.8	9.44	06/16/2023	14:22
9	O3289-11	1	Solid	20.01	20	22.8	9.17	06/16/2023	14:25
10	O3289-12	1	Solid	20.02	20	22.5	10.74	06/16/2023	14:30
11	O3289-14	1	Solid	20.03	20	22.8	8.85	06/16/2023	14:33
12	O3289-18	1	Solid	20.04	20	22.8	8.77	06/16/2023	14:36
13	O3289-21	1	Solid	20.02	20	22.9	8.62	06/16/2023	14:40
14	O3289-24	1	Solid	20.01	20	22.6	8.49	06/16/2023	14:45
15	O3296-02	1	Solid	20.02	20	22.2	6.26	06/16/2023	14:50
16	CCV2	1	Water	NA	NA	20.8	12.02	06/16/2023	15:00
17	O3296-08	1	Solid	20.03	20	22.0	7.85	06/16/2023	15:05
18	O3296-14	1	Solid	20.02	20	22.9	8.17	06/16/2023	15:10
19	O3296-20	1	Solid	20.03	20	22.6	8.11	06/16/2023	15:15
20	O3297-02	1	Solid	20.04	20	22.4	8.39	06/16/2023	15:20

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	03298-02	1	Solid	20.02	20	22.4	8.32	06/16/2023	15:22
22	03298-08	1	Solid	20.01	20	22.7	8.51	06/16/2023	15:25
23	03298-08DUP	1	Solid	20.02	20	22.8	8.52	06/16/2023	15:26
24	CCV3	1	Water	NA	NA	20.7	2.01	06/16/2023	15:33

WORKLIST(Hardcopy Internal Chain)

126001

WorkList Name : CORRRSIVITY O2804

WorkList ID : 170962

Department : Wet-Chemistry

Date : 06-16-2023 13:18:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3289-03	SB-01-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/12/2023	9045D
O3289-06	SB-02-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/13/2023	9045D
O3289-09	SB-03-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/12/2023	9045D
O3289-11	SB-04-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/13/2023	9045D
O3289-12	SB-05-0-20	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/12/2023	9045D
O3289-14	SB-05-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/13/2023	9045D
O3289-18	SB-06-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/13/2023	9045D
O3289-21	SB-07-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/12/2023	9045D
O3289-24	SB-08-COMP	Solid	Corrosivity	Cool 4 deg C	LIRO01	D31	06/12/2023	9045D
O3296-02	COMPOSIT-R-02	Solid	Corrosivity	Cool 4 deg C	SOUN03	D21	06/15/2023	9045D
O3296-08	COMPOSIT-R-01	Solid	Corrosivity	Cool 4 deg C	SOUN03	D21	06/15/2023	9045D
O3296-14	COMPOSIT-TW-01	Solid	Corrosivity	Cool 4 deg C	SOUN03	D21	06/15/2023	9045D
O3296-20	COMPOSIT-TW-07	Solid	Corrosivity	Cool 4 deg C	SOUN03	D21	06/15/2023	9045D
O3297-02	COMPOSIT-TP-6-6-9	Solid	Corrosivity	Cool 4 deg C	SOUN03	D12	06/15/2023	9045D
O3298-02	COMPOSIT-TW-04	Solid	Corrosivity	Cool 4 deg C	SOUN03	D11	06/15/2023	9045D
O3298-08	COMPOSIT-TW-05	Solid	Corrosivity	Cool 4 deg C	SOUN03	D11	06/15/2023	9045D

13:25

Date/Time 06-16-23

Raw Sample Received by: 10 (CWC)

Raw Sample Relinquished by: JT (CWC)

17:00

Date/Time 06-16-23

Raw Sample Received by: JT (CWC)

Raw Sample Relinquished by: 10 (CWC)