



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/19/2023

OVENTEMP IN Celsius(°C): 106
Time IN: 13:25
In Date: 06/17/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:15
Out Date: 06/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126005

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3192-01	EZCK1	1	1.17	8.21	9.38	8.57	90.1	
O3192-02	EZCK2	2	1.14	8.80	9.94	8.99	89.2	
O3192-03	EZCK3	3	1.15	8.73	9.88	8.72	86.7	
O3192-04	EZCK4	4	1.15	8.84	9.99	9.22	91.3	
O3192-05	EZCK5	5	1.15	8.59	9.74	8.94	90.7	
O3192-06	EZCK6	6	1.13	8.39	9.52	8.68	90.0	
O3192-07	EZCK7	7	1.15	8.13	9.28	8.4	89.2	
O3192-08	EZCK8	8	1.13	8.16	9.29	8.27	87.5	
O3192-09	EZCK9	9	1.14	8.67	9.81	8.43	84.1	
O3192-10	EZCL0	10	1.15	8.62	9.77	8.12	80.9	
O3192-11	EZCL1	11	1.15	8.41	9.56	8.62	88.8	
O3192-12	EZCL2	12	1.13	8.78	9.91	8.94	89.0	
O3192-13	EZCL3	13	1.14	8.62	9.76	8.71	87.8	
O3192-14	EZCL4	14	1.13	8.66	9.79	8.84	89.0	
O3192-15	EZCL5	15	1.14	8.85	9.99	8.75	86.0	
O3192-16	EZCL6	16	1.14	8.45	9.59	8.27	84.4	
O3192-17	EZCL7	17	1.13	8.84	9.97	8.66	85.2	
O3192-18	EZCL7D	18	1.13	8.84	9.97	8.66	85.2	
O3192-19	EZCL7S	19	1.13	8.84	9.97	8.66	85.2	
O3192-20	EZCL8	20	1.14	8.84	9.98	9.17	90.8	
O3192-21	EZCL9	21	1.13	8.68	9.81	8.93	89.9	
O3192-22	EZCM0	22	1.15	8.68	9.83	8.86	88.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$