



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/19/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 13:55
In Date: 06/17/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:22
Out Date: 06/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126006

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3193-01	EZCM1	1	1.14	8.53	9.67	9.00	92.1	
O3193-02	EZCM2	2	1.15	8.52	9.67	8.9	91.0	
O3193-03	EZCM3	3	1.13	8.86	9.99	9.11	90.1	
O3193-04	EZCM4	4	1.14	8.11	9.25	8.78	94.2	
O3193-05	EZCM5	5	1.13	8.79	9.92	9.41	94.2	
O3193-06	EZCM6	6	1.13	8.85	9.98	9.05	89.5	
O3193-07	EZCM7	7	1.13	8.55	9.68	8.95	91.5	
O3193-08	EZCM8	8	1.14	8.81	9.95	9.32	92.8	
O3193-09	EZCM9	9	1.15	8.24	9.39	8.42	88.2	
O3193-10	EZCN0	10	1.14	8.85	9.99	8.82	86.8	
O3193-11	EZCN1	11	1.13	8.43	9.56	8.69	89.7	
O3193-12	EZCN2	12	1.14	8.63	9.77	8.81	88.9	
O3193-13	EZCN3	13	1.13	8.86	9.99	9.04	89.3	
O3193-14	EZCN4	14	1.15	8.29	9.44	8.48	88.4	
O3193-15	EZCN5	15	1.14	8.60	9.74	8.55	86.2	
O3193-16	EZCN6	16	1.14	8.82	9.96	8.92	88.2	
O3193-17	EZCN7	17	1.15	8.74	9.89	9.06	90.5	
O3193-18	EZCN8	18	1.14	8.32	9.46	8.46	88.0	
O3193-19	EZCN9	19	1.14	8.15	9.29	8.34	88.3	
O3193-20	EZCP0	20	1.14	8.74	9.88	8.87	88.4	
O3193-21	EZCP0D	21	1.14	8.74	9.88	8.87	88.4	
O3193-22	EZCP0S	22	1.14	8.74	9.88	8.87	88.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$