



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/19/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:10
In Date: 06/17/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 06/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126010

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3197-01	EZC82	1	1.15	8.43	9.58	8.95	92.5	
O3197-02	EZC83	2	1.15	8.46	9.61	8.8	90.4	
O3197-03	EZC84	3	1.15	8.84	9.99	9.42	93.6	
O3197-04	EZC85	4	1.15	8.74	9.89	9.46	95.1	
O3197-05	EZC86	5	1.15	8.82	9.97	9.36	93.1	
O3197-06	EZC87	6	1.14	8.61	9.75	9.24	94.1	
O3197-07	EZC88	7	1.14	8.31	9.45	8.74	91.5	
O3197-08	EZC89	8	1.12	8.87	9.99	9.35	92.8	
O3197-09	EZC89D	9	1.12	8.87	9.99	9.35	92.8	
O3197-10	EZC89S	10	1.12	8.87	9.99	9.35	92.8	
O3197-11	EZC90	11	1.12	8.43	9.55	8.92	92.5	
O3197-12	EZBT0	12	1.12	8.73	9.85	9.00	90.3	
O3197-13	EZBT1	13	1.13	8.65	9.78	8.66	87.1	
O3197-14	EZBT2	14	1.13	8.64	9.77	8.38	83.9	
O3197-15	EZBT3	15	1.13	8.54	9.67	8.84	90.3	
O3197-16	EZBT4	16	1.13	8.72	9.85	8.63	86.0	
O3197-17	EZBT5	17	1.13	8.75	9.88	8.67	86.2	
O3197-18	EZBT6	18	1.15	8.84	9.99	9.26	91.7	
O3197-19	EZBT7	19	1.13	8.48	9.61	8.89	91.5	
O3197-20	EZBT8	20	1.13	8.78	9.91	9.36	93.7	
O3197-21	EZBT9	21	1.12	8.50	9.62	8.74	89.6	
O3197-22	EZBW0	22	1.13	8.86	9.99	9.06	89.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$