



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB126013

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/17/2023	09:30
2	CAL2	1	Water	NA	NA	20.7	7.00	06/17/2023	09:31
3	CAL3	1	Water	NA	NA	20.7	10.02	06/17/2023	09:32
4	ICV	1	Water	NA	NA	20.7	6.99	06/17/2023	09:35
5	CCV1	1	Water	NA	NA	20.7	2.01	06/17/2023	09:40
6	O3305-01	1	Water	NA	NA	20.9	7.29	06/17/2023	09:45
7	O3305-01DUP	1	Water	NA	NA	20.8	7.30	06/17/2023	09:46
8	O3311-01	1	Water	NA	NA	22.6	8.72	06/17/2023	09:50
9	O3315-04	1	Water	NA	NA	24.6	5.45	06/17/2023	09:55
10	O3315-05	1	Water	NA	NA	24.8	4.97	06/17/2023	10:00
11	CCV2	1	Water	NA	NA	20.7	12.02	06/17/2023	10:05

126013

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 03305 WorkList ID : 170989 Department : Wet-Chemistry Date : 06-17-2023 09:01:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3305-01	NEGATIVE-MH-BT-TRACK-89-	Water	pH	Cool 4 deg C	TULL02	G11	06/16/2023	9040C
O3311-01	1646	Water	pH	Cool 4 deg C	PSEG03	G11	06/16/2023	9040C
O3315-04	34344	Water	pH	Cool 4 deg C	PSEG03	G11	06/16/2023	9040C
O3315-05	60104	Water	pH	Cool 4 deg C	PSEG03	G11	06/16/2023	9040C

06.17.23 09:20

Date/Time
Raw Sample Received by: SP/CD
Raw Sample Relinquished by: SP/CD

06.17.23

Date/Time
Raw Sample Received by: SP/CD
Raw Sample Relinquished by: SP/CD