



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126014

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/17/2023	10:00
2	CAL2	1	Water	NA	NA	20.7	7.00	06/17/2023	10:01
3	CAL3	1	Water	NA	NA	20.6	10.02	06/17/2023	10:04
4	ICV	1	Water	NA	NA	20.7	6.99	06/17/2023	10:05
5	CCV1	1	Water	NA	NA	20.7	2.02	06/17/2023	10:10
6	O3308-02	1	Solid	20.02	20	23.6	8.47	06/17/2023	10:15
7	O3308-05	1	Solid	20.03	20	24.0	8.35	06/17/2023	10:20
8	O3320-02	1	Solid	20.02	20	22.3	6.23	06/17/2023	10:25
9	O3320-05	1	Solid	20.01	20	22.4	6.36	06/17/2023	10:26
10	O3320-08	1	Solid	20.02	20	22.3	7.20	06/17/2023	10:30
11	O3320-11	1	Solid	20.03	20	22.4	7.25	06/17/2023	10:33
12	O3320-14	1	Solid	20.04	20	22.2	7.40	06/17/2023	10:35
13	O3320-17	1	Solid	20.02	20	22.2	7.30	06/17/2023	10:40
14	O3320-20	1	Solid	20.03	20	22.0	7.54	06/17/2023	10:45
15	O3321-02	1	Solid	20.02	20	21.9	7.64	06/17/2023	10:47
16	CCV2	1	Water	NA	NA	20.6	12.02	06/17/2023	10:50
17	O3321-05	1	Solid	20.02	20	21.8	6.49	06/17/2023	11:00
18	O3321-08	1	Solid	20.03	20	21.3	7.72	06/17/2023	11:06
19	O3321-11	1	Solid	20.04	20	21.7	7.73	06/17/2023	11:10
20	O3321-14	1	Solid	20.02	20	21.8	7.48	06/17/2023	11:15

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O3321-17	1	Solid	20.03	20	21.7	8.17	06/17/2023	11:20
22	O3321-17DUP	1	Solid	20.04	20	21.8	8.18	06/17/2023	11:21
23	CCV3	1	Water	NA	NA	20.7	2.01	06/17/2023	11:25

126014

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORRISIVITY O3221 WorkList ID : 170987 Department : Wet-Chemistry Date : 06-17-2023 08:59:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3308-02	SP-LEFT	Solid	Corrosivity	Cool 4 deg C	TULL02	G11	06/16/2023	9045D
O3308-05	SP-RIGHT	Solid	Corrosivity	Cool 4 deg C	TULL02	G11	06/16/2023	9045D
O3320-02	WC-19	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/16/2023	9045D
O3320-05	WC-10	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/16/2023	9045D
O3320-08	WC-12	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/16/2023	9045D
O3320-11	WC-13	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/16/2023	9045D
O3320-14	WC-14	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/16/2023	9045D
O3320-17	WC-16	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/16/2023	9045D
O3320-20	WC-14A	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/16/2023	9045D
O3321-02	WC-25	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/15/2023	9045D
O3321-05	WC-24	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/15/2023	9045D
O3321-08	WC-23	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/15/2023	9045D
O3321-11	WC-22	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/15/2023	9045D
O3321-14	WC-21	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/15/2023	9045D
O3321-17	WC-20	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/15/2023	9045D

Date/Time 06-17-23 09:30
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: AL SM

Date/Time 06-17-23 16:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)