



## Extraction and Analytical Summary Report

Analysis Method: 1664A  
Test: TPH  
Run Number: LB126015  
Analysis Date: 06/17/2023  
BalanceID: WC SC-6  
OvenID: EXT OVEN-3

ANALYST: jignesh  
REVIEWED BY: Iwona  
Extraction Date: 06/17/2023  
Extraction IN Time: 12:45  
Extraction OUT Time: 13:30  
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB126015BL	LB126015BL	WATER	1.3	1000	100	3.0524	3.0524	3.02	3.0525	3.0525	0.0001	0.1
2	LB126015BS	LB126015BS	WATER	1.3	1000	100	3.1487	3.1487	3.01	3.1655	3.1655	0.0168	16.8
3	LB126015BSD	LB126015BSD	WATER	1.3	1000	100	3.0605	3.0605	3.03	3.0772	3.0772	0.0167	16.7
4	O3226-01	GRAB	WATER	1.6	1000	100	3.0997	3.0997	3.05	3.1246	3.1246	0.0249	24.9

QC Batch# LB126015Test: TPHAnalysis Date: 06/17/2023**Chemicals Used:**

Chemical Name	Chemical Lot #
HEXANE	W3034
pH Paper 0-14	M4909
Sodium Sulfate	EP2347
1:1 HCL	WP101334
Silica Gel	W2839
Sand	NA

**Standards Used:**

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP101332
LCSWD	5.00 ML	WP101333
MS/MSD	NA	NA

**BALANCE CALIBRATION / OVEN Dessicator Data****Analytical Balance ID # : WC SC-6****Before Analysis**

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 14:46  
1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 14:00  
Bal Check Time: 13:10 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 15:25  
Out Time1: 14:45

**After Analysis**

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 16:31  
1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 16:00  
Bal Check Time: 17:15 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 17:05  
Out Time2: 16:30

WORKLIST(Hardcopy Internal Chain)

MM 126015

WorkList Name : TPH 03226

WorkList ID : 171000

Department : Wet-Chemistry

Date : 06-17-2023 12:31:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3226-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	D22	06/14/2023	1664A

121.25

Date/Time 06.17.23

Raw Sample Received by: CP SM

Raw Sample Relinquished by: CP SM

Date/Time 06.17.23

Raw Sample Received by: CP SM

Raw Sample Relinquished by: 181000