



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/19/2023

OVENTEMP IN Celsius(°C): 106
Time IN: 13:40
In Date: 06/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 06/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126022

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3205-01	EZCB0	1	1.15	8.62	9.77	8.68	87.4	
O3205-02	EZCB1	2	1.14	8.64	9.78	8.43	84.4	
O3205-03	EZCB2	3	1.14	8.85	9.99	8.6	84.3	
O3205-04	EZCB4	4	1.14	8.47	9.61	8.99	92.7	
O3205-05	EZCB5	5	1.14	8.60	9.74	9.02	91.6	
O3205-06	EZCB6	6	1.14	8.41	9.55	8.62	88.9	
O3205-07	EZCB7	7	1.14	8.80	9.94	8.5	83.6	
O3205-08	EZCB8	8	1.15	8.70	9.85	8.38	83.1	
O3205-09	EZCB9	9	1.14	8.28	9.42	7.87	81.3	
O3205-10	EZCC0	10	1.15	8.48	9.63	8.89	91.3	
O3205-11	EZCC1	11	1.14	8.73	9.87	8.9	88.9	
O3205-12	EZCC2	12	1.15	8.61	9.76	8.32	83.3	
O3205-13	EZCC3	13	1.15	8.59	9.74	8.96	90.9	
O3205-14	EZCC4	14	1.14	8.59	9.73	8.79	89.1	
O3205-15	EZCC5	15	1.15	8.41	9.56	8.58	88.3	
O3205-16	EZCC6	16	1.14	8.43	9.57	8.32	85.2	
O3205-17	EZCC7	17	1.13	8.54	9.67	8.68	88.4	
O3205-18	EZCC8	18	1.15	8.63	9.78	8.57	86.0	
O3205-19	EZCC9	19	1.14	8.85	9.99	8.48	82.9	
O3205-20	EZCD6	20	1.13	8.68	9.81	8.61	86.2	
O3205-21	EZCD6D	21	1.13	8.68	9.81	8.61	86.2	
O3205-22	EZCD6S	22	1.13	8.68	9.81	8.61	86.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$