


LB126032

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9159e-004

C= 3.3430e-002

Rho= 0.9998830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.076	0.076	222	0.000	222	0				
0.2	0.200	0.196	-0.004	848	0.0 %	848					-2
2.5	2.500	2.395	-0.105	12325	0.0 %	12325					-4
5.0	5.000	4.994	-0.006	25894	0.0 %	25894					0
7.5	7.500	7.547	0.047	39215	0.0 %	39215					0
10.0	10.000	9.992	-0.008	51980	0.0 %	51980					0

LB126032 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jun 2023 07:57:38

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	222	-	0.0000	7470A	PPB	19 Jun 2023 08:05:34	S	Std
0.2 - 1	50.2	848	-	0.2000	7470A	PPB	19 Jun 2023 08:12:17	S	Std
2.5 - 1	52.5	12325	-	2.5000	7470A	PPB	19 Jun 2023 08:16:36	S	Std
5.0 - 1	55	25894	-	5.0000	7470A	PPB	19 Jun 2023 08:21:28	S	Std
7.5 - 1	57.5	39215	-	7.5000	7470A	PPB	19 Jun 2023 08:23:45	S	Std
10.0 - 1	510	51980	-	10.0000	7470A	PPB	19 Jun 2023 08:28:41	S	Std
ICV27 - 1	ICV27	19467	3.7631	-	7470A	PPB	19 Jun 2023 08:34:02	U	SMPL
ICB27 - 1	ICB27	130	0.0583	-	7470A	PPB	19 Jun 2023 08:38:50	U	SMPL
CCV61 - 1	CCV61	28462	5.4865	-	7470A	PPB	19 Jun 2023 08:41:06	U	SMPL
CCB61 - 1	CCB61	-87	0.0168	-	7470A	PPB	19 Jun 2023 08:43:21	U	SMPL
CRA - 1	CRA	800	0.1867	-	7470A	PPB	19 Jun 2023 08:45:39	U	SMPL
HighStd - 1	HighStd	51222	9.8470	-	7470A	PPB	19 Jun 2023 08:47:54	U	SMPL
ChkStd - 1	ChkStd	35417	6.8190	-	7470A	PPB	19 Jun 2023 08:50:10	U	SMPL
PB153511BL - 1	PBW	47	0.0424	-	7470A	PPB	19 Jun 2023 08:54:45	U	SMPL
PB153511BS - 1	LCSW	18886	3.6518	-	7470A	PPB	19 Jun 2023 08:57:01	U	SMPL
O3255-02 - 1	COMPOSIT-SB-24	733	0.1739	-	7470A	PPB	19 Jun 2023 09:01:51	U	SMPL
O3255-02DUP - 1	COMPOSIT-SB-24DUP	558	0.1403	-	7470A	PPB	19 Jun 2023 09:04:07	U	SMPL
O3255-02MS - 1	COMPOSIT-SB-24MS	18008	3.4836	-	7470A	PPB	19 Jun 2023 09:06:24	U	SMPL
O3255-02MSD - 1	COMPOSIT-SB-24MSD	19312	3.7334	-	7470A	PPB	19 Jun 2023 09:08:42	U	SMPL
O3255-08 - 1	COMPOSIT-SB-34	802	0.1871	-	7470A	PPB	19 Jun 2023 09:13:41	U	SMPL
CCV62 - 1	CCV62	26939	5.1947	-	7470A	PPB	19 Jun 2023 09:15:56	U	SMPL
CCB62 - 1	CCB62	-64	0.0212	-	7470A	PPB	19 Jun 2023 09:18:11	U	SMPL
O3256-02 - 1	COMPOSIT-TP6-01-1-5	868	0.1997	-	7470A	PPB	19 Jun 2023 09:20:29	U	SMPL
O3262-02 - 1	WS-SOIL1-0615	1072	0.2388	-	7470A	PPB	19 Jun 2023 09:22:44	U	SMPL
O3262-05 - 1	WS-SOIL2-0615	1935	0.4042	-	7470A	PPB	19 Jun 2023 09:25:00	U	SMPL
O3262-08 - 1	WS-SOIL3-0615	604	0.1492	-	7470A	PPB	19 Jun 2023 09:27:16	U	SMPL
O3282-02 - 1	WC-NEX-001	786	0.1840	-	7470A	PPB	19 Jun 2023 09:29:32	U	SMPL
O3282-07 - 1	WC-NEX-002	1284	0.2794	-	7470A	PPB	19 Jun 2023 09:31:49	U	SMPL
PB153512BL - 1	PBW	16	0.0365	-	7470A	PPB	19 Jun 2023 09:34:05	U	SMPL
PB153512BS - 1	LCSW	19189	3.7098	-	7470A	PPB	19 Jun 2023 09:36:23	U	SMPL
O3073-02 - 1	COMP	1179	0.2593	-	7470A	PPB	19 Jun 2023 09:38:40	U	SMPL
O3153-04 - 1	NO3-TRANS-WATER	427	0.1152	-	7470A	PPB	19 Jun 2023 09:40:57	U	SMPL
CCV63 - 1	CCV63	27450	5.2926	-	7470A	PPB	19 Jun 2023 09:43:15	U	SMPL
CCB63 - 1	CCB63	-131	0.0083	-	7470A	PPB	19 Jun 2023 09:45:30	U	SMPL
O3251-02 - 1	MW-01	2750	0.5603	-	7470A	PPB	19 Jun 2023 09:52:51	U	SMPL
O3251-02DUP - 1	MW-01DUP	2745	0.5593	-	7470A	PPB	19 Jun 2023 09:55:07	U	SMPL
O3251-02MS - 1	MW-01MS	20181	3.8999	-	7470A	PPB	19 Jun 2023 09:57:23	U	SMPL
O3251-02MSD - 1	MW-01MSD	21925	4.2340	-	7470A	PPB	19 Jun 2023 09:59:39	U	SMPL
O3251-03 - 1	MW-01	696	0.1668	-	7470A	PPB	19 Jun 2023 10:01:56	U	SMPL
O3251-04 - 1	MW-01-DUP	1330	0.2882	-	7470A	PPB	19 Jun 2023 10:04:15	U	SMPL
O3251-05 - 1	MW-01-DUP	603	0.1490	-	7470A	PPB	19 Jun 2023 10:06:31	U	SMPL
O3251-06 - 1	MW-02	1524	0.3254	-	7470A	PPB	19 Jun 2023 10:08:49	U	SMPL
O3251-07 - 1	MW-02	934	0.2124	-	7470A	PPB	19 Jun 2023 10:11:06	U	SMPL
O3251-08 - 1	TWP-05	3758	0.7534	-	7470A	PPB	19 Jun 2023 10:13:22	U	SMPL
CCV64 - 1	CCV64	27170	5.2389	-	7470A	PPB	19 Jun 2023 10:15:38	U	SMPL
CCB64 - 1	CCB64	-136	0.0074	-	7470A	PPB	19 Jun 2023 10:17:55	U	SMPL
O3251-09 - 1	TWP-05	1401	0.3018	-	7470A	PPB	19 Jun 2023 10:20:13	U	SMPL
O3251-10 - 1	TWP-01	2650	0.5411	-	7470A	PPB	19 Jun 2023 10:22:29	U	SMPL
O3251-11 - 1	TWP-01	1476	0.3162	-	7470A	PPB	19 Jun 2023 10:24:45	U	SMPL
O3274-01 - 1	TW-1	1632	0.3461	-	7470A	PPB	19 Jun 2023 10:27:01	U	SMPL
O3278-02 - 1	RW5A-IDWGW-20230614	933	0.2122	-	7470A	PPB	19 Jun 2023 10:29:17	U	SMPL
PB153476TB - 1	PB153476TB	227	0.0769	-	7470A	PPB	19 Jun 2023 10:38:07	U	SMPL
O3255-02LX5 - 1		-11	0.0313	-	7470A	PPB	19 Jun 2023 10:40:24	U	SMPL
O3255-02A - 1		18236	3.5273	-	7470A	PPB	19 Jun 2023 10:42:41	U	SMPL
O3251-02LX5 - 1		430	0.1158	-	7470A	PPB	19 Jun 2023 10:44:58	U	SMPL
O3251-02A - 1		23239	4.4858	-	7470A	PPB	19 Jun 2023 10:47:14	U	SMPL
CCV65 - 1	CCV65	25735	4.9640	-	7470A	PPB	19 Jun 2023 10:49:30	U	SMPL
CCB65 - 1	CCB65	-109	0.0125	-	7470A	PPB	19 Jun 2023 10:51:47	U	SMPL