



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/20/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 06/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:33
Out Date: 06/20/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126037

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3084-01	DCFL8	1	1.14	8.45	9.59	8.67	89.1	
O3084-02	DCFL9	2	1.11	6.65	7.76	7.14	90.7	
O3084-03	DCFM0	3	1.13	8.72	9.85	8.95	89.7	
O3084-04	DCFM1	4	1.13	8.50	9.63	8.7	89.1	
O3084-05	DCFM2	5	1.11	8.53	9.64	9.02	92.7	
O3084-06	DCFM3	6	1.13	8.00	9.13	8.9	97.1	
O3084-07	DCFM4	7	1.11	8.74	9.85	9.16	92.1	
O3084-08	DCFM5	8	1.1	8.85	9.95	8.7	85.9	
O3084-09	DCFM6	9	1.08	7.99	9.07	7.74	83.4	
O3084-10	DCFM7	10	1.13	8.53	9.66	8.72	89.0	
O3084-11	DCFM8	11	1.13	8.85	9.98	8.8	86.7	
O3084-12	DCFM9	12	1.12	8.60	9.72	9.06	92.3	
O3084-13	DCFN0	13	1.1	8.50	9.6	8.8	90.6	
O3084-14	DCFN1	14	1.13	8.61	9.74	8.82	89.3	
O3084-15	DCFN2	15	1.1	8.42	9.52	8.42	86.9	
O3084-16	DCFN3	16	1.12	8.79	9.91	8.97	89.3	
O3084-17	DCFN4	17	1.12	8.45	9.57	9.00	93.3	
O3084-18	DCFN5	18	1.11	8.74	9.85	8.88	88.9	
O3084-19	DCFN6	19	1.11	8.87	9.98	8.97	88.6	
O3084-20	DCFQ7	20	1.11	4.14	5.25	4.85	90.3	
O3084-21	DCFQ7MS	21	1.11	4.14	5.25	4.85	90.3	
O3084-22	DCFQ7MSD	22	1.11	4.14	5.25	4.85	90.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$