

LB126047


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

62972



A= 0.0000e+000

B= 1.6182e-004

C= -5.7391e-002

Rho= 0.9993811

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.082	0.082	861	0.000	861					
0.2	0.200	0.204	0.004	1613	0.0 %	1613					2
2.5	2.500	2.588	0.088	16346	0.0 %	16346					4
5.0	5.000	4.741	-0.259	29655	0.0 %	29655					5
7.5	7.500	7.453	-0.047	46414	0.0 %	46414					1
10.0	10.000	10.132	0.132	62972	0.0 %	62972					1

LB126047 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jun 2023 13:32:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	861	-	0.0000	7471B	PPB	19 Jun 2023 13:37:48	S	Std
0.2 - 1	SD	1613	-	0.2000	7471B	PPB	19 Jun 2023 13:40:05	S	Std
2.5 - 1	SD	16346	-	2.5000	7471B	PPB	19 Jun 2023 13:42:23	S	Std
5.0 - 1	SD	29655	-	5.0000	7471B	PPB	19 Jun 2023 13:44:40	S	Std
7.5 - 1	SD	46414	-	7.5000	7471B	PPB	19 Jun 2023 13:46:59	S	Std
10.0 - 1	SD	62972	-	10.0000	7471B	PPB	19 Jun 2023 13:51:55	S	Std
ICV28 - 1	ICV28	26890	4.2938	-	7471B	PPB	19 Jun 2023 13:54:53	U	SMPL
ICB28 - 1	ICB28	30	-0.0525	-	7471B	PPB	19 Jun 2023 13:59:47	U	SMPL
CCV66 - 1	CCV66	33589	5.3778	-	7471B	PPB	19 Jun 2023 14:02:02	U	SMPL
CCB66 - 1	CCB66	-112	-0.0755	-	7471B	PPB	19 Jun 2023 14:04:17	U	SMPL
CRA - 1	CRA	1544	0.1925	-	7471B	PPB	19 Jun 2023 14:06:34	U	SMPL
HighStd - 1	HighStd	62387	10.0378	-	7471B	PPB	19 Jun 2023 14:08:50	U	SMPL
ChkStd - 1	ChkStd	42438	6.8097	-	7471B	PPB	19 Jun 2023 14:11:05	U	SMPL
PB153522BL - 1	PBS	555	0.0324	-	7471B	PPB	19 Jun 2023 14:13:23	U	SMPL
PB153522BS - 1	LCSS	23589	3.7597	-	7471B	PPB	19 Jun 2023 14:15:42	U	SMPL
O3262-01 - 1	WS-SOIL1-0615	2447	0.3386	-	7471B	PPB	19 Jun 2023 14:17:59	U	SMPL
O3262-04 - 1	WS-SOIL2-0615	4788	0.7174	-	7471B	PPB	19 Jun 2023 14:20:18	U	SMPL
O3262-04DUP - 1	WS-SOIL2-0615DUP	4253	0.6308	-	7471B	PPB	19 Jun 2023 14:22:35	U	SMPL
O3262-04MS - 1	WS-SOIL2-0615MS	32819	5.2532	-	7471B	PPB	19 Jun 2023 14:24:52	U	SMPL
O3262-04MSD - 1	WS-SOIL2-0615MSD	32677	5.2303	-	7471B	PPB	19 Jun 2023 14:27:08	U	SMPL
CCV67 - 1	CCV67	30557	4.8872	-	7471B	PPB	19 Jun 2023 14:29:25	U	SMPL
CCB67 - 1	CCB67	341	-0.0022	-	7471B	PPB	19 Jun 2023 14:34:20	U	SMPL
O3262-07 - 1	WS-SOIL3-0615	5370	0.8116	-	7471B	PPB	19 Jun 2023 14:36:35	U	SMPL
O3279-01 - 1	TOPSOIL-OU3-DENALI-1	4482	0.6679	-	7471B	PPB	19 Jun 2023 14:38:51	U	SMPL
O3282-01 - 1	WC-NEX-001	211250	34.1261	-	7471B	PPB	19 Jun 2023 14:41:06	U	SMPL
O3282-06 - 1	WC-NEX-002	143457	23.1562	-	7471B	PPB	19 Jun 2023 14:43:22	U	SMPL
O3287-01 - 1	OK-1-061523	8308	1.2870	-	7471B	PPB	19 Jun 2023 14:45:51	U	SMPL
O3288-01 - 1	PAT-2-061523	9125	1.4192	-	7471B	PPB	19 Jun 2023 14:48:18	U	SMPL
O3295-01 - 1	WC-1	1939	0.2564	-	7471B	PPB	19 Jun 2023 14:50:35	U	SMPL
O3296-01 - 1	COMPOSIT-R-02	10851	1.6985	-	7471B	PPB	19 Jun 2023 14:52:52	U	SMPL
O3296-07 - 1	COMPOSIT-R-01	16807	2.6622	-	7471B	PPB	19 Jun 2023 14:55:09	U	SMPL
O3296-13 - 1	COMPOSIT-TWO-01	23617	3.7642	-	7471B	PPB	19 Jun 2023 14:57:25	U	SMPL
CCV68 - 1	CCV68	28245	4.5131	-	7471B	PPB	19 Jun 2023 14:59:42	U	SMPL
CCB68 - 1	CCB68	423	0.0111	-	7471B	PPB	19 Jun 2023 15:04:43	U	SMPL
O3296-19 - 1	COMPOSIT-TW-07	15550	2.4588	-	7471B	PPB	19 Jun 2023 15:06:58	U	SMPL
O3297-01 - 1	COMPOSIT-TP-6-6-9	33218	5.3178	-	7471B	PPB	19 Jun 2023 15:09:14	U	SMPL
O3298-01 - 1	COMPOSIT-TW-04	22193	3.5338	-	7471B	PPB	19 Jun 2023 15:11:31	U	SMPL
O3298-07 - 1	COMPOSIT-TW-05	24163	3.8526	-	7471B	PPB	19 Jun 2023 15:13:50	U	SMPL
O3302-01 - 1	SB-1	1614	0.2038	-	7471B	PPB	19 Jun 2023 15:16:07	U	SMPL
O3302-02 - 1	SB-2	1407	0.1703	-	7471B	PPB	19 Jun 2023 15:18:25	U	SMPL
O3302-03 - 1	SB-4	1581	0.1984	-	7471B	PPB	19 Jun 2023 15:20:42	U	SMPL
O3302-04 - 1	SB-3	2665	0.3738	-	7471B	PPB	19 Jun 2023 15:23:00	U	SMPL
PB153523BL - 1	PBS	811	0.0738	-	7471B	PPB	19 Jun 2023 15:25:19	U	SMPL
PB153523BS - 1	LCSS	23536	3.7511	-	7471B	PPB	19 Jun 2023 15:27:35	U	SMPL
CCV69 - 1	CCV69	33070	5.2938	-	7471B	PPB	19 Jun 2023 15:29:51	U	SMPL
CCB69 - 1	CCB69	653	0.0483	-	7471B	PPB	19 Jun 2023 15:32:09	U	SMPL
O3289-01 - 1	SB-01-0-2.0	11035	1.7282	-	7471B	PPB	19 Jun 2023 15:34:26	U	SMPL
O3289-01DUP - 1	SB-01-0-2.0DUP	12750	2.0058	-	7471B	PPB	19 Jun 2023 15:36:42	U	SMPL
O3289-01MS - 1	SB-01-0-2.0MS	47025	7.5520	-	7471B	PPB	19 Jun 2023 15:38:58	U	SMPL
O3289-01MSD - 1	SB-01-0-2.0MSD	40721	6.5319	-	7471B	PPB	19 Jun 2023 15:41:14	U	SMPL
O3289-02 - 1	SB-01-4.0-6.0	11702	1.8362	-	7471B	PPB	19 Jun 2023 15:43:32	U	SMPL
O3289-04 - 1	SB-02-0-2.0	9435	1.4693	-	7471B	PPB	19 Jun 2023 15:45:51	U	SMPL
O3289-05 - 1	SB-02-12.0-14.0	692	0.0546	-	7471B	PPB	19 Jun 2023 15:48:09	U	SMPL
O3289-07 - 1	SB-03-0-2.0	12970	2.0414	-	7471B	PPB	19 Jun 2023 15:50:26	U	SMPL
O3289-08 - 1	SB-03-4.0-6.0	9811	1.5302	-	7471B	PPB	19 Jun 2023 15:52:43	U	SMPL
O3289-10 - 1	SB-04-12.0-14.0	1629	0.2062	-	7471B	PPB	19 Jun 2023 15:54:59	U	SMPL
CCV70 - 1	CCV70	29427	4.7044	-	7471B	PPB	19 Jun 2023 15:57:15	U	SMPL
CCB70 - 1	CCB70	312	-0.0069	-	7471B	PPB	19 Jun 2023 15:59:31	U	SMPL
O3289-13 - 1	SB-05-12.0-14.0	6490	0.9928	-	7471B	PPB	19 Jun 2023 16:01:48	U	SMPL
O3289-15 - 1	SB-06-0-2.0	18689	2.9668	-	7471B	PPB	19 Jun 2023 16:04:04	U	SMPL
O3289-16 - 1	SB-06-12.0-14.0	764	0.0662	-	7471B	PPB	19 Jun 2023 16:06:20	U	SMPL
O3289-17 - 1	SB-06-12.0-14.0-DUP	873	0.0839	-	7471B	PPB	19 Jun 2023 16:08:37	U	SMPL
O3289-19 - 1	SB-07-0-2.0	5949	0.9052	-	7471B	PPB	19 Jun 2023 16:10:53	U	SMPL
O3289-20 - 1	SB-07-4.0-6.0	2473	0.3428	-	7471B	PPB	19 Jun 2023 16:13:10	U	SMPL
O3289-22 - 1	SB-08-0-2.0	11806	1.8530	-	7471B	PPB	19 Jun 2023 16:15:26	U	SMPL
O3289-23 - 1	SB-08-4.0-6.0	590	0.0381	-	7471B	PPB	19 Jun 2023 16:17:43	U	SMPL
O3308-01 - 1	SP-LEFT	8679	1.3470	-	7471B	PPB	19 Jun 2023 16:20:00	U	SMPL
O3308-04 - 1	SP-RIGHT	9935	1.5502	-	7471B	PPB	19 Jun 2023 16:22:17	U	SMPL
CCV71 - 1	CCV71	29546	4.7236	-	7471B	PPB	19 Jun 2023 16:24:37	U	SMPL
CCB71 - 1	CCB71	282	-0.0118	-	7471B	PPB	19 Jun 2023 16:26:53	U	SMPL
O3289-01LX5 - 1		895	0.0874	-	7471B	PPB	19 Jun 2023 16:29:11	U	SMPL
O3289-01A - 1		25366	4.0472	-	7471B	PPB	19 Jun 2023 16:31:28	U	SMPL
O3262-04LX5 - 1		1433	0.1745	-	7471B	PPB	19 Jun 2023 16:33:44	U	SMPL
O3262-04A - 1		38803	6.2215	-	7471B	PPB	19 Jun 2023 16:36:03	U	SMPL
O3282-01DLX10 - 1		17793	2.8218	-	7471B	PPB	19 Jun 2023 16:38:20	U	SMPL
O3282-06DLX5 - 1		30774	4.9223	-	7471B	PPB	19 Jun 2023 16:40:39	U	SMPL
CCV72 - 1	CCV72	29423	4.7037	-	7471B	PPB	19 Jun 2023 16:42:58	U	SMPL
CCB72 - 1	CCB72	22	-0.0538	-	7471B	PPB	19 Jun 2023 16:45:17	U	SMPL