



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126060

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/19/2023	12:37
2	CAL2	1	Water	NA	NA	20.7	7.00	06/19/2023	12:38
3	CAL3	1	Water	NA	NA	20.7	10.02	06/19/2023	12:40
4	ICV	1	Water	NA	NA	20.8	7.01	06/19/2023	12:44
5	CCV1	1	Water	NA	NA	20.7	2.01	06/19/2023	12:45
6	O3327-02	1	Solid	20.02	20	23.8	8.58	06/19/2023	12:50
7	O3327-02DUP	1	Solid	20.01	20	23.9	8.59	06/19/2023	12:51
8	O3327-08	1	Solid	20.03	20	23.3	8.42	06/19/2023	12:55
9	O3327-14	1	Solid	20.04	20	23.7	8.18	06/19/2023	13:00
10	O3327-20	1	Solid	20.02	20	23.8	8.17	06/19/2023	13:02
11	O3328-02	1	Solid	20.03	20	23.7	8.73	06/19/2023	13:05
12	O3328-08	1	Solid	20.01	20	23.6	8.39	06/19/2023	13:10
13	O3328-14	1	Solid	20.02	20	23.7	8.52	06/19/2023	13:11
14	O3328-20	1	Solid	20.03	20	23.6	8.26	06/19/2023	13:15
15	O3329-02	1	Solid	20.02	20	23.8	8.35	06/19/2023	13:20
16	CCV2	1	Water	NA	NA	20.7	12.02	06/19/2023	13:22
17	O3329-08	1	Solid	20.01	20	23.9	8.42	06/19/2023	13:25
18	O3329-14	1	Solid	20.04	20	23.8	8.47	06/19/2023	13:30
19	O3329-20	1	Solid	20.02	20	23.8	8.65	06/19/2023	13:35
20	O3330-02	1	Solid	20.02	20	23.9	8.64	06/19/2023	13:40

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O3330-08	1	Solid	20.03	20	23.9	8.50	06/19/2023	13:42
22	O3330-14	1	Solid	20.05	20	24.0	8.24	06/19/2023	13:45
23	O3331-02	1	Solid	20.04	20	23.6	8.19	06/19/2023	13:50
24	O3331-09	1	Solid	20.02	20	23.6	8.22	06/19/2023	14:00
25	O3331-15	1	Solid	20.01	20	23.8	8.30	06/19/2023	14:05
26	O3331-21	1	Solid	20.02	20	24.0	8.30	06/19/2023	14:10
27	CCV3	1	Water	NA	NA	20.8	2.01	06/19/2023	14:15
28	O3331-27	1	Solid	20.03	20	24.0	8.58	06/19/2023	14:25
29	O3331-27DUP	1	Solid	20.04	20	24.1	8.59	06/19/2023	14:26
30	CCV4	1	Water	NA	NA	20.7	12.02	06/19/2023	14:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity o3331

WorkList ID : 171079

Department : Wet-Chemistry

Date : 06-19-2023 12:11:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3327-02	COMPOSIT-TW-03	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3327-08	COMPOSIT-TW-02	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3327-14	COMPOSIT-TW-06	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3327-20	COMPOSIT-SB-14	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3328-02	COMPOSIT-SW02-8-15	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3328-08	COMPOSIT-SW04-0-8	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3328-14	COMPOSIT-SW02-0-8	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3328-20	COMPOSIT-SW04-8-15	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3329-02	COMPOSIT-SW03-0-8	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3329-08	COMPOSIT-SW03-8-15	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3329-14	COMPOSIT-SW01-0-8	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3329-20	COMPOSIT-SW01-8-15	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3330-02	COMPOSITE-MC03	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3330-08	COMPOSITE-MC01	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3330-14	COMPOSITE-MC02	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3331-02	COMPOSITE-SB8	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3331-09	COMPOSITE-SB17	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3331-15	COMPOSITE-SB9	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3331-21	COMPOSITE-SB30-35	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D
O3331-27	COMPOSITE-SB18-27	Solid	Corrosivity	Cool 4 deg C	SOUN03	G11	06/16/2023	9045D

Date/Time 06/19/23 12:30

Raw Sample Received by: J.D. WOOD

Raw Sample Relinquished by: JTCsm

Date/Time 06/19/23

Raw Sample Received by: JTCsm

Raw Sample Relinquished by: J.D. WOOD