

LB126062


7471B

INSTRUMENT ID: 8V1

Linear

 μ Abs.:

66800



A= 0.0000e+000

B= 1.4982e-004

C= 5.2479e-004

Rho= 0.9998902

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.010	0.010	63	0.000	63					-2
0.2	0.200	0.197	-0.003	1310	0.0 %	1310					-2
2.5	2.500	2.446	-0.054	16322	0.0 %	16322					2
5.0	5.000	5.103	0.103	34060	0.0 %	34060					-1
7.5	7.500	7.436	-0.064	49627	0.0 %	49627					0
10.0	10.000	10.008	0.008	66800	0.0 %	66800					

LB126062 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Jun 2023 08:12:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	63	-	0.0000	7471B	PPB	20 Jun 2023 08:20:13	S	Std
0.2 - 1	50.2	1310	-	0.2000	7471B	PPB	20 Jun 2023 08:22:31	S	Std
2.5 - 1	52.5	16322	-	2.5000	7471B	PPB	20 Jun 2023 08:24:48	S	Std
5.0 - 1	55	34060	-	5.0000	7471B	PPB	20 Jun 2023 08:27:05	S	Std
7.5 - 1	52.5	49627	-	7.5000	7471B	PPB	20 Jun 2023 08:29:23	S	Std
10.0 - 1	510	66800	-	10.0000	7471B	PPB	20 Jun 2023 08:34:26	S	Std
ICV29 - 1	ICV29	24658	3.6948	-	7471B	PPB	20 Jun 2023 08:39:40	U	SMPL
ICB29 - 1	ICB29	-81	-0.0116	-	7471B	PPB	20 Jun 2023 08:41:55	U	SMPL
CCV73 - 1	CCV73	33660	5.0434	-	7471B	PPB	20 Jun 2023 08:46:46	U	SMPL
CCB73 - 1	CCB73	-103	-0.0149	-	7471B	PPB	20 Jun 2023 08:49:02	U	SMPL
CRA - 1	CRA	1330	0.1998	-	7471B	PPB	20 Jun 2023 08:51:19	U	SMPL
HighStd - 1	HighStd	66601	9.9786	-	7471B	PPB	20 Jun 2023 08:53:35	U	SMPL
ChkStd - 1	ChkStd	45349	6.7947	-	7471B	PPB	20 Jun 2023 08:55:50	U	SMPL
PB153604BL - 1	PBS	-274	-0.0405	-	7471B	PPB	20 Jun 2023 08:58:09	U	SMPL
PB153604BS - 1	LCSS	25803	3.8663	-	7471B	PPB	20 Jun 2023 09:03:01	U	SMPL
O3316-01 - 1	SU-4-061623	1819	0.2730	-	7471B	PPB	20 Jun 2023 09:05:17	U	SMPL
O3320-01 - 1	WC-19	2920	0.4380	-	7471B	PPB	20 Jun 2023 09:07:35	U	SMPL
O3320-04 - 1	WC-10	1740	0.2612	-	7471B	PPB	20 Jun 2023 09:09:52	U	SMPL
O3320-07 - 1	WC-12	843	0.1268	-	7471B	PPB	20 Jun 2023 09:12:10	U	SMPL
O3320-10 - 1	WC-13	1356	0.2037	-	7471B	PPB	20 Jun 2023 09:14:25	U	SMPL
CCV74 - 1	CCV74	32933	4.9345	-	7471B	PPB	20 Jun 2023 09:16:41	U	SMPL
CCB74 - 1	CCB74	-75	-0.0107	-	7471B	PPB	20 Jun 2023 09:18:56	U	SMPL
O3320-13 - 1	WC-14	1170	0.1758	-	7471B	PPB	20 Jun 2023 09:21:13	U	SMPL
O3320-16 - 1	WC-16	1184	0.1779	-	7471B	PPB	20 Jun 2023 09:23:29	U	SMPL
O3320-19 - 1	WC-14A	1077	0.1619	-	7471B	PPB	20 Jun 2023 09:25:44	U	SMPL
O3321-01 - 1	WC-25	759	0.1142	-	7471B	PPB	20 Jun 2023 09:28:01	U	SMPL
O3321-04 - 1	WC-24	2598	0.3898	-	7471B	PPB	20 Jun 2023 09:30:17	U	SMPL
O3321-07 - 1	WC-23	693	0.1043	-	7471B	PPB	20 Jun 2023 09:32:33	U	SMPL
O3321-10 - 1	WC-22	1037	0.1559	-	7471B	PPB	20 Jun 2023 09:34:51	U	SMPL
O3321-13 - 1	WC-21	1977	0.2967	-	7471B	PPB	20 Jun 2023 09:37:08	U	SMPL
O3321-16 - 1	WC-20	1180	0.1773	-	7471B	PPB	20 Jun 2023 09:39:25	U	SMPL
O3322-01 - 1	HD-02-061623	1522	0.2285	-	7471B	PPB	20 Jun 2023 09:41:41	U	SMPL
CCV75 - 1	CCV75	32875	4.9258	-	7471B	PPB	20 Jun 2023 09:57:54	U	SMPL
CCB75 - 1	CCB75	-47	-0.0065	-	7471B	PPB	20 Jun 2023 10:00:09	U	SMPL
O3327-01 - 1	COMPOSIT-TW-03	19373	2.9030	-	7471B	PPB	20 Jun 2023 10:02:26	U	SMPL
O3327-01DUP - 1	COMPOSIT-TW-03DUP	20437	3.0624	-	7471B	PPB	20 Jun 2023 10:04:42	U	SMPL
O3327-01MS - 1	COMPOSIT-TW-03MS	51055	7.6495	-	7471B	PPB	20 Jun 2023 10:06:58	U	SMPL
O3327-01MSD - 1	COMPOSIT-TW-03MSD	49377	7.3981	-	7471B	PPB	20 Jun 2023 10:09:15	U	SMPL
O3327-07 - 1	COMPOSIT-TW-02	7492	1.1230	-	7471B	PPB	20 Jun 2023 10:11:33	U	SMPL
O3327-13 - 1	COMPOSIT-TW-06	6796	1.0187	-	7471B	PPB	20 Jun 2023 10:13:52	U	SMPL
O3327-19 - 1	COMPOSIT-SB-14	36810	5.5154	-	7471B	PPB	20 Jun 2023 10:16:09	U	SMPL
PB153605BL - 1	PBS	-164	-0.0240	-	7471B	PPB	20 Jun 2023 10:18:27	U	SMPL
O3328-01 - 1	COMPOSIT-SW02-8-15	2976	0.4464	-	7471B	PPB	20 Jun 2023 10:23:02	U	SMPL
CCV76 - 1	CCV76	34480	5.1663	-	7471B	PPB	20 Jun 2023 10:25:19	U	SMPL
CCB76 - 1	CCB76	-32	-0.0043	-	7471B	PPB	20 Jun 2023 10:27:35	U	SMPL
O3328-07 - 1	COMPOSIT-SW04-0-8	18519	2.7750	-	7471B	PPB	20 Jun 2023 10:29:53	U	SMPL
O3328-13 - 1	COMPOSIT-SW02-0-8	16938	2.5382	-	7471B	PPB	20 Jun 2023 10:32:09	U	SMPL
O3328-19 - 1	COMPOSIT-SW04-8-15	15211	2.2794	-	7471B	PPB	20 Jun 2023 10:34:26	U	SMPL
O3329-01 - 1	COMPOSIT-SW03-0-8	23499	3.5211	-	7471B	PPB	20 Jun 2023 10:36:44	U	SMPL
O3329-01DUP - 1	COMPOSIT-SW03-0-8DUP	25322	3.7942	-	7471B	PPB	20 Jun 2023 10:39:01	U	SMPL
O3329-01MS - 1	COMPOSIT-SW03-0-8MS	46583	6.9795	-	7471B	PPB	20 Jun 2023 10:44:28	U	SMPL
O3329-01MSD - 1	COMPOSIT-SW03-0-8MSD	47321	7.0901	-	7471B	PPB	20 Jun 2023 10:46:45	U	SMPL
O3329-07 - 1	COMPOSIT-SW03-8-15	17080	2.5594	-	7471B	PPB	20 Jun 2023 10:49:04	U	SMPL
O3329-13 - 1	COMPOSIT-SW01-0-8	16070	2.4081	-	7471B	PPB	20 Jun 2023 10:51:24	U	SMPL
O3329-19 - 1	COMPOSIT-SW01-8-15	9319	1.3967	-	7471B	PPB	20 Jun 2023 10:53:41	U	SMPL
CCV77 - 1	CCV77	34820	5.2172	-	7471B	PPB	20 Jun 2023 10:55:58	U	SMPL
CCB77 - 1	CCB77	-21	-0.0026	-	7471B	PPB	20 Jun 2023 10:58:14	U	SMPL
O3330-01 - 1	COMPOSITE-MC03	25525	3.8247	-	7471B	PPB	20 Jun 2023 11:00:32	U	SMPL
O3330-07 - 1	COMPOSITE-MC01	16654	2.4956	-	7471B	PPB	20 Jun 2023 11:02:47	U	SMPL
O3330-13 - 1	COMPOSITE-MC02	10614	1.5907	-	7471B	PPB	20 Jun 2023 11:05:05	U	SMPL
O3331-01 - 1	COMPOSITE-SB8	24501	3.6712	-	7471B	PPB	20 Jun 2023 11:07:22	U	SMPL
O3331-08 - 1	COMPOSITE-SB17	29936	4.4855	-	7471B	PPB	20 Jun 2023 11:09:38	U	SMPL
O3331-14 - 1	COMPOSITE-SB9	30115	4.5123	-	7471B	PPB	20 Jun 2023 11:11:56	U	SMPL
O3331-20 - 1	COMPOSITE-SB30-35	15997	2.3972	-	7471B	PPB	20 Jun 2023 11:14:14	U	SMPL
O3331-26 - 1	COMPOSTIE-SB18-27	22271	3.3371	-	7471B	PPB	20 Jun 2023 11:16:34	U	SMPL
CCV78 - 1	CCV78	34745	5.2060	-	7471B	PPB	20 Jun 2023 11:18:52	U	SMPL
CCB78 - 1	CCB78	-26	-0.0034	-	7471B	PPB	20 Jun 2023 11:21:09	U	SMPL
PB153605BS - 1	LCSS	27069	4.0560	-	7471B	PPB	20 Jun 2023 11:23:28	U	SMPL
O3327-01LX5 - 1		3991	0.5985	-	7471B	PPB	20 Jun 2023 11:25:45	U	SMPL
O3327-01A - 1		45981	6.8894	-	7471B	PPB	20 Jun 2023 11:28:03	U	SMPL
O3329-01LX5 - 1		4523	0.6782	-	7471B	PPB	20 Jun 2023 11:30:19	U	SMPL
O3329-01A - 1		50644	7.5880	-	7471B	PPB	20 Jun 2023 11:32:39	U	SMPL
CCV79 - 1	CCV79	34117	5.1119	-	7471B	PPB	20 Jun 2023 11:34:55	U	SMPL
CCB79 - 1	CCB79	-73	-0.0104	-	7471B	PPB	20 Jun 2023 11:37:14	U	SMPL