

LB126075


7470A

INSTRUMENT 1A: CV2

Linear

 μ Abs.:

45993



A= 0.0000e+000

B= 2.1784e-004

C= -3.3560e-002

Rho= 0.9999693

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%d
0.0	0.000	-0.027	-0.027	29	0.000	29	0				5
0.2	0.200	0.222	0.022	1172	0.0 %	1172					1
2.5	2.500	2.523	0.023	11737	0.0 %	11737					1
5.0	5.000	4.959	-0.041	22920	0.0 %	22920					1
7.5	7.500	7.537	0.037	34755	0.0 %	34755					0
10.0	10.000	9.986	-0.014	45993	0.0 %	45993					0

LB126075 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jun 2023 14:34:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	29	-	0.0000	7470A	PPB	20 Jun 2023 14:40:36	S	Std
0.2 - 1	SP.2	1172	-	0.2000	7470A	PPB	20 Jun 2023 14:42:53	S	Std
2.5 - 1	SP.5	11737	-	2.5000	7470A	PPB	20 Jun 2023 14:45:11	S	Std
5.0 - 1	SP	22920	-	5.0000	7470A	PPB	20 Jun 2023 14:47:28	S	Std
7.5 - 1	SP.5	34755	-	7.5000	7470A	PPB	20 Jun 2023 14:52:56	S	Std
10.0 - 1	SP.10	45993	-	10.0000	7470A	PPB	20 Jun 2023 14:58:02	S	Std
ICV30 - 1	ICV30	19353	4.1823	-	7470A	PPB	20 Jun 2023 15:00:56	U	SMPL
ICB30 - 1	ICB30	-173	-0.0712	-	7470A	PPB	20 Jun 2023 15:03:11	U	SMPL
CCV80 - 1	CCV80	23664	5.1214	-	7470A	PPB	20 Jun 2023 15:05:27	U	SMPL
CCB80 - 1	CCB80	-140	-0.0641	-	7470A	PPB	20 Jun 2023 15:07:43	U	SMPL
CRA - 1	CRA	955	0.1745	-	7470A	PPB	20 Jun 2023 15:10:00	U	SMPL
HighStd - 1	HighStd	45867	9.9581	-	7470A	PPB	20 Jun 2023 15:12:15	U	SMPL
ChkStd - 1	ChkStd	31421	6.8112	-	7470A	PPB	20 Jun 2023 15:14:31	U	SMPL
PB153565BL - 1	PBW	232	0.0170	-	7470A	PPB	20 Jun 2023 15:16:48	U	SMPL
PB153565BS - 1	LCSW	16777	3.6211	-	7470A	PPB	20 Jun 2023 15:21:50	U	SMPL
O3289-03 - 1	SB-01-COMP	-29	-0.0399	-	7470A	PPB	20 Jun 2023 15:24:06	U	SMPL
O3289-03DUP - 1	SB-01-COMPDUP	50	-0.0227	-	7470A	PPB	20 Jun 2023 15:26:23	U	SMPL
O3289-03MS - 1	SB-01-COMPMS	17570	3.7939	-	7470A	PPB	20 Jun 2023 15:28:40	U	SMPL
O3289-03MSD - 1	SB-01-COMPMSD	17176	3.7080	-	7470A	PPB	20 Jun 2023 15:30:57	U	SMPL
O3289-06 - 1	SB-02-COMP	659	0.1100	-	7470A	PPB	20 Jun 2023 15:33:14	U	SMPL
CCV81 - 1	CCV81	23820	5.1554	-	7470A	PPB	20 Jun 2023 15:35:31	U	SMPL
CCB81 - 1	CCB81	-140	-0.0641	-	7470A	PPB	20 Jun 2023 15:37:46	U	SMPL
O3289-09 - 1	SB-03-COMP	783	0.1370	-	7470A	PPB	20 Jun 2023 15:40:03	U	SMPL
O3289-11 - 1	SB-04-COMP	365	0.0460	-	7470A	PPB	20 Jun 2023 15:42:19	U	SMPL
O3289-12 - 1	SB-05-0-2.0	295	0.0307	-	7470A	PPB	20 Jun 2023 15:44:35	U	SMPL
O3289-14 - 1	SB-05-COMP	238	0.0183	-	7470A	PPB	20 Jun 2023 15:46:51	U	SMPL
O3289-18 - 1	SB-06-COMP	121	-0.0072	-	7470A	PPB	20 Jun 2023 15:49:07	U	SMPL
O3289-21 - 1	SB-07-COMP	130	-0.0052	-	7470A	PPB	20 Jun 2023 15:51:24	U	SMPL
O3289-24 - 1	SB-08-COMP	105	-0.0107	-	7470A	PPB	20 Jun 2023 15:53:40	U	SMPL
O3296-02 - 1	COMPOSIT-R-02	56	-0.0214	-	7470A	PPB	20 Jun 2023 15:55:57	U	SMPL
O3296-08 - 1	COMPOSIT-R-01	44	-0.0240	-	7470A	PPB	20 Jun 2023 15:58:14	U	SMPL
O3296-14 - 1	COMPOSIT-TW-01	109	-0.0098	-	7470A	PPB	20 Jun 2023 16:00:30	U	SMPL
CCV82 - 1	CCV82	23776	5.1458	-	7470A	PPB	20 Jun 2023 16:02:46	U	SMPL
CCB82 - 1	CCB82	-30	-0.0401	-	7470A	PPB	20 Jun 2023 16:05:02	U	SMPL
O3296-20 - 1	COMPOSIT-TW-07	-27	-0.0394	-	7470A	PPB	20 Jun 2023 16:07:19	U	SMPL
O3297-02 - 1	COMPOSIT-TP-6-6-9	-4	-0.0344	-	7470A	PPB	20 Jun 2023 16:09:35	U	SMPL
O3298-02 - 1	COMPOSIT-TW-04	-23	-0.0386	-	7470A	PPB	20 Jun 2023 16:11:51	U	SMPL
O3298-08 - 1	COMPOSIT-TW-05	3	-0.0329	-	7470A	PPB	20 Jun 2023 16:14:06	U	SMPL
PB153566BL - 1	PBW	-10	-0.0357	-	7470A	PPB	20 Jun 2023 16:16:23	U	SMPL
PB153566BS - 1	LCSW	17219	3.7174	-	7470A	PPB	20 Jun 2023 16:18:39	U	SMPL
O3308-02 - 1	SP-LEFT	-127	-0.0612	-	7470A	PPB	20 Jun 2023 16:20:56	U	SMPL
O3308-05 - 1	SP-RIGHT	-7	-0.0351	-	7470A	PPB	20 Jun 2023 16:23:14	U	SMPL
O3320-02 - 1	WC-19	46	-0.0235	-	7470A	PPB	20 Jun 2023 16:25:31	U	SMPL
O3320-05 - 1	WC-10	26	-0.0279	-	7470A	PPB	20 Jun 2023 16:27:47	U	SMPL
CCV83 - 1	CCV83	23479	5.0811	-	7470A	PPB	20 Jun 2023 16:30:03	U	SMPL
CCB83 - 1	CCB83	44	-0.0240	-	7470A	PPB	20 Jun 2023 16:32:19	U	SMPL
O3320-08 - 1	WC-12	100	-0.0118	-	7470A	PPB	20 Jun 2023 16:34:36	U	SMPL
O3320-11 - 1	WC-13	-29	-0.0399	-	7470A	PPB	20 Jun 2023 16:36:52	U	SMPL
O3320-14 - 1	WC-14	-4	-0.0344	-	7470A	PPB	20 Jun 2023 16:39:07	U	SMPL
O3320-17 - 1	WC-16	-10	-0.0357	-	7470A	PPB	20 Jun 2023 16:41:23	U	SMPL
O3320-20 - 1	WC-14A	-21	-0.0381	-	7470A	PPB	20 Jun 2023 16:43:39	U	SMPL
O3321-02 - 1	WC-25	-28	-0.0397	-	7470A	PPB	20 Jun 2023 16:45:56	U	SMPL
O3321-05 - 1	WC-24	-13	-0.0364	-	7470A	PPB	20 Jun 2023 16:48:13	U	SMPL
O3321-08 - 1	WC-23	56	-0.0214	-	7470A	PPB	20 Jun 2023 16:50:31	U	SMPL
O3321-11 - 1	WC-22	75	-0.0172	-	7470A	PPB	20 Jun 2023 16:52:48	U	SMPL
O3321-14 - 1	WC-21	17	-0.0299	-	7470A	PPB	20 Jun 2023 16:55:04	U	SMPL
CCV84 - 1	CCV84	21037	4.5491	-	7470A	PPB	20 Jun 2023 16:57:20	U	SMPL
CCB84 - 1	CCB84	74	-0.0174	-	7470A	PPB	20 Jun 2023 16:59:36	U	SMPL
O3321-17 - 1	WC-20	7	-0.0320	-	7470A	PPB	20 Jun 2023 17:01:53	U	SMPL
O3321-17DUP - 1	WC-20DUP	27	-0.0277	-	7470A	PPB	20 Jun 2023 17:04:08	U	SMPL
O3321-17MS - 1	WC-20MS	17883	3.8621	-	7470A	PPB	20 Jun 2023 17:06:24	U	SMPL
O3321-17MSD - 1	WC-20MSD	15790	3.4061	-	7470A	PPB	20 Jun 2023 17:08:40	U	SMPL
PB153567BL - 1	PBW	-84	-0.0519	-	7470A	PPB	20 Jun 2023 17:10:56	U	SMPL
PB153567BS - 1	LCSW	16266	3.5098	-	7470A	PPB	20 Jun 2023 17:13:14	U	SMPL
O3279-02 - 1	TOPSOIL-OU3-DENALI-1	150	-0.0009	-	7470A	PPB	20 Jun 2023 17:15:30	U	SMPL
O3279-02DUP - 1	TOPSOIL-OU3-DENALI-1DUP	409	0.0555	-	7470A	PPB	20 Jun 2023 17:17:47	U	SMPL
O3279-02MS - 1	TOPSOIL-OU3-DENALI-1MS	16123	3.4787	-	7470A	PPB	20 Jun 2023 17:20:07	U	SMPL
O3279-02MSD - 1	TOPSOIL-OU3-DENALI-1MSD	16281	3.5131	-	7470A	PPB	20 Jun 2023 17:22:23	U	SMPL
CCV85 - 1	CCV85	22630	4.8961	-	7470A	PPB	20 Jun 2023 17:24:41	U	SMPL
CCB85 - 1	CCB85	106	-0.0105	-	7470A	PPB	20 Jun 2023 17:26:58	U	SMPL
O3289-03LX5 - 1		-13	-0.0364	-	7470A	PPB	20 Jun 2023 17:29:15	U	SMPL
O3289-03A - 1		19716	4.2614	-	7470A	PPB	20 Jun 2023 17:31:32	U	SMPL
O3321-17LX5 - 1		-113	-0.0582	-	7470A	PPB	20 Jun 2023 17:33:48	U	SMPL
O3321-17A - 1		20043	4.3326	-	7470A	PPB	20 Jun 2023 17:36:05	U	SMPL
O3279-02LX5 - 1		-153	-0.0669	-	7470A	PPB	20 Jun 2023 17:38:22	U	SMPL
O3279-02A - 1		20396	4.4095	-	7470A	PPB	20 Jun 2023 17:40:39	U	SMPL
PB153510TB - 1	PB153510TB	-31	-0.0403	-	7470A	PPB	20 Jun 2023 17:42:56	U	SMPL
PB153530TB - 1	PB153530TB	10	-0.0314	-	7470A	PPB	20 Jun 2023 17:45:14	U	SMPL
PB153567TB - 1	PB153532TB	7	-0.0320	-	7470A	PPB	20 Jun 2023 17:47:31	U	SMPL
CCV86 - 1	CCV86	22962	4.9685	-	7470A	PPB	20 Jun 2023 17:49:48	U	SMPL
CCB86 - 1	CCB86	-98	-0.0549	-	7470A	PPB	20 Jun 2023 17:52:04	U	SMPL