



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126082

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/20/2023	15:15
2	CAL2	1	Water	NA	NA	20.7	7.00	06/20/2023	15:16
3	CAL3	1	Water	NA	NA	20.7	10.02	06/20/2023	15:19
4	ICV	1	Water	NA	NA	20.6	6.98	06/20/2023	15:25
5	CCV1	1	Water	NA	NA	20.7	2.01	06/20/2023	15:30
6	O3352-01	1	Solid	20.02	20	24.7	10.22	06/20/2023	15:33
7	O3352-01DUP	1	Solid	20.01	20	24.8	10.23	06/20/2023	15:34
8	CCV2	1	Water	NA	NA	20.8	12.02	06/20/2023	15:35

171129

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O3352

WorkList ID : 171129

Department : Wet-Chemistry

Date : 06-20-2023 15:03:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3352-01	BU-3-062023	Solid	pH	Cool 4 deg C	PSEG05	H11	06/20/2023	9045D

171129

Date/Time 06/20/23

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time 06/20/23

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

171129
JDC