



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126083

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/20/2023	15:15
2	CAL2	1	Water	NA	NA	20.7	7.00	06/20/2023	15:16
3	CAL3	1	Water	NA	NA	20.7	10.02	06/20/2023	15:19
4	ICV	1	Water	NA	NA	20.6	6.98	06/20/2023	15:25
5	CCV1	1	Water	NA	NA	20.7	2.01	06/20/2023	15:30
6	O3342-01	1	Water	NA	NA	21.7	6.42	06/20/2023	15:45
7	O3342-01DUP	1	Water	NA	NA	21.8	6.43	06/20/2023	15:46
8	O3345-03	1	Water	NA	NA	25.3	7.26	06/20/2023	15:50
9	O3345-04	1	Water	NA	NA	23.7	6.51	06/20/2023	16:00
10	CCV2	1	Water	NA	NA	20.8	12.03	06/20/2023	16:02

17126083

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O3342 W

WorkList ID : 171130

Department : Wet-Chemistry

Date : 06-20-2023 15:04:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3342-01	C 50047	Water	pH	Cool 4 deg C	PSEG03	H11	06/20/2023	9040C
O3345-03	A COMP-2-0620	Water	pH	Cool 4 deg C	PSEG03	H11	06/20/2023	9040C
O3345-04	A 5774	Water	pH	Cool 4 deg C	PSEG03	H11	06/20/2023	9040C

15115

Date/Time 06/20/23

Raw Sample Received by: JWC

Raw Sample Relinquished by: JWC

Date/Time 06/20/23

Raw Sample Received by: JWC

Raw Sample Relinquished by: JWC

171200

JWC