



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB126084

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/20/2023	15:15
2	CAL2	1	Water	NA	NA	20.7	7.00	06/20/2023	15:16
3	CAL3	1	Water	NA	NA	20.7	10.02	06/20/2023	15:19
4	ICV	1	Water	NA	NA	20.6	6.98	06/20/2023	15:25
5	CCV1	1	Water	NA	NA	20.8	2.01	06/20/2023	16:00
6	O3333-02	1	Solid	20.02	20	22.6	10.38	06/20/2023	16:05
7	O3333-02DUP	1	Solid	20.03	20	22.7	10.39	06/20/2023	16:06
8	O3336-02	1	Solid	20.03	20	22.6	8.56	06/20/2023	16:10
9	O3336-05	1	Solid	20.04	20	22.7	8.23	06/20/2023	16:15
10	O3336-08	1	Solid	20.02	20	22.5	7.74	06/20/2023	16:20
11	O3336-11	1	Solid	20.01	20	22.4	7.19	06/20/2023	16:21
12	O3336-14	1	Solid	20.02	20	22.3	7.09	06/20/2023	16:25
13	O3336-17	1	Solid	20.03	20	22.4	7.68	06/20/2023	16:30
14	O3337-02	1	Solid	20.01	20	22.4	6.94	06/20/2023	16:32
15	O3337-05	1	Solid	20.02	20	22.3	7.09	06/20/2023	16:35
16	CCV2	1	Water	NA	NA	20.9	12.02	06/20/2023	16:40
17	O3337-08	1	Solid	20.03	20	22.4	7.30	06/20/2023	16:43
18	O3337-11	1	Solid	20.02	20	22.2	6.90	06/20/2023	16:45
19	O3337-14	1	Solid	20.01	20	22.3	7.18	06/20/2023	16:50
20	O3337-17	1	Solid	20.04	20	21.7	7.16	06/20/2023	16:55

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O3337-20	1	Solid	20.02	20	21.9	7.30	06/20/2023	17:00
22	O3345-01	1	Solid	20.02	20	24.6	5.55	06/20/2023	17:05
23	CCV3	1	Water	NA	NA	20.8	2.01	06/20/2023	17:10

WORKLIST(Hardcopy Internal Chain)

126084

WorkList Name : corrosivity o3336 WorkList ID : 17117 Department : Wet-Chemistry Date : 06-20-2023 08:50:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3333-02	DOMINICK-SP	Solid	Corrosivity	Cool 4 deg C	TULL02	G11	06/19/2023	9045D
O3336-02	WC-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3336-05	WC-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3336-08	WC-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3336-11	WC-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3336-14	WC-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3336-17	WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3337-02	WC-18	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3337-05	WC-17	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3337-08	WC-9	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3337-11	WC-8	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3337-14	WC-7	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3337-17	WC-15	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3337-20	WC-11	Solid	Corrosivity	Cool 4 deg C	PSEG03	G11	06/19/2023	9045D
O3345-01	COMP-1-0620	Solid	Corrosivity	Cool 4 deg C	PSEG03	H11	06/20/2023	9045D

Date/Time 06/20/23 15:10
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC

Date/Time 06/20/23 17:30
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC