

Analysis Method: 9034

ANALYST: Rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB126087

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2976
IODINE SOLUTION .025N 1L	W2924
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB153622BL		1	5.00	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	06/21/2023	09:00
2	O3321-02		1	5.06	50	2.00	0.00	1.90	1.90	0.10	0.02	1.58	06/21/2023	09:02
3	O3321-02DUP		1	5.06	50	2.00	0.00	1.90	1.90	0.10	0.02	1.58	06/21/2023	09:05
4	O3321-05		1	5.03	50	2.00	0.00	1.84	1.84	0.16	0.08	6.36	06/21/2023	09:07
5	O3321-08		1	5.05	50	2.00	0.00	1.82	1.82	0.18	0.10	7.92	06/21/2023	09:10
6	O3321-11		1	5.04	50	2.00	0.00	1.82	1.82	0.18	0.10	7.94	06/21/2023	09:12
7	O3321-14		1	5.03	50	2.00	0.00	1.88	1.88	0.12	0.04	3.18	06/21/2023	09:15
8	O3321-17		1	5.02	50	2.00	0.00	1.88	1.88	0.12	0.04	3.19	06/21/2023	09:18
9	O3327-02		1	5.03	50	2.00	0.00	1.86	1.86	0.14	0.06	4.77	06/21/2023	09:20
10	O3327-08		1	5.06	50	2.00	0.00	1.82	1.82	0.18	0.10	7.91	06/21/2023	09:23
11	O3327-14		1	5.02	50	2.00	0.00	1.84	1.84	0.16	0.08	6.37	06/21/2023	09:25
12	O3327-20		1	5.03	50	2.00	0.00	1.84	1.84	0.16	0.08	6.36	06/21/2023	09:28

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Analysis Method: 9034

ANALYST: Rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB126087

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2976
IODINE SOLUTION .025N 1L	W2924
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
13	O3336-02		1	5.02	50	2.00	0.00	1.84	1.84	0.16	0.08	6.37	06/21/2023	09:30
14	O3336-05		1	5.03	50	2.00	0.00	1.88	1.88	0.12	0.04	3.18	06/21/2023	09:33
15	O3336-08		1	5.06	50	2.00	0.00	1.82	1.82	0.18	0.10	7.91	06/21/2023	09:35
16	O3336-11		1	5.06	50	2.00	0.00	1.82	1.82	0.18	0.10	7.91	06/21/2023	09:38
17	O3336-14		1	5.05	50	2.00	0.00	1.86	1.86	0.14	0.06	4.75	06/21/2023	09:40
18	O3336-17		1	5.07	50	2.00	0.00	1.82	1.82	0.18	0.10	7.89	06/21/2023	09:42

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume