



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:50
In Date: 06/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 06/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126091

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3270-01	EZDS9	1	1.12	8.53	9.65	8.83	90.4	
O3270-02	EZDT0	2	1.15	8.41	9.56	8.68	89.5	
O3270-03	EZDT1	3	1.13	8.68	9.81	8.69	87.1	
O3270-04	EZDT2	4	1.13	8.36	9.49	8.71	90.7	
O3270-05	EZDT3	5	1.14	8.58	9.72	9.11	92.9	
O3270-06	EZDT4	6	1.15	8.52	9.67	8.75	89.2	
O3270-07	EZDT5	7	1.14	8.53	9.67	8.66	88.2	
O3270-08	EZDT6	8	1.14	8.28	9.42	8.44	88.2	
O3270-09	EZDT7	9	1.13	8.38	9.51	8.45	87.4	
O3270-10	EZDT8	10	1.13	8.46	9.59	8.82	90.9	
O3270-11	EZDT9	11	1.13	8.60	9.73	8.93	90.7	
O3270-12	EZDW0	12	1.13	8.41	9.54	8.64	89.3	
O3270-13	EZDW1	13	1.16	8.47	9.63	8.2	83.1	
O3270-14	EZDW2	14	1.14	8.69	9.83	8.54	85.2	
O3270-15	EZDW3	15	1.12	8.66	9.78	8.87	89.5	
O3270-16	EZDW4	16	1.14	8.32	9.46	8.47	88.1	
O3270-17	EZDW5	17	1.12	8.79	9.91	8.88	88.3	
O3270-18	EZDW6	18	1.15	8.74	9.89	8.53	84.4	
O3270-19	EZE07	19	1.14	8.73	9.87	8.54	84.8	
O3270-20	EZE09	20	1.13	8.54	9.67	8.68	88.4	
O3270-21	EZE09D	21	1.12	8.55	9.67	8.68	88.4	
O3270-22	EZE09S	22	1.12	8.55	9.67	8.68	88.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$