

LB126093


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

65175



A= 0.0000e+000

B= 1.5275e-004

C= -4.7830e-002

Rho= 0.9995268

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.015	-0.015	214	0.000	214					
0.2	0.200	0.207	0.007	1668	0.0 %	1668					
2.5	2.500	2.562	0.062	17083	0.0 %	17083					
5.0	5.000	4.841	-0.159	32002	0.0 %	32002					
7.5	7.500	7.698	0.198	50710	0.0 %	50710					
10.0	10.000	9.908	-0.092	65175	0.0 %	65175					

0/0D
 3
 2
 3
 3
 1

LB126093 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jun 2023 09:57:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	214	-	0.000	7471B	PPB	21 Jun 2023 09:58:55	S	Std
0.2 - 1	50.2	1668	-	0.200	7471B	PPB	21 Jun 2023 10:01:13	S	Std
2.5 - 1	52.5	17083	-	2.500	7471B	PPB	21 Jun 2023 10:03:31	S	Std
5.0 - 1	55	32002	-	5.000	7471B	PPB	21 Jun 2023 10:05:48	S	Std
7.5 - 1	57.5	50710	-	7.500	7471B	PPB	21 Jun 2023 10:08:05	S	Std
10.0 - 1	511	65175	-	10.000	7471B	PPB	21 Jun 2023 10:13:09	S	Std
ICV32 - 1	ICV32	26344	3.9763	-	7471B	PPB	21 Jun 2023 10:17:06	U	SMPL
ICB32 - 1	ICB32	3	-0.0474	-	7471B	PPB	21 Jun 2023 10:19:21	U	SMPL
CCV90 - 1	CCV90	32035	4.8456	-	7471B	PPB	21 Jun 2023 10:21:38	U	SMPL
CCB90 - 1	CCB90	-43	-0.0544	-	7471B	PPB	21 Jun 2023 10:23:56	U	SMPL
CRA - 1	CRA	1650	0.2042	-	7471B	PPB	21 Jun 2023 10:26:14	U	SMPL
HighStd - 1	HighStd	65333	9.9319	-	7471B	PPB	21 Jun 2023 10:28:30	U	SMPL
ChkStd - 1	ChkStd	47216	7.1645	-	7471B	PPB	21 Jun 2023 10:30:45	U	SMPL
PB153641BL - 1	PBS	-280	-0.0906	-	7471B	PPB	21 Jun 2023 10:33:04	U	SMPL
PB153641BS - 1	LCSS	26008	3.9249	-	7471B	PPB	21 Jun 2023 10:39:18	U	SMPL
O3333-01 - 1	DOMINICK-SP	29648	4.4810	-	7471B	PPB	21 Jun 2023 10:41:34	U	SMPL
O3335-01 - 1	NB-7-061923	688	0.0573	-	7471B	PPB	21 Jun 2023 10:43:54	U	SMPL
O3335-01DUP - 1	NB-7-061923DUP	948	0.0970	-	7471B	PPB	21 Jun 2023 10:46:14	U	SMPL
O3335-01MS - 1	NB-7-061923MS	24535	3.6999	-	7471B	PPB	21 Jun 2023 10:48:31	U	SMPL
O3335-01MSD - 1	NB-7-061923MSD	25064	3.7807	-	7471B	PPB	21 Jun 2023 10:50:47	U	SMPL
CCV91 - 1	CCV91	31869	4.8202	-	7471B	PPB	21 Jun 2023 10:53:04	U	SMPL
CCB91 - 1	CCB91	14	-0.0457	-	7471B	PPB	21 Jun 2023 10:55:22	U	SMPL
O3336-01 - 1	WC-6	1712	0.2137	-	7471B	PPB	21 Jun 2023 10:57:39	U	SMPL
O3336-04 - 1	WC-5	2782	0.3771	-	7471B	PPB	21 Jun 2023 10:59:55	U	SMPL
O3336-07 - 1	WC-4	1084	0.1178	-	7471B	PPB	21 Jun 2023 11:02:11	U	SMPL
O3336-10 - 1	WC-3	3260	0.4501	-	7471B	PPB	21 Jun 2023 11:04:27	U	SMPL
O3336-13 - 1	WC-2	2524	0.3377	-	7471B	PPB	21 Jun 2023 11:06:43	U	SMPL
O3336-16 - 1	WC-1	865	0.0843	-	7471B	PPB	21 Jun 2023 11:09:00	U	SMPL
O3337-01 - 1	WC-18	1076	0.1165	-	7471B	PPB	21 Jun 2023 11:11:19	U	SMPL
O3337-04 - 1	WC-17	934	0.0948	-	7471B	PPB	21 Jun 2023 11:13:35	U	SMPL
O3337-07 - 1	WC-9	1439	0.1720	-	7471B	PPB	21 Jun 2023 11:15:52	U	SMPL
O3337-10 - 1	WC-8	1133	0.1252	-	7471B	PPB	21 Jun 2023 11:18:08	U	SMPL
CCV92 - 1	CCV92	32135	4.8609	-	7471B	PPB	21 Jun 2023 11:20:24	U	SMPL
CCB92 - 1	CCB92	83	-0.0352	-	7471B	PPB	21 Jun 2023 11:22:40	U	SMPL
O3337-13 - 1	WC-7	737	0.0647	-	7471B	PPB	21 Jun 2023 11:24:58	U	SMPL
O3337-16 - 1	WC-15	1822	0.2305	-	7471B	PPB	21 Jun 2023 11:27:14	U	SMPL
O3337-19 - 1	WC-11	1000	0.1049	-	7471B	PPB	21 Jun 2023 11:29:29	U	SMPL
O3335-01LX5 - 1		163	-0.0229	-	7471B	PPB	21 Jun 2023 11:36:00	U	SMPL
O3335-01A - 1		28080	4.2414	-	7471B	PPB	21 Jun 2023 11:38:16	U	SMPL
CCV93 - 1	CCV93	32209	4.8722	-	7471B	PPB	21 Jun 2023 11:40:33	U	SMPL
CCB93 - 1	CCB93	37	-0.0422	-	7471B	PPB	21 Jun 2023 11:42:54	U	SMPL