


LB126

7470A

Linear

 μ Abs.:

48539



A= 0.0000e+000

B= 2.0659e-004

C= 1.4062e-002

Rho= 0.9999134

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.014	0.014	-2	0.000	-2	0				-7
0.2	0.200	0.187	-0.013	836	0.0 %	836					0
2.5	2.500	2.494	-0.006	12004	0.0 %	12004					0
5.0	5.000	5.057	0.057	24411	0.0 %	24411					1
7.5	7.500	7.407	-0.093	35783	0.0 %	35783					1
10.0	10.000	10.042	0.042	48539	0.0 %	48539					0

LB126101 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Jun 2023 14:20:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.0000	7470A	PPB	21 Jun 2023 14:26:05	S	Std
0.2 - 1	50.2	836	-	0.2000	7470A	PPB	21 Jun 2023 14:28:23	S	Std
2.5 - 1	52.5	12004	-	2.5000	7470A	PPB	21 Jun 2023 14:30:40	S	Std
5.0 - 1	55	24411	-	5.0000	7470A	PPB	21 Jun 2023 14:32:58	S	Std
7.5 - 1	57.5	35783	-	7.5000	7470A	PPB	21 Jun 2023 14:35:14	S	Std
10.0 - 1	516	48539	-	10.0000	7470A	PPB	21 Jun 2023 14:37:32	S	Std
ICV33 - 1	ICV33	18518	3.8397	-	7470A	PPB	21 Jun 2023 14:40:24	U	SMPL
ICB33 - 1	ICB33	-228	-0.0330	-	7470A	PPB	21 Jun 2023 14:42:39	U	SMPL
CCV94 - 1	CCV94	24146	5.0024	-	7470A	PPB	21 Jun 2023 14:44:55	U	SMPL
CCB94 - 1	CCB94	-216	-0.0306	-	7470A	PPB	21 Jun 2023 14:47:10	U	SMPL
CRA - 1	CRA	770	0.1731	-	7470A	PPB	21 Jun 2023 14:49:28	U	SMPL
HighStd - 1	HighStd	45482	9.4103	-	7470A	PPB	21 Jun 2023 14:51:44	U	SMPL
ChkStd - 1	ChkStd	33097	6.8516	-	7470A	PPB	21 Jun 2023 14:54:00	U	SMPL
PB153620BL - 1	PBW	-262	-0.0401	-	7470A	PPB	21 Jun 2023 14:56:18	U	SMPL
PB153620BS - 1	LCSW	17703	3.6714	-	7470A	PPB	21 Jun 2023 14:58:36	U	SMPL
O3327-02 - 1	COMPOSIT-TW-03	-132	-0.0132	-	7470A	PPB	21 Jun 2023 15:00:53	U	SMPL
O3327-08 - 1	COMPOSIT-TW-02	-39	0.0060	-	7470A	PPB	21 Jun 2023 15:03:10	U	SMPL
O3327-14 - 1	COMPOSIT-TW-06	-21	0.0097	-	7470A	PPB	21 Jun 2023 15:05:27	U	SMPL
O3327-20 - 1	COMPOSIT-SB-14	40	0.0223	-	7470A	PPB	21 Jun 2023 15:07:44	U	SMPL
O3328-02 - 1	COMPOSIT-SW02-8-15	37	0.0217	-	7470A	PPB	21 Jun 2023 15:10:00	U	SMPL
CCV95 - 1	CCV95	23447	4.8580	-	7470A	PPB	21 Jun 2023 15:12:15	U	SMPL
CCB95 - 1	CCB95	-198	-0.0268	-	7470A	PPB	21 Jun 2023 15:14:31	U	SMPL
O3328-08 - 1	COMPOSIT-SW04-0-8	37	0.0217	-	7470A	PPB	21 Jun 2023 15:16:48	U	SMPL
O3328-14 - 1	COMPOSIT-SW02-0-8	56	0.0256	-	7470A	PPB	21 Jun 2023 15:19:04	U	SMPL
O3328-20 - 1	COMPOSIT-SW04-8-15	-2	0.0136	-	7470A	PPB	21 Jun 2023 15:21:19	U	SMPL
O3329-02 - 1	COMPOSIT-SW03-0-8	41	0.0225	-	7470A	PPB	21 Jun 2023 15:23:35	U	SMPL
O3329-08 - 1	COMPOSIT-SW03-8-15	14	0.0170	-	7470A	PPB	21 Jun 2023 15:25:51	U	SMPL
O3329-14 - 1	COMPOSIT-SW01-0-8	35	0.0213	-	7470A	PPB	21 Jun 2023 15:28:08	U	SMPL
O3329-20 - 1	COMPOSIT-SW01-8-15	22	0.0186	-	7470A	PPB	21 Jun 2023 15:30:24	U	SMPL
O3330-02 - 1	COMPOSITE-MC03	17	0.0176	-	7470A	PPB	21 Jun 2023 15:32:42	U	SMPL
O3330-08 - 1	COMPOSITE-MC01	11	0.0163	-	7470A	PPB	21 Jun 2023 15:34:59	U	SMPL
O3330-14 - 1	COMPOSITE-MC02	45	0.0234	-	7470A	PPB	21 Jun 2023 15:37:15	U	SMPL
CCV96 - 1	CCV96	23878	4.9471	-	7470A	PPB	21 Jun 2023 15:39:31	U	SMPL
CCB96 - 1	CCB96	-190	-0.0252	-	7470A	PPB	21 Jun 2023 15:41:47	U	SMPL
O3331-02 - 1	COMPOSITE-SB8	38	0.0219	-	7470A	PPB	21 Jun 2023 15:44:04	U	SMPL
O3331-09 - 1	COMPOSITE-SB17	4	0.0149	-	7470A	PPB	21 Jun 2023 15:46:20	U	SMPL
O3331-15 - 1	COMPOSITE-SB9	8	0.0157	-	7470A	PPB	21 Jun 2023 15:48:36	U	SMPL
O3331-21 - 1	COMPOSITE-SB30-35	35	0.0213	-	7470A	PPB	21 Jun 2023 15:50:52	U	SMPL
O3331-27 - 1	COMPOSTIE-SB18-27	55	0.0254	-	7470A	PPB	21 Jun 2023 15:53:08	U	SMPL
O3331-27DUP - 1	COMPOSTIE-SB18-27DUP	55	0.0254	-	7470A	PPB	21 Jun 2023 15:55:25	U	SMPL
O3331-27MS - 1	COMPOSTIE-SB18-27MS	17474	3.6241	-	7470A	PPB	21 Jun 2023 15:57:41	U	SMPL
O3331-27MSD - 1	COMPOSTIE-SB18-27MSD	17580	3.6460	-	7470A	PPB	21 Jun 2023 16:04:44	U	SMPL
PB153623BL - 1	PBW	-35	0.0068	-	7470A	PPB	21 Jun 2023 16:07:01	U	SMPL
PB153623BS - 1	LCSW	17899	3.7119	-	7470A	PPB	21 Jun 2023 16:09:18	U	SMPL
CCV97 - 1	CCV97	23800	4.9310	-	7470A	PPB	21 Jun 2023 16:14:52	U	SMPL
CCB97 - 1	CCB97	-181	-0.0233	-	7470A	PPB	21 Jun 2023 16:17:08	U	SMPL
O3311-01 - 1	1646	2322	0.4938	-	7470A	PPB	21 Jun 2023 16:24:10	U	SMPL
O3311-01DUP - 1	1646DUP	2266	0.4822	-	7470A	PPB	21 Jun 2023 16:26:26	U	SMPL
O3311-01MS - 1	1646MS	15221	3.1586	-	7470A	PPB	21 Jun 2023 16:28:42	U	SMPL
O3311-01MSD - 1	1646MSD	14807	3.0731	-	7470A	PPB	21 Jun 2023 16:30:58	U	SMPL
O3311-03 - 1	6923	533	0.1242	-	7470A	PPB	21 Jun 2023 16:33:16	U	SMPL
O3315-04 - 1	34344	376	0.0917	-	7470A	PPB	21 Jun 2023 16:35:34	U	SMPL
O3315-05 - 1	60104	325	0.0812	-	7470A	PPB	21 Jun 2023 16:37:51	U	SMPL
CCV98 - 1	CCV98	23344	4.8368	-	7470A	PPB	21 Jun 2023 16:40:08	U	SMPL
CCB98 - 1	CCB98	-135	-0.0138	-	7470A	PPB	21 Jun 2023 16:42:25	U	SMPL
PB153577TB - 1	PB153577TB	-35	0.0068	-	7470A	PPB	21 Jun 2023 16:44:42	U	SMPL
O3331-27LX5 - 1		-31	0.0077	-	7470A	PPB	21 Jun 2023 16:46:57	U	SMPL
O3331-27A - 1		18834	3.9050	-	7470A	PPB	21 Jun 2023 16:49:14	U	SMPL
O3311-01LX5 - 1		366	0.0897	-	7470A	PPB	21 Jun 2023 16:51:29	U	SMPL
O3311-01A - 1		19376	4.0170	-	7470A	PPB	21 Jun 2023 16:53:45	U	SMPL
CCV99 - 1	CCV99	23785	4.9279	-	7470A	PPB	21 Jun 2023 16:56:01	U	SMPL
CCB99 - 1	CCB99	-186	-0.0244	-	7470A	PPB	21 Jun 2023 16:58:17	U	SMPL