



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126102

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	06/21/2023	10:00
2	O3301-01		1	12	100.00	>212.0	>212.0	06/21/2023	10:30
3	O3301-01DUP		1	14	100.00	>212.0	>212.0	06/21/2023	11:00
4	O3305-01		1	12	100.00	>212.0	>212.0	06/21/2023	11:30
5	O3311-01		1	12	100.00	>212.0	>212.0	06/21/2023	12:00
6	O3311-03		1	14	100.00	>212.0	>212.0	06/21/2023	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB126162

WorkList Name : fp 6-19

WorkList ID : 171035

Department : Wet-Chemistry

Date : 06-19-2023 08:40:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3305-01	NEGATIVE-MH-BT-TRACK-89-I	Water	Flash Point	Cool 4 deg C	TULL02	G11	06/16/2023	1010B
O3301-01 D	1619	Water	Flash Point	Cool 4 deg C	PSEG03	G11	06/16/2023	1010B
O3311-01 F	1646	Water	Flash Point	Cool 4 deg C	PSEG03	G11	06/16/2023	1010B
O3311-03 C	6923	Water	Flash Point	Cool 4 deg C	PSEG03	G11	06/16/2023	1010B

Date/Time 06/24/2023 09:58
 Raw Sample Received by: RM (wv)
 Raw Sample Relinquished by: JDC (CD C)

Date/Time 06/24/2023 12:35
 Raw Sample Received by: RM (CD C)
 Raw Sample Relinquished by: RM (wv)