



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 06/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126104

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3343-01	WASTE	70	1.14	8.55	9.69	8.39	84.8	
O3343-02	VOC	71	1.12	8.65	9.77	8.88	89.7	
O3343-03	1	72	1.12	8.73	9.85	8.4	83.4	
O3343-04	2	73	1.19	8.43	9.62	8.46	86.2	
O3343-05	3	74	1.16	8.65	9.81	8.34	83.0	
O3343-06	4	75	1.19	8.62	9.81	8.27	82.1	
O3343-07	5	76	1.12	8.86	9.98	8.78	86.5	
O3355-01	SB-01-6.5-7.0	1	1.13	8.47	9.6	9.21	95.4	
O3355-02	SB-01-COMP	2	1.14	8.49	9.63	8.89	91.3	
O3355-03	SB-02-6.5-7.0	3	1.11	8.43	9.54	9.06	94.3	
O3355-04	SB-02-COMP	4	1.13	8.61	9.74	9.44	96.5	
O3355-05	SB-03-6.5-7.0	5	1.18	8.56	9.74	8.67	87.5	
O3355-06	SB-03-COMP	6	1.13	8.51	9.64	8.45	86.0	
O3355-07	SB-04-6.5-7.0	7	1.15	8.69	9.84	8.26	81.8	
O3355-08	SB-04-COMP	8	1.16	8.54	9.7	8.24	82.9	
O3355-09	SB-05-6.5-7.0	9	1.12	8.45	9.57	9.06	94.0	
O3355-10	SB-05-COMP	10	1.15	8.37	9.52	9.02	94.0	
O3355-11	SB-06-6.5-7.0	11	1.15	8.65	9.8	8.98	90.5	
O3355-12	SB-06-COMP	12	1.16	8.83	9.99	9.18	90.8	
O3355-13	SB-07-6.5-7.0	13	1.16	8.81	9.97	9.6	95.8	
O3355-14	SB-07-COMP	14	1.18	8.64	9.82	9.37	94.8	
O3357-01	WC-SCAN-01	15	1.14	8.69	9.83	9.16	92.3	
O3357-04	WC-SCAN-01-G	16	1.1	8.61	9.71	8.98	91.5	
O3357-06	WC-SCAN-02	17	1.14	8.70	9.84	9.04	90.8	
O3357-09	WC-SCAN-02-G	18	1.13	8.53	9.66	9.2	94.6	
O3357-11	WC-SCAN-03	19	1.13	8.44	9.57	9.22	95.9	
O3357-14	WC-SCAN-03-G	20	1.12	8.55	9.67	9.17	94.2	
O3357-16	WC-SCAN-04	21	1.13	8.71	9.84	9.63	97.6	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 06/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126104

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3357-19	WC-SCAN-04-G	22	1.13	8.37	9.5	9.29	97.5	
O3357-21	WC-SCAN-EPH-01	23	1.13	8.52	9.65	8.94	91.7	
O3357-22	WC-SCAN-EPH-02	24	1.15	8.50	9.65	9.12	93.8	
O3357-23	WC-SCAN-EPH-03	25	1.16	8.59	9.75	9.02	91.5	
O3357-24	WC-SCAN-EPH-04	26	1.12	8.55	9.67	8.92	91.2	
O3357-25	WC-SCAN-EPH-05	27	1.09	8.89	9.98	9.19	91.1	
O3357-26	WC-SCAN-EPH-06	28	1.11	8.87	9.98	9.44	93.9	
O3357-27	WC-SCAN-EPH-07	29	1.13	8.57	9.7	9.22	94.4	
O3357-28	WC-SCAN-EPH-08	30	1.1	8.65	9.75	9.25	94.2	
O3357-29	WC-SCAN-EPH-09	31	1.12	8.86	9.98	9.6	95.7	
O3357-30	WC-SCAN-EPH-10	32	1.12	8.80	9.92	9.32	93.2	
O3357-31	WC-SCAN-EPH-11	33	1.14	8.64	9.78	9.64	98.4	
O3357-32	WC-SCAN-EPH-12	34	1.14	8.83	9.97	9.66	96.5	
O3357-33	WC-SCAN-EPH-13	35	1.11	8.88	9.99	9.25	91.7	
O3358-01	SB-1-(0-2)	36	1.15	8.83	9.98	8.85	87.2	
O3358-02	SB-1-(8-10)	37	1.11	8.57	9.68	8.97	91.7	
O3358-03	SB-1-(34-36)	38	1.18	8.56	9.74	9.34	95.3	
O3358-04	SB-4-(0-2)	39	1.14	8.73	9.87	8.3	82.0	
O3358-05	SB-4-(8-10)	40	1.13	8.81	9.94	9.44	94.3	
O3358-06	SB-4-(22-26)	41	1.1	8.82	9.92	9.38	93.9	
O3358-07	SB-6-(0-2)	42	1.1	8.58	9.68	7.37	73.1	
O3358-08	SB-6-(8-10)	43	1.11	8.71	9.82	9.11	91.8	
O3358-09	SB-6-(20-23)	44	1.13	8.71	9.84	7.28	70.6	
O3358-12	DUP-01	45	1.11	8.78	9.89	8.12	79.8	
O3360-01	CE-91-062023	46	1.12	8.68	9.8	9.52	96.8	
O3360-03	CE-91-062023	47	1.12	8.68	9.8	9.5	96.5	
O3360-05	CE-88N-062023	48	1.14	8.75	9.89	9.47	95.2	
O3360-07	CE-88N-062023	49	1.12	8.69	9.81	9.12	92.1	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 06/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126104

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3361-01	COMPOSIT-SB-28-29	50	1.13	8.85	9.98	9.46	94.1	
O3361-03	EPH-SB-28-29	51	1.14	8.46	9.6	9.14	94.6	
O3361-07	COMPOSIT-SB-31-36	52	1.17	8.66	9.83	9.42	95.3	
O3361-09	EPH-SB-31-36	53	1.14	8.42	9.56	9.11	94.7	
O3361-13	COMPOSIT-SB-21-23	54	1.15	8.69	9.84	9.44	95.4	
O3361-15	EPH-SB-21-23	55	1.14	8.83	9.97	9.55	95.2	
O3361-19	COMPOSIT-SB-7-22	56	1.16	8.51	9.67	9.22	94.7	
O3361-21	EPH-SB-7-22	57	1.13	8.80	9.93	9.5	95.1	
O3362-01	COMPOSIT-SB-32-37	58	1.14	8.81	9.95	9.55	95.5	
O3362-03	EPH-SB-32-37	59	1.13	8.78	9.91	9.26	92.6	
O3362-08	COMPOSIT-SB-12-13	60	1.11	8.59	9.7	9.14	93.5	
O3362-10	EPH-SB-12-13	61	1.12	8.86	9.98	9.53	94.9	
O3362-14	COMPOSIT-SB-19-20	62	1.12	8.82	9.94	9.44	94.3	
O3362-16	EPH-SB-19-20	63	1.13	8.68	9.81	9.35	94.7	
O3362-20	COMPOSIT-SB-25	64	1.16	8.63	9.79	9.51	96.8	
O3362-22	EPH-SB-25	65	1.12	8.72	9.84	9.4	95.0	
O3363-01	SP	77	1.15	8.55	9.7	8.82	89.7	
O3363-02	SP-EPH-2	78	1.15	8.55	9.7	8.82	89.7	
O3363-03	CONCRETE-BLOCKING	79	1.00	1.00	2.00	2.00	100.0	CONCRETE - BLOCKING SAMPLE
O3365-01	CONC-1	80	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3365-02	CONC-2	81	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
O3365-03	TP2	82	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-05	TP2-A	83	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-06	TP2-B	84	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-07	TP2-C	85	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-08	TP2-D	86	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-09	TP2-E	87	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-10	TP2-F	88	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 06/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126104

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3365-11	TP2-G	89	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-12	TP2-H	90	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-13	TP4	91	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-15	TP4-A	92	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-16	TP4-B	93	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-17	TP4-C	94	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-18	TP4-D	95	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-19	TP4-E	96	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-20	TP4-F	97	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-21	TP4-G	98	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-22	TP4-H	99	1.00	1.00	2.00	2.00	100.0	poly-stone SAMPLE
O3365-24	TP1	100	1.12	8.47	9.59	9.08	94.0	
O3365-25	TP1-A	101	1.14	8.76	9.9	9.34	93.6	
O3365-26	TP1-B	102	1.18	8.69	9.87	9.41	94.7	
O3365-27	TP1-C	103	1.18	8.59	9.77	9.36	95.2	
O3365-28	TP1-D	104	1.12	8.77	9.89	9.28	93.0	
O3365-29	TP1-E	105	1.15	8.83	9.98	9.44	93.9	
O3365-30	TP1-F	106	1.11	8.88	9.99	9.54	94.9	
O3365-31	TP1-G	107	1.19	8.41	9.6	9.1	94.1	
O3365-32	TP1-H	108	1.13	8.81	9.94	9.48	94.8	
O3365-33	TP3	109	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-35	TP3-A	110	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-36	TP3-B	111	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-37	TP3-C	112	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-38	TP3-D	113	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-39	TP3-E	114	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-40	TP3-F	115	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3365-41	TP3-G	116	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/22/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 06/22/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB126104

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O3365-42	TP3-H	117	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
O3366-01	SB37	66	1.15	8.84	9.99	7.8	75.2	
O3366-02	WC37	67	1.13	8.85	9.98	8.17	79.5	
O3366-03	SB38	68	1.13	8.76	9.89	8.69	86.3	
O3366-04	WC38	69	1.09	8.88	9.97	8.6	84.6	
O3367-01	EO-3-062123	118	1.18	8.53	9.71	8.99	91.6	
O3367-02	EO-3-062123-EPH-2	119	1.15	8.82	9.97	9.16	90.8	
O3368-01	CF-28	120	1.17	8.80	9.97	9.17	90.9	
O3368-02	CF-29	121	1.1	8.65	9.75	8.42	84.6	
O3369-01	ORA-1650	122	1.18	8.60	9.78	8.44	84.4	
O3369-02	ORA-1650-EPH-2	123	1.18	8.60	9.78	8.44	84.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$