

LB126107


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56124



A= 0.0000e+000

B= 1.7862e-004 slope

C= -3.7811e-002 y-intercept

Rho= 0.9998903

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.020	-0.020	102	0.000	102					1
0.05	0.050					302	0				-4
0.20	0.200	0.192	-0.008	1285	0.0 %	1285					1
2.50	2.500	2.480	-0.020	14096	0.0 %	14096					2
5.00	5.000	5.117	0.117	28860	0.0 %	28860					-1
7.50	7.500	7.444	-0.056	41885	0.0 %	41885					0
10.0	10.000	9.987	-0.013	56124	0.0 %	56124					

LB126107
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	102	-		0 SFAM01.1	PPB	6/21/2023 17:42	Std
	0.2 S01	1285	-		0.2 SFAM01.1	PPB	6/21/2023 17:44	Std
	2.5 S02	14096	-		2.5 SFAM01.1	PPB	6/21/2023 17:49	Std
	5 S03	28860	-		5 SFAM01.1	PPB	6/21/2023 17:51	Std
	7.5 S04	41885	-		7.5 SFAM01.1	PPB	6/21/2023 17:53	Std
	10 S05	56124	-		10 SFAM01.1	PPB	6/21/2023 17:58	Std
ICV034	ICV034	22806	4.0358 -		SFAM01.1	PPB	6/21/2023 18:01	SMPL
ICB034	ICB034	-10	-0.0396 -		SFAM01.1	PPB	6/21/2023 18:04	SMPL
CCV001	CCV001	28964	5.1357 -		SFAM01.1	PPB	6/21/2023 18:06	SMPL
CCB001	CCB001	-24	-0.0421 -		SFAM01.1	PPB	6/21/2023 18:08	SMPL
PB153674BL	PBS674	95	-0.0208 -		SFAM01.1	PPB	6/21/2023 18:11	SMPL
O3102-01	MEZEL6	257	0.0081 -		SFAM01.1	PPB	6/21/2023 18:13	SMPL
O3102-02	MEZEL7	643	0.077 -		SFAM01.1	PPB	6/21/2023 18:15	SMPL
O3102-03	MEZEL8	250	0.0068 -		SFAM01.1	PPB	6/21/2023 18:17	SMPL
O3102-04	MEZEL8D	330	0.0211 -		SFAM01.1	PPB	6/21/2023 18:20	SMPL
O3102-05	MEZEL8S	12003	2.1062 -		SFAM01.1	PPB	6/21/2023 18:22	SMPL
O3102-06	MEZEL9	232	0.0036 -		SFAM01.1	PPB	6/21/2023 18:24	SMPL
O3102-07	MEZEM0	1268	0.1887 -		SFAM01.1	PPB	6/21/2023 18:26	SMPL
O3102-08	MEZEM4	3886	0.6563 -		SFAM01.1	PPB	6/21/2023 18:29	SMPL
O3102-09	MEZEM5	1859	0.2942 -		SFAM01.1	PPB	6/21/2023 18:31	SMPL
O3102-10	MEZEM6	1705	0.2667 -		SFAM01.1	PPB	6/21/2023 18:33	SMPL
O3102-11	MEZEN0	204	-0.0014 -		SFAM01.1	PPB	6/21/2023 18:35	SMPL
PB153675BL	PBS675	66	-0.026 -		SFAM01.1	PPB	6/21/2023 18:38	SMPL
O3112-19	X3G79	14795	2.6049 -		SFAM01.1	PPB	6/21/2023 18:40	SMPL
O3112-20	X3G79D	14933	2.6295 -		SFAM01.1	PPB	6/21/2023 18:45	SMPL
O3112-21	X3G79S	28261	5.0102 -		SFAM01.1	PPB	6/21/2023 18:51	SMPL
CCV002	CCV002	29189	5.1759 -		SFAM01.1	PPB	6/21/2023 18:54	SMPL
CCB002	CCB002	-32	-0.0435 -		SFAM01.1	PPB	6/21/2023 18:56	SMPL