



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB126113

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/22/2023	12:10
2	O3315-01		1	4	9.00	48.2	47.9	06/22/2023	12:40
3	O3315-01DUP		1	4	9.00	48.2	47.9	06/22/2023	13:10
4	O3345-02		1	11	100.00	>212.0	>212.0	06/22/2023	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

2626113

WorkList Name : FP S 19

WorkList ID : 171080

Department : Wet-Chemistry

Date : 06-19-2023 12:20:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3315-01	31263	Solid	Flash Point	Cool 4 deg C	PSEG03	G11	06/16/2023	1010B
O3345-02	5773	Solid	Flash Point	Cool 4 deg C	PSEG03	H11	06/20/2023	1010B

Date/Time 06/22/2023 12:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06/22/2023 14:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]