

LB126126


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41123



A= 0.0000e+000

B= 2.4625e-004

C= -6.3120e-002

Rho= 0.9997528

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	-0.002	-0.002	248	0.000	248	0				
0.2	0.200	0.173	-0.027	960	0.0 %	960					-14
2.5	2.500	2.639	0.139	10971	0.0 %	10971					6
5.0	5.000	4.887	-0.113	20101	0.0 %	20101					-2
7.5	7.500	7.440	-0.060	30468	0.0 %	30468					-1
10.0	10.000	10.064	0.064	41123	0.0 %	41123					1

LB126120 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jun 2023 12:15:08

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	248	-	0.0000	7470A	PPB	22 Jun 2023 12:19:23	S	Std
0.2 - 1	502	960	-	0.2000	7470A	PPB	22 Jun 2023 12:21:39	S	Std
2.5 - 1	52.5	10971	-	2.5000	7470A	PPB	22 Jun 2023 12:23:56	S	Std
5.0 - 1	55	20101	-	5.0000	7470A	PPB	22 Jun 2023 12:26:13	S	Std
7.5 - 1	57.5	30468	-	7.5000	7470A	PPB	22 Jun 2023 12:28:29	S	Std
10.0 - 1	510	41123	-	10.0000	7470A	PPB	22 Jun 2023 12:36:00	S	Std
ICV35 - 1	ICV35	17974	4.3631	-7470A	PPB	22 Jun 2023 12:38:46	U	SMPL	
ICB35 - 1	ICB35	148	-0.0267	-7470A	PPB	22 Jun 2023 12:41:00	U	SMPL	
CCV03 - 1	CCV03	20284	4.9319	-7470A	PPB	22 Jun 2023 12:46:03	U	SMPL	
CCB03 - 1	CCB03	233	-0.0057	-7470A	PPB	22 Jun 2023 12:48:18	U	SMPL	
CRA - 1	CRA	933	0.1666	-7470A	PPB	22 Jun 2023 12:50:33	U	SMPL	
HighStd - 1	HighStd	40492	9.9082	-7470A	PPB	22 Jun 2023 12:52:48	U	SMPL	
ChkStd - 1	ChkStd	29853	7.2883	-7470A	PPB	22 Jun 2023 12:55:04	U	SMPL	
PB153672BL - 1	PBW	160	-0.0237	-7470A	PPB	22 Jun 2023 12:59:57	U	SMPL	
PB153672BS - 1	LCSW	16626	4.0311	-7470A	PPB	22 Jun 2023 13:02:13	U	SMPL	
O3333-02 - 1	DOMINICK-SP	-27	-0.0698	-7470A	PPB	22 Jun 2023 13:04:29	U	SMPL	
O3336-02 - 1	WC-6	84	-0.0424	-7470A	PPB	22 Jun 2023 13:06:46	U	SMPL	
O3336-05 - 1	WC-5	61	-0.0481	-7470A	PPB	22 Jun 2023 13:09:03	U	SMPL	
O3336-08 - 1	WC-4	115	-0.0348	-7470A	PPB	22 Jun 2023 13:11:20	U	SMPL	
O3336-11 - 1	WC-3	125	-0.0323	-7470A	PPB	22 Jun 2023 13:13:36	U	SMPL	
CCV04 - 1	CCV04	20844	5.0698	-7470A	PPB	22 Jun 2023 13:15:51	U	SMPL	
CCB04 - 1	CCB04	33	-0.0550	-7470A	PPB	22 Jun 2023 13:26:51	U	SMPL	
O3336-14 - 1	WC-2	119	-0.0338	-7470A	PPB	22 Jun 2023 13:29:05	U	SMPL	
O3336-17 - 1	WC-1	123	-0.0328	-7470A	PPB	22 Jun 2023 13:31:21	U	SMPL	
O3337-02 - 1	WC-18	131	-0.0309	-7470A	PPB	22 Jun 2023 13:33:36	U	SMPL	
O3337-05 - 1	WC-17	193	-0.0156	-7470A	PPB	22 Jun 2023 13:35:52	U	SMPL	
O3337-08 - 1	WC-9	113	-0.0353	-7470A	PPB	22 Jun 2023 13:38:08	U	SMPL	
O3337-11 - 1	WC-8	136	-0.0296	-7470A	PPB	22 Jun 2023 13:40:24	U	SMPL	
O3337-14 - 1	WC-7	109	-0.0363	-7470A	PPB	22 Jun 2023 13:42:40	U	SMPL	
O3337-17 - 1	WC-15	111	-0.0358	-7470A	PPB	22 Jun 2023 13:44:57	U	SMPL	
O3337-20 - 1	WC-11	131	-0.0309	-7470A	PPB	22 Jun 2023 13:47:14	U	SMPL	
O3345-01 - 1	COMP-1-0620	165	-0.0225	-7470A	PPB	22 Jun 2023 13:49:29	U	SMPL	
CCV05 - 1	CCV05	20815	5.0627	-7470A	PPB	22 Jun 2023 13:51:45	U	SMPL	
CCB05 - 1	CCB05	35	-0.0545	-7470A	PPB	22 Jun 2023 13:55:53	U	SMPL	
O3345-01DUP - 1	COMP-1-0620DUP	258	0.0004	-7470A	PPB	22 Jun 2023 13:58:09	U	SMPL	
O3345-02 - 1	5773	-43	-0.0737	-7470A	PPB	22 Jun 2023 14:04:57	U	SMPL	
O3345-03 - 1	COMP-2-0620	-249	-0.1244	-7470A	PPB	22 Jun 2023 14:11:48	U	SMPL	
O3345-03DUP - 1	COMP-2-0620DUP	-39	-0.0727	-7470A	PPB	22 Jun 2023 14:14:04	U	SMPL	
O3345-03MS - 1	COMP-2-0620MS	18137	4.4032	-7470A	PPB	22 Jun 2023 14:25:33	U	SMPL	
O3345-03MSD - 1	COMP-2-0620MSD	15933	3.8605	-7470A	PPB	22 Jun 2023 14:27:48	U	SMPL	
O3345-01MS - 1	COMP-1-0620MS	14750	3.5691	-7470A	PPB	22 Jun 2023 14:30:38	U	SMPL	
O3345-01MSD - 1	COMP-1-0620MSD	17525	4.2525	-7470A	PPB	22 Jun 2023 14:38:02	U	SMPL	
PB153673BL - 1	PBW	40	-0.0533	-7470A	PPB	22 Jun 2023 14:40:48	U	SMPL	
PB153673BS - 1	LCSW	15830	3.8351	-7470A	PPB	22 Jun 2023 14:48:30	U	SMPL	
CCV06 - 1	CCV06	18900	4.5911	-7470A	PPB	22 Jun 2023 14:51:09	U	SMPL	
CCB06 - 1	CCB06	-16	-0.0671	-7470A	PPB	22 Jun 2023 14:53:23	U	SMPL	
O3345-04 - 1	5774	188	-0.0168	-7470A	PPB	22 Jun 2023 14:55:41	U	SMPL	
PB153608TB - 1		36	-0.0543	-7470A	PPB	22 Jun 2023 14:57:56	U	SMPL	
PB153666TB - 1		655	0.0982	-7470A	PPB	22 Jun 2023 15:00:12	U	SMPL	
O3345-01LX5 - 1		98	-0.0390	-7470A	PPB	22 Jun 2023 15:02:27	U	SMPL	
O3345-01A - 1		16001	3.8772	-7470A	PPB	22 Jun 2023 15:04:43	U	SMPL	
O3345-03LX5 - 1		53	-0.0501	-7470A	PPB	22 Jun 2023 15:06:59	U	SMPL	
O3345-03A - 1		16132	3.9095	-7470A	PPB	22 Jun 2023 15:09:16	U	SMPL	
CCV07 - 1	CCV07	18590	4.5148	-7470A	PPB	22 Jun 2023 15:11:32	U	SMPL	
CCB07 - 1	CCB07	30	-0.0557	-7470A	PPB	22 Jun 2023 15:13:49	U	SMPL	