

Analysis Method: 9034

ANALYST: Rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB126123

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2976
IODINE SOLUTION .025N 1L	W2924
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB153678BL		1	5.00	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	06/22/2023	09:00
2	O3331-02		1	5.06	50	2.00	0.00	1.90	1.90	0.10	0.02	1.58	06/22/2023	09:03
3	O3331-02DUP		1	5.06	50	2.00	0.00	1.90	1.90	0.10	0.02	1.58	06/22/2023	09:06
4	O3331-09		1	5.02	50	2.00	0.00	1.86	1.86	0.14	0.06	4.78	06/22/2023	09:08
5	O3331-15		1	5.05	50	2.00	0.00	1.82	1.82	0.18	0.10	7.92	06/22/2023	09:10
6	O3331-21		1	5.04	50	2.00	0.00	1.84	1.84	0.16	0.08	6.35	06/22/2023	09:13
7	O3331-27		1	5.03	50	2.00	0.00	1.88	1.88	0.12	0.04	3.18	06/22/2023	09:15
8	O3355-02		1	5.04	50	2.00	0.00	1.84	1.84	0.16	0.08	6.35	06/22/2023	09:18
9	O3355-04		1	5.05	50	2.00	0.00	1.82	1.82	0.18	0.10	7.92	06/22/2023	09:20
10	O3355-06		1	5.06	50	2.00	0.00	1.88	1.88	0.12	0.04	3.16	06/22/2023	09:22
11	O3355-08		1	5.02	50	2.00	0.00	1.82	1.82	0.18	0.10	7.97	06/22/2023	09:25
12	O3355-10		1	5.06	50	2.00	0.00	1.86	1.86	0.14	0.06	4.74	06/22/2023	09:28

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB126123

ANALYST: Rubina

SUPERVISOR REVIEW BY: Iwona

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2976
IODINE SOLUTION .025N 1L	W2924
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
13	03355-12		1	5.02	50	2.00	0.00	1.82	1.82	0.18	0.10	7.97	06/22/2023	09:30
14	03355-14		1	5.06	50	2.00	0.00	1.86	1.86	0.14	0.06	4.74	06/22/2023	09:33
15	03366-02		1	5.05	50	2.00	0.00	1.82	1.82	0.18	0.10	7.92	06/22/2023	09:36
16	03366-04		1	5.05	50	2.00	0.00	1.82	1.82	0.18	0.10	7.92	06/22/2023	09:38

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume