

LB126131


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

63297



A= 0.0000e+000

B= 1.5853e-004

C= -1.2105e-002

Rho= 0.9997053

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/v D
0.0	0.000	-0.006	-0.006	40	0.000	40					
0.2	0.200	0.237	0.037	1574	0.0 %	1574					19
2.5	2.500	2.571	0.071	16292	0.0 %	16292					3
5.0	5.000	4.808	-0.192	30407	0.0 %	30407					-4
7.5	7.500	7.567	0.067	47808	0.0 %	47808					1
10.0	10.000	10.022	0.022	63297	0.0 %	63297					0

LB126131 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jun 2023 16:31:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0000	7471B	PPB	22 Jun 2023 16:38:30	S	Std
0.2 - 1	50.2	1574	-	0.2000	7471B	PPB	22 Jun 2023 16:40:47	S	Std
2.5 - 1	52.5	16292	-	2.5000	7471B	PPB	22 Jun 2023 16:43:04	S	Std
5.0 - 1	55	30407	-	5.0000	7471B	PPB	22 Jun 2023 16:45:20	S	Std
7.5 - 1	57.5	47808	-	7.5000	7471B	PPB	22 Jun 2023 16:47:38	S	Std
10.0 - 1	516	63297	-	10.0000	7471B	PPB	22 Jun 2023 16:49:56	S	Std
ICV36 - 1	ICV36	25617	4.0490	-	7471B	PPB	22 Jun 2023 16:52:55	U	SMPL
ICB36 - 1	ICB36	-62	-0.0219	-	7471B	PPB	22 Jun 2023 16:55:09	U	SMPL
CCV08 - 1	CCV08	30388	4.8053	-	7471B	PPB	22 Jun 2023 16:57:27	U	SMPL
CCB08 - 1	CCB08	-143	-0.0348	-	7471B	PPB	22 Jun 2023 16:59:42	U	SMPL
CRA - 1	CRA	1527	0.2300	-	7471B	PPB	22 Jun 2023 17:01:59	U	SMPL
HighStd - 1	HighStd	62404	9.8808	-	7471B	PPB	22 Jun 2023 17:04:13	U	SMPL
ChkStd - 1	ChkStd	43256	6.8453	-	7471B	PPB	22 Jun 2023 17:06:29	U	SMPL
PB153704BL - 1	PBS	424	0.0551	-	7471B	PPB	22 Jun 2023 17:08:48	U	SMPL
PB153704BS - 1	LCSS	21771	3.4393	-	7471B	PPB	22 Jun 2023 17:11:05	U	SMPL
O3352-01 - 1	BU-3-062023	1737	0.2633	-	7471B	PPB	22 Jun 2023 17:13:21	U	SMPL
O3357-01 - 1	WC-SCAN-01	8147	1.2794	-	7471B	PPB	22 Jun 2023 17:15:40	U	SMPL
O3357-06 - 1	WC-SCAN-02	7909	1.2417	-	7471B	PPB	22 Jun 2023 17:17:56	U	SMPL
O3357-11 - 1	WC-SCAN-03	2767	0.4265	-	7471B	PPB	22 Jun 2023 17:20:14	U	SMPL
O3357-16 - 1	WC-SCAN-04	5916	0.9258	-	7471B	PPB	22 Jun 2023 17:22:30	U	SMPL
CCV09 - 1	CCV09	29949	4.7357	-	7471B	PPB	22 Jun 2023 17:24:45	U	SMPL
CCB09 - 1	CCB09	-151	-0.0360	-	7471B	PPB	22 Jun 2023 17:27:01	U	SMPL
O3358-01 - 1	SB-1(0-2)	16987	2.6809	-	7471B	PPB	22 Jun 2023 17:29:18	U	SMPL
O3358-02 - 1	SB-1(8-10)	1077	0.1586	-	7471B	PPB	22 Jun 2023 17:31:33	U	SMPL
O3358-03 - 1	SB-1(34-36)	83999	13.3043	-	7471B	PPB	22 Jun 2023 17:33:50	U	SMPL
O3358-04 - 1	SB-4(0-2)	939	0.1368	-	7471B	PPB	22 Jun 2023 17:36:06	U	SMPL
O3358-05 - 1	SB-4(8-10)	604	0.0836	-	7471B	PPB	22 Jun 2023 17:38:27	U	SMPL
O3358-06 - 1	SB-4(22-26)	590	0.0814	-	7471B	PPB	22 Jun 2023 17:40:43	U	SMPL
O3358-07 - 1	SB-6(0-2)	5539	0.8660	-	7471B	PPB	22 Jun 2023 17:42:59	U	SMPL
O3358-08 - 1	SB-6(8-10)	1573	0.2373	-	7471B	PPB	22 Jun 2023 17:45:16	U	SMPL
O3358-09 - 1	SB-6(20-23)	1661	0.2512	-	7471B	PPB	22 Jun 2023 17:47:33	U	SMPL
O3358-12 - 1	DUP-01	37298	5.9008	-	7471B	PPB	22 Jun 2023 17:49:48	U	SMPL
CCV10 - 1	CCV10	30308	4.7926	-	7471B	PPB	22 Jun 2023 17:52:03	U	SMPL
CCB10 - 1	CCB10	-154	-0.0365	-	7471B	PPB	22 Jun 2023 17:54:21	U	SMPL
O3360-03 - 1	CE-91-062023	91562	14.5033	-	7471B	PPB	22 Jun 2023 17:56:38	U	SMPL
O3360-07 - 1	CE-88N-062023	6123	0.9586	-	7471B	PPB	22 Jun 2023 17:58:54	U	SMPL
O3361-01 - 1	COMPOSIT-SB-28-29	50825	8.0452	-	7471B	PPB	22 Jun 2023 18:01:13	U	SMPL
O3361-07 - 1	COMPOSIT-SB-31-36	29932	4.7330	-	7471B	PPB	22 Jun 2023 18:03:29	U	SMPL
O3361-07DUP - 1	COMPOSIT-SB-31-36DUP	29560	4.6741	-	7471B	PPB	22 Jun 2023 18:10:44	U	SMPL
O3361-07MS - 1	COMPOSIT-SB-31-36MS	52880	8.3710	-	7471B	PPB	22 Jun 2023 18:13:02	U	SMPL
O3361-07MSD - 1	COMPOSIT-SB-31-36MSD	52351	8.2871	-	7471B	PPB	22 Jun 2023 18:15:20	U	SMPL
PB153705BL - 1	PBS	269	0.0305	-	7471B	PPB	22 Jun 2023 18:17:39	U	SMPL
PB153705BS - 1	LCSS	23422	3.7010	-	7471B	PPB	22 Jun 2023 18:22:31	U	SMPL
O3361-13 - 1	COMPOSIT-SB-21-23	18848	2.9759	-	7471B	PPB	22 Jun 2023 18:24:46	U	SMPL
CCV11 - 1	CCV11	30102	4.7600	-	7471B	PPB	22 Jun 2023 18:27:03	U	SMPL
CCB11 - 1	CCB11	-85	-0.0256	-	7471B	PPB	22 Jun 2023 18:29:20	U	SMPL
O3361-19 - 1	COMPOSIT-SB-7-22	73127	11.5808	-	7471B	PPB	22 Jun 2023 18:31:38	U	SMPL
O3362-01 - 1	COMPOSIT-SB-32-37	26780	4.2333	-	7471B	PPB	22 Jun 2023 18:33:53	U	SMPL
O3362-08 - 1	COMPOSIT-SB-12-13	29626	4.6845	-	7471B	PPB	22 Jun 2023 18:36:11	U	SMPL
O3362-14 - 1	COMPOSIT-SB-19-20	29712	4.6982	-	7471B	PPB	22 Jun 2023 18:38:28	U	SMPL
O3362-20 - 1	COMPOSIT-SB-25	19068	3.0108	-	7471B	PPB	22 Jun 2023 18:40:46	U	SMPL
O3363-01 - 1	SP	23435	3.7031	-	7471B	PPB	22 Jun 2023 18:43:04	U	SMPL
O3363-03 - 1	CONCRETE-BLOCKING	2309	0.3539	-	7471B	PPB	22 Jun 2023 18:45:22	U	SMPL
O3365-03 - 1	TP2	941	0.1371	-	7471B	PPB	22 Jun 2023 18:47:40	U	SMPL
O3365-13 - 1	TP4	1030	0.1512	-	7471B	PPB	22 Jun 2023 18:49:56	U	SMPL
O3365-24 - 1	TP1	1570	0.2368	-	7471B	PPB	22 Jun 2023 18:52:12	U	SMPL
CCV12 - 1	CCV12	29623	4.6840	-	7471B	PPB	22 Jun 2023 18:54:27	U	SMPL
CCB12 - 1	CCB12	-88	-0.0261	-	7471B	PPB	22 Jun 2023 18:56:42	U	SMPL
O3365-24DUP - 1	TP1DUP	1102	0.1626	-	7471B	PPB	22 Jun 2023 18:58:59	U	SMPL
O3365-24MS - 1	TP1MS	24389	3.8543	-	7471B	PPB	22 Jun 2023 19:01:15	U	SMPL
O3365-24MSD - 1	TP1MSD	23145	3.6571	-	7471B	PPB	22 Jun 2023 19:03:30	U	SMPL
O3365-33 - 1	TP3	918	0.1334	-	7471B	PPB	22 Jun 2023 19:05:47	U	SMPL
O3366-01 - 1	SB37	64174	10.1614	-	7471B	PPB	22 Jun 2023 19:08:05	U	SMPL
O3366-03 - 1	SB38	11447	1.8026	-	7471B	PPB	22 Jun 2023 19:10:21	U	SMPL
O3367-01 - 1	EO-3-062123	2056	0.3138	-	7471B	PPB	22 Jun 2023 19:12:40	U	SMPL
O3368-01 - 1	CF-28	4472	0.6968	-	7471B	PPB	22 Jun 2023 19:14:58	U	SMPL
O3368-02 - 1	CF-29	4844	0.7558	-	7471B	PPB	22 Jun 2023 19:17:14	U	SMPL
O3369-01 - 1	ORA-1650	7216	1.1319	-	7471B	PPB	22 Jun 2023 19:19:30	U	SMPL
CCV13 - 1	CCV13	29970	4.7391	-	7471B	PPB	22 Jun 2023 19:21:46	U	SMPL
CCB13 - 1	CCB13	-84	-0.0254	-	7471B	PPB	22 Jun 2023 19:24:02	U	SMPL
O3361-07LX5 - 1		4947	0.7721	-	7471B	PPB	22 Jun 2023 19:26:20	U	SMPL
O3361-07A - 1		49854	7.8913	-	7471B	PPB	22 Jun 2023 19:28:36	U	SMPL
O3365-24LX5 - 1		-147	-0.0354	-	7471B	PPB	22 Jun 2023 19:30:53	U	SMPL
O3365-24A - 1		25247	3.9903	-	7471B	PPB	22 Jun 2023 19:33:12	U	SMPL
O3358-03DLX5 - 1		15840	2.4990	-	7471B	PPB	22 Jun 2023 19:35:33	U	SMPL
O3360-03DLX5 - 1		16834	2.6566	-	7471B	PPB	22 Jun 2023 19:37:50	U	SMPL
O3361-19DLX5 - 1		15332	2.4185	-	7471B	PPB	22 Jun 2023 19:40:07	U	SMPL
O3366DLX2 - 1		32554	5.1487	-	7471B	PPB	22 Jun 2023 19:42:24	U	SMPL
CCV14 - 1	CCV14	29314	4.6351	-	7471B	PPB	22 Jun 2023 19:44:41	U	SMPL
CCB14 - 1	CCB14	-132	-0.0330	-	7471B	PPB	22 Jun 2023 19:46:58	U	SMPL