



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB126132

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2982
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/22/2023	16:10
2	CAL2	1	Water	NA	NA	20.8	7.00	06/22/2023	16:11
3	CAL3	1	Water	NA	NA	20.8	10.02	06/22/2023	16:12
4	ICV	1	Water	NA	NA	20.6	6.97	06/22/2023	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	06/22/2023	16:20
6	O3392-01	1	Solid	20.02	20	22.7	6.08	06/22/2023	16:21
7	O3392-01DUP	1	Solid	20.03	20	22.8	6.09	06/22/2023	16:22
8	O3392-03	1	Solid	20.04	20	22.8	5.92	06/22/2023	16:25
9	O3392-05	1	Solid	20.02	20	22.9	6.70	06/22/2023	16:30
10	O3392-07	1	Solid	20.04	20	22.9	7.02	06/22/2023	16:33
11	CCV2	1	Water	NA	NA	20.7	12.02	06/22/2023	16:35

U3126132

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O3392 WorkList ID : 171222 Department : Wet-Chemistry Date : 06-22-2023 15:49:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O3392-01	D RY-1	Solid	pH	Cool 4 deg C	NEWP02	H11	06/22/2023	9045D
O3392-03	C SRY-1	Solid	pH	Cool 4 deg C	NEWP02	H11	06/22/2023	9045D
O3392-05	D SRY-3	Solid	pH	Cool 4 deg C	NEWP02	H11	06/22/2023	9045D
O3392-07	H SB-64	Solid	pH	Cool 4 deg C	NEWP02	H11	06/22/2023	9045D

Date/Time 06/22/23 15:50
Raw Sample Received by: 40 (SOC)
Raw Sample Relinquished by: 2.C (3m)

Date/Time 06/22/23 18:00
Raw Sample Received by: 7-C (3m)
Raw Sample Relinquished by: 40 (SOC)